

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051420\
 Data File : BF120287.D
 Acq On : 14 May 2020 13:43
 Operator : MA/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 14 14:15:21 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 14 12:47:38 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
2	1,4-Dioxane	0.409	0.461	-12.7	124	0.00
3	Pyridine	1.285	1.221	5.0	88	0.00
4	n-Nitrosodimethylamine	0.503	0.495	1.6	90	0.00
5 S	2-Fluorophenol	1.028	1.027	0.1	89	0.00
6	Aniline	1.662	1.499	9.8	80	0.00
7 S	Phenol-d6	1.300	1.151	11.5	80	0.00
8	2-Chlorophenol	1.199	1.104	7.9	82	0.00
9	Benzaldehyde	0.801	0.795	0.7	86	0.00
10 C	Phenol	1.517	1.385	8.7	81	0.00
11	bis(2-Chloroethyl)ether	1.212	1.274	-5.1	97	0.00
12	1,3-Dichlorobenzene	1.282	1.266	1.2	87	0.00
13 C	1,4-Dichlorobenzene	1.321	1.332	-0.8	86	0.00
14	1,2-Dichlorobenzene	1.189	1.126	5.3	77	0.00
15	Benzyl Alcohol	0.923	0.871	5.6	78	0.00
16	2,2'-oxybis(1-Chloropropane	1.910	1.762	7.7	76	0.00
17	2-Methylphenol	1.012	0.960	5.1	77	0.00
18	Hexachloroethane	0.503	0.451	10.3	75	0.00
19 P	n-Nitroso-di-n-propylamine	0.852	0.766	10.1	75	0.00
20	3+4-Methylphenols	1.315	1.233	6.2	75	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	77	0.00
22	Acetophenone	0.410	0.418	-2.0	76	0.00
23 S	Nitrobenzene-d5	0.292	0.299	-2.4	75	0.00
24	Nitrobenzene	0.314	0.304	3.2	70	0.00
25	Isophorone	0.608	0.623	-2.5	77	0.00
26 C	2-Nitrophenol	0.164	0.174	-6.1	77	0.00
27	2,4-Dimethylphenol	0.239	0.278	-16.3	85	0.00
28	bis(2-Chloroethoxy)methane	0.382	0.435	-13.9	85	0.00
29 C	2,4-Dichlorophenol	0.246	0.276	-12.2	83	0.00
30	1,2,4-Trichlorobenzene	0.272	0.284	-4.4	79	0.00
31	Naphthalene	0.844	0.876	-3.8	78	0.00
32	Benzoic acid	0.157	0.178	-13.4	80	0.00
33	4-Chloroaniline	0.385	0.384	0.3	74	0.00
34 C	Hexachlorobutadiene	0.157	0.174	-10.8	84	0.00
35	Caprolactam	0.080	0.092	-15.0	85	0.00
36 C	4-Chloro-3-methylphenol	0.266	0.295	-10.9	85	0.00
37	2-Methylnaphthalene	0.578	0.613	-6.1	81	0.00
38	1-Methylnaphthalene	0.560	0.587	-4.8	81	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
40	1,2,4,5-Tetrachlorobenzene	0.504	0.540	-7.1	86	0.00
41 P	Hexachlorocyclopentadiene	0.178	0.183	-2.8	78	0.00
42 S	2,4,6-Tribromophenol	0.163	0.199	-22.1	101	0.00
43 C	2,4,6-Trichlorophenol	0.339	0.324	4.4	75	0.00
44	2,4,5-Trichlorophenol	0.389	0.384	1.3	78	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051420\
 Data File : BF120287.D
 Acq On : 14 May 2020 13:43
 Operator : MA/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 14 14:15:21 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 14 12:47:38 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	0.973	1.059	-8.8	84	0.00
46	1,1'-Biphenyl	1.310	1.409	-7.6	85	0.00
47	2-Chloronaphthalene	1.051	1.094	-4.1	83	0.00
48	2-Nitroaniline	0.384	0.399	-3.9	85	0.00
49	Acenaphthylene	1.606	1.660	-3.4	84	0.00
50	Dimethylphthalate	1.264	1.185	6.2	78	0.00
51	2,6-Dinitrotoluene	0.282	0.287	-1.8	82	0.00
52 C	Acenaphthene	0.962	1.004	-4.4	89	0.00
53	3-Nitroaniline	0.336	0.342	-1.8	87	0.00
54 P	2,4-Dinitrophenol	0.133	0.131	1.5	85	0.00
55	Dibenzofuran	1.386	1.464	-5.6	90	0.00
56 P	4-Nitrophenol	0.181	0.186	-2.8	85	0.00
57	2,4-Dinitrotoluene	0.331	0.349	-5.4	89	0.00
58	Fluorene	1.056	1.166	-10.4	92	0.00
59	2,3,4,6-Tetrachlorophenol	0.293	0.307	-4.8	87	0.00
60	Diethylphthalate	1.221	1.309	-7.2	90	0.00
61	4-Chlorophenyl-phenylether	0.549	0.588	-7.1	92	0.00
62	4-Nitroaniline	0.313	0.351	-12.1	93	0.00
63	Azobenzene	1.154	1.419	-23.0	100	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	0.101	0.098	3.0	92	0.00
66 c	n-Nitrosodiphenylamine	0.553	0.559	-1.1	99	0.00
67	4-Bromophenyl-phenylether	0.194	0.195	-0.5	99	0.00
68	Hexachlorobenzene	0.203	0.201	1.0	100	0.00
69	Atrazine	0.166	0.165	0.6	100	0.00
70 C	Pentachlorophenol	0.086	0.086	0.0	95	0.00
71	Phenanthrene	0.840	0.854	-1.7	102	0.00
72	Anthracene	0.900	0.923	-2.6	102	0.00
73	Carbazole	0.841	0.863	-2.6	102	0.00
74	Di-n-butylphthalate	1.013	1.058	-4.4	102	0.00
75 C	Fluoranthene	0.901	0.942	-4.6	102	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
77	Benzidine	0.584	0.615	-5.3	96	0.00
78	Pyrene	1.426	1.430	-0.3	92	0.00
79 S	Terphenyl-d14	0.905	0.895	1.1	94	0.00
80	Butylbenzylphthalate	0.671	0.684	-1.9	93	0.00
81	Benzo(a)anthracene	1.055	1.034	2.0	91	0.00
82	3,3'-Dichlorobenzidine	0.398	0.410	-3.0	90	0.00
83	Chrysene	1.133	1.144	-1.0	92	0.00
84	Bis(2-ethylhexyl)phthalate	0.835	0.832	0.4	93	0.00
85 c	Di-n-octyl phthalate	1.503	1.539	-2.4	92	0.00
86	Indeno(1,2,3-cd)pyrene	1.028	0.915	11.0	84	0.00
87 I	Perylene-d12	1.000	1.000	0.0	77	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051420\
Data File : BF120287.D
Acq On : 14 May 2020 13:43
Operator : MA/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 14 14:15:21 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 14 12:47:38 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.092	1.241	-13.6	82	0.00
89	Benzo(k)fluoranthene	0.946	1.002	-5.9	82	0.00
90 C	Benzo(a)pyrene	0.980	1.015	-3.6	78	0.00
91	Dibenzo(a,h)anthracene	0.868	0.946	-9.0	84	0.00
92	Benzo(q,h,i)perylene	0.875	0.932	-6.5	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0