


Data Path : Z:\HPCHEM1\BNA F\DATA\BF051517\
 Data File : BF095123.D
 Acq On : 15 May 2017 13:41
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 03 17:24:51 2017



AutoFind: Scans 2316, 2317, 2318; Background Corrected with Scan 2311

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.7	14927	PASS
68	69	0.00	2	1.1	165	PASS
69	198	0.00	100	35.6	14892	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	50.3	21031	PASS
197	198	0.00	2	0.6	234	PASS
198	198	100	100	100.0	41786	PASS
199	198	5	9	6.9	2878	PASS
275	198	10	60	24.6	10300	PASS
365	198	1	100	2.4	1010	PASS
441	442	0.01	24	15.6	5909	PASS
442	198	50	100	90.5	37824	PASS
443	442	15	24	18.1	6840	PASS