

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051520\
 Data File : BF120331.D
 Acq On : 15 May 2020 19:43
 Operator : MA/SJ
 Sample : L2663-02
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 10025

Quant Time: May 16 04:38:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 15 12:31:31 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.62	152	297934	20.00	ng	0.00
21) Naphthalene-d8	7.90	136	1200678	20.00	ng	0.00
39) Acenaphthene-d10	9.65	164	658118	20.00	ng	0.00
64) Phenanthrene-d10	11.13	188	1141183	20.00	ng	0.00
76) Chrysene-d12	13.77	240	726150	20.00	ng	0.00
87) Perylene-d12	15.16	264	586080	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.22	112	1240152	80.99	ng	0.00
7) Phenol-d6	6.26	99	1635386	84.46	ng	-0.01
23) Nitrobenzene-d5	7.18	82	1009854	57.59	ng	-0.01
42) 2,4,6-Tribromophenol	10.44	330	428517	79.86	ng	0.00
45) 2-Fluorobiphenyl	8.97	172	2068592	64.58	ng	0.00
79) Terphenyl-d14	12.72	244	2080344	63.29	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.37	163	312328	7.507	ng	98
75) Fluoranthene	12.34	202	207868	4.044	ng	96
78) Pyrene	12.57	202	190689	3.684	ng	97
81) Benzo(a)anthracene	13.76	228	99608	2.601	ng	99
83) Chrysene	13.79	228	95544	2.323	ng	96
85) Di-n-octyl phthalate	14.36	149	156677	2.872	ng	# 61
88) Benzo(b)fluoranthene	14.77	252	108837m	3.400	ng	
90) Benzo(a)pyrene	15.10	252	76944	2.681	ng	# 90
92) Benzo(g,h,i)perylene	16.87	276	52883	2.062	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051520\
Data File : BF120331.D
Acq On : 15 May 2020 19:43
Operator : MA/SJ
Sample : L2663-02
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
10025

Quant Time: May 16 04:38:39 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 15 12:31:31 2020
Response via : Initial Calibration

