

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051520\
 Data File : BF120314.D
 Acq On : 15 May 2020 11:47
 Operator : MA/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 15 12:33:17 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 15 12:31:31 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
2	1,4-Dioxane	0.409	0.504	-23.2	132	0.00
3	Pyridine	1.285	1.421	-10.6	100	0.00
4	n-Nitrosodimethylamine	0.503	0.554	-10.1	98	0.00
5 S	2-Fluorophenol	1.028	1.100	-7.0	93	0.00
6	Aniline	1.662	1.698	-2.2	89	0.00
7 S	Phenol-d6	1.300	1.322	-1.7	90	0.00
8	2-Chlorophenol	1.199	1.206	-0.6	88	0.00
9	Benzaldehyde	0.801	0.804	-0.4	85	0.00
10 C	Phenol	1.517	1.540	-1.5	88	0.00
11	bis(2-Chloroethyl)ether	1.212	1.168	3.6	86	0.00
12	1,3-Dichlorobenzene	1.282	1.313	-2.4	88	0.00
13 C	1,4-Dichlorobenzene	1.321	1.365	-3.3	86	0.00
14	1,2-Dichlorobenzene	1.189	1.234	-3.8	83	0.00
15	Benzyl Alcohol	0.923	0.989	-7.2	86	0.00
16	2,2'-oxybis(1-Chloropropane	1.910	2.043	-7.0	86	0.00
17	2-Methylphenol	1.012	1.078	-6.5	85	0.00
18	Hexachloroethane	0.503	0.552	-9.7	90	0.00
19 P	n-Nitroso-di-n-propylamine	0.852	0.930	-9.2	89	0.00
20	3+4-Methylphenols	1.315	1.520	-15.6	90	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	82	0.00
22	Acetophenone	0.410	0.471	-14.9	91	0.00
23 S	Nitrobenzene-d5	0.292	0.329	-12.7	87	0.00
24	Nitrobenzene	0.314	0.340	-8.3	84	0.00
25	Isophorone	0.608	0.626	-3.0	83	0.00
26 C	2-Nitrophenol	0.164	0.171	-4.3	81	0.00
27	2,4-Dimethylphenol	0.239	0.244	-2.1	80	0.00
28	bis(2-Chloroethoxy)methane	0.382	0.380	0.5	79	0.00
29 C	2,4-Dichlorophenol	0.246	0.254	-3.3	82	0.00
30	1,2,4-Trichlorobenzene	0.272	0.277	-1.8	82	0.00
31	Naphthalene	0.844	0.876	-3.8	83	0.00
32	Benzoic acid	0.157	0.172	-9.6	82	0.00
33	4-Chloroaniline	0.385	0.379	1.6	78	0.00
34 C	Hexachlorobutadiene	0.157	0.157	0.0	81	0.00
35	Caprolactam	0.080	0.079	1.3	78	0.00
36 C	4-Chloro-3-methylphenol	0.266	0.258	3.0	79	0.00
37	2-Methylnaphthalene	0.578	0.593	-2.6	84	0.00
38	1-Methylnaphthalene	0.560	0.552	1.4	81	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	0.504	0.453	10.1	79	0.00
41 P	Hexachlorocyclopentadiene	0.178	0.157	11.8	73	0.00
42 S	2,4,6-Tribromophenol	0.163	0.163	0.0	91	0.00
43 C	2,4,6-Trichlorophenol	0.339	0.334	1.5	84	0.00
44	2,4,5-Trichlorophenol	0.389	0.375	3.6	84	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051520\
 Data File : BF120314.D
 Acq On : 15 May 2020 11:47
 Operator : MA/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 15 12:33:17 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 15 12:31:31 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	0.973	1.001	-2.9	87	0.00
46	1,1'-Biphenyl	1.310	1.229	6.2	82	0.00
47	2-Chloronaphthalene	1.051	0.966	8.1	80	0.00
48	2-Nitroaniline	0.384	0.347	9.6	81	0.00
49	Acenaphthylene	1.606	1.484	7.6	83	0.00
50	Dimethylphthalate	1.264	1.114	11.9	81	0.00
51	2,6-Dinitrotoluene	0.282	0.246	12.8	77	0.00
52 C	Acenaphthene	0.962	0.961	0.1	93	0.00
53	3-Nitroaniline	0.336	0.343	-2.1	96	0.00
54 P	2,4-Dinitrophenol	0.133	0.137	-3.0	98	0.00
55	Dibenzofuran	1.386	1.337	3.5	90	0.00
56 P	4-Nitrophenol	0.181	0.173	4.4	87	0.00
57	2,4-Dinitrotoluene	0.331	0.320	3.3	90	0.00
58	Fluorene	1.056	1.026	2.8	88	0.00
59	2,3,4,6-Tetrachlorophenol	0.293	0.274	6.5	86	0.00
60	Diethylphthalate	1.221	1.132	7.3	86	0.00
61	4-Chlorophenyl-phenylether	0.549	0.525	4.4	90	0.00
62	4-Nitroaniline	0.313	0.308	1.6	90	0.00
63	Azobenzene	1.154	1.075	6.8	84	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	84	0.00
65	4,6-Dinitro-2-methylphenol	0.101	0.117	-15.8	90	0.00
66 c	n-Nitrosodiphenylamine	0.553	0.589	-6.5	86	0.00
67	4-Bromophenyl-phenylether	0.194	0.228	-17.5	95	0.00
68	Hexachlorobenzene	0.203	0.236	-16.3	97	0.00
69	Atrazine	0.166	0.184	-10.8	91	0.00
70 C	Pentachlorophenol	0.086	0.092	-7.0	82	0.00
71	Phenanthrene	0.840	0.945	-12.5	93	0.00
72	Anthracene	0.900	1.024	-13.8	93	0.00
73	Carbazole	0.841	0.923	-9.8	89	0.00
74	Di-n-butylphthalate	1.013	1.143	-12.8	90	0.00
75 C	Fluoranthene	0.901	1.028	-14.1	91	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
77	Benzidine	0.584	0.646	-10.6	110	0.00
78	Pyrene	1.426	1.399	1.9	98	0.00
79 S	Terphenyl-d14	0.905	0.863	4.6	98	0.00
80	Butylbenzylphthalate	0.671	0.667	0.6	99	0.00
81	Benzo(a)anthracene	1.055	1.037	1.7	100	0.00
82	3,3'-Dichlorobenzidine	0.398	0.440	-10.6	105	0.00
83	Chrysene	1.133	1.157	-2.1	101	0.00
84	Bis(2-ethylhexyl)phthalate	0.835	0.798	4.4	97	0.00
85 c	Di-n-octyl phthalate	1.503	1.530	-1.8	100	0.00
86	Indeno(1,2,3-cd)pyrene	1.028	0.943	8.3	94	0.00
87 I	Perylene-d12	1.000	1.000	0.0	102	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF051520\
Data File : BF120314.D
Acq On : 15 May 2020 11:47
Operator : MA/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 15 12:33:17 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 15 12:31:31 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.092	1.159	-6.1	101	0.00
89	Benzo(k)fluoranthene	0.946	0.948	-0.2	103	0.00
90 C	Benzo(a)pyrene	0.980	0.992	-1.2	101	0.00
91	Dibenzo(a,h)anthracene	0.868	0.803	7.5	95	0.00
92	Benzo(q,h,i)perylene	0.875	0.766	12.5	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0