

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051521\
 Data File : BF123949.D
 Acq On : 15 May 2021 17:20
 Operator : JU/CG
 Sample : M2392-14
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-05-COMP

Manual Integrations
 APPROVED

mohammad
 5/17/2021 11:30:16 AM

Quant Time: May 16 05:20:28 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF051421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 14 19:05:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.904	152	39031	20.00	ng	0.00
21) Naphthalene-d8	8.186	136	143752	20.00	ng	0.00
39) Acenaphthene-d10	9.939	164	75507	20.00	ng	0.00
64) Phenanthrene-d10	11.427	188	111309	20.00	ng	0.00
76) Chrysene-d12	14.074	240	95245	20.00	ng	0.00
86) Perylene-d12	15.568	264	68426	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.510	112	176805	63.60	ng	0.00
7) Phenol-d6	6.516	99	225889	62.64	ng	0.00
23) Nitrobenzene-d5	7.463	82	172939	51.83	ng	0.00
42) 2,4,6-Tribromophenol	10.727	330	53195	58.66	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	269282	43.89	ng	0.00
79) Terphenyl-d14	13.015	244	205716	32.71	ng	0.00
Target Compounds						Qvalue
50) Dimethylphthalate	9.651	163	39363	6.53	ng	# 99
52) Acenaphthene	9.974	154	23184	4.34	ng	96
58) Fluorene	10.486	166	18126	3.13	ng	89
71) Phenanthrene	11.451	178	285511	40.47	ng	99
72) Anthracene	11.504	178	71858	10.05	ng	97
73) Carbazole	11.651	167	29977	4.79	ng	98
75) Fluoranthene	12.645	202	501630	68.67	ng	98
78) Pyrene	12.874	202	459880	54.39	ng	99
81) Benzo(a)anthracene	14.062	228	277710	40.32	ng	96
83) Chrysene	14.104	228	235750	35.58	ng	98
87) Indeno(1,2,3-cd)pyrene	17.068	276	90469	17.12	ng	97
88) Benzo(b)fluoranthene	15.133	252	296612m	56.04	ng	
89) Benzo(k)fluoranthene	15.156	252	72778m	15.09	ng	
90) Benzo(a)pyrene	15.503	252	187017	40.87	ng	99
91) Dibenzo(a,h)anthracene	17.068	278	21700m	4.88	ng	
92) Benzo(g,h,i)perylene	17.521	276	81845	18.34	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051521\
Data File : BF123949.D
Acq On : 15 May 2021 17:20
Operator : JU/CG
Sample : M2392-14
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampled :
SB-05-COMP

Manual Integrations
APPROVED

mohammad
5/17/2021 11:30:16 AM

Quant Time: May 16 05:20:28 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF051421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 14 19:05:57 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.904 min Scan# 819
Delta R.T. 0.000 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20

Instrument :
BNA_F
ClientSampleId :
SB-05-COMP

Tgt Ion:152 Resp: 39031
Ion Ratio Lower Upper
152 100
150 149.6 127.0 190.6
115 60.4 54.3 81.5




Manual Integrations
APPROVED


mohammad
5/17/2021 11:30:16 AM



#5
2-Fluorophenol
Concen: 63.60 ng
RT: 5.510 min Scan# 582
Delta R.T. -0.000 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20

Tgt Ion:112 Resp: 176805
Ion Ratio Lower Upper
112 100
64 79.4 61.4 92.0
63 43.3 33.0 49.4









#7
Phenol-d6
Concen: 62.64 ng
RT: 6.516 min Scan# 753
Delta R.T. -0.006 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20

Tgt Ion: 99 Resp: 225889
Ion Ratio Lower Upper
99 100
42 27.2 19.8 29.8
71 43.6 31.1 46.7

Instrument :
BNA_F
ClientSampleId :
SB-05-COMP




Manual Integrations
APPROVED


mohammad
5/17/2021 11:30:16 AM



#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.186 min Scan# 1037
Delta R.T. 0.000 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20

Tgt Ion: 136 Resp: 143752
Ion Ratio Lower Upper
136 100
137 11.0 8.7 13.1
54 12.5 9.8 14.6
68 5.3 4.1 6.1





#23

Nitrobenzene-d5

Concen: 51.83 ng

RT: 7.463 min Scan# 914

Delta R.T. -0.000 min

Lab File: BF123949.D

Acq: 15 May 2021 17:20

Tgt Ion: 82 Resp: 172939

Ion	Ratio	Lower	Upper
82	100		
128	39.0	29.4	44.2
54	67.9	54.1	81.1

Instrument :





BNA_F

ClientSampleId :

SB-05-COMP

Manual Integrations
APPROVED

mohammad
5/17/2021 11:30:16 AM



#39

Acenaphthene-d10

Concen: 20.00 ng

RT: 9.939 min Scan# 1335




Delta R.T. -0.006 min


Lab File: BF123949.D

Acq: 15 May 2021 17:20

Tgt Ion:164 Resp: 75507


Ion	Ratio	Lower	Upper
164	100		
162	104.8	80.7	121.1
160	45.5	37.3	55.9





#42
2,4,6-Tribromophenol
Concen: 58.66 ng
RT: 10.727 min Scan# 1469
Delta R.T. -0.000 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20




Instrument :
BNA_F
ClientSampleId :
SB-05-COMP




Tgt Ion: 330 Resp: 53195
Ion Ratio Lower Upper
330 100
332 97.0 77.1 115.7
141 37.3 27.9 41.9

Manual Integrations
APPROVED



mohammad
5/17/2021 11:30:16 AM




#45
2-Fluorobiphenyl
Concen: 43.89 ng
RT: 9.263 min Scan# 1220
Delta R.T. -0.000 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20



Tgt Ion: 172 Resp: 269282
Ion Ratio Lower Upper
172 100
171 35.8 28.1 42.1
170 23.2 19.2 28.8









#50
Dimethylphthalate
Concen: 6.53 ng
RT: 9.651 min Scan# 1286
Delta R.T. -0.012 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20

Instrument :
BNA_F
ClientSampled :
SB-05-COMP

Tgt Ion:163 Resp: 39363
Ion Ratio Lower Upper
163 100
194 3.1 3.1 4.7#
164 10.1 8.2 12.2




Manual Integrations
APPROVED


mohammad
5/17/2021 11:30:16 AM



#52
Acenaphthene
Concen: 4.34 ng
RT: 9.974 min Scan# 1341
Delta R.T. -0.000 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20

Tgt Ion:154 Resp: 23184
Ion Ratio Lower Upper
154 100
153 110.8 88.1 132.1
152 46.6 43.4 65.0









#58
Fluorene
Concen: 3.13 ng
RT: 10.486 min Scan# 1428
Delta R.T. -0.006 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20

Tgt Ion	Ratio	Lower	Upper
166	100		
165	102.8	73.5	110.3
167	15.0	10.7	16.1

Instrument :
BNA_F
ClientSampleId :
SB-05-COMP




Manual Integrations
APPROVED


mohammad
5/17/2021 11:30:16 AM



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.427 min Scan# 1588
Delta R.T. 0.000 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20


Tgt Ion	Ratio	Lower	Upper
188	100		
94	8.1	6.9	10.3
80	10.4	8.2	12.2





#71
Phenanthrene
Concen: 40.47 ng
RT: 11.451 min Scan# 1592
Delta R.T. -0.000 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20




Instrument :
BNA_F
ClientSampleId :
SB-05-COMP




Tgt Ion:178 Resp: 285511
Ion Ratio Lower Upper
178 100
176 19.8 15.6 23.4
179 15.3 12.5 18.7

Manual Integrations
APPROVED



mohammad
5/17/2021 11:30:16 AM




#72
Anthracene
Concen: 10.05 ng
RT: 11.504 min Scan# 1601
Delta R.T. -0.000 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20



Tgt Ion:178 Resp: 71858
Ion Ratio Lower Upper
178 100
176 19.6 15.4 23.0
179 13.7 12.6 18.8









#73
 Carbazole
 Concen: 4.79 ng
 RT: 11.651 min Scan# 1626
 Delta R.T. -0.006 min
 Lab File: BF123949.D
 Acq: 15 May 2021 17:20

Instrument :
 BNA_F
 ClientSampled :
 SB-05-COMP

Tgt Ion:167	Resp:	29977
Ion Ratio	Lower	Upper
167	100	
166	20.4	17.0 25.6
139	14.3	10.4 15.6




Manual Integrations
 APPROVED


mohammad
 5/17/2021 11:30:16 AM




#75
 Fluoranthene
 Concen: 68.67 ng
 RT: 12.645 min Scan# 1795
 Delta R.T. 0.006 min
 Lab File: BF123949.D
 Acq: 15 May 2021 17:20

Tgt Ion:202	Resp:	501630
Ion Ratio	Lower	Upper
202	100	
101	11.7	0.0 30.9
203	16.9	0.0 37.5









#76
Chrysene-d12
Concen: 20.00 ng
RT: 14.074 min Scan# 2038
Delta R.T. 0.006 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20





Tgt Ion:240 Resp: 95245
Ion Ratio Lower Upper
240 100
120 10.6 7.2 10.8
236 22.5 19.0 28.4



#78
Pyrene
Concen: 54.39 ng
RT: 12.874 min Scan# 1834
Delta R.T. 0.006 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20




Tgt Ion:202 Resp: 459880
Ion Ratio Lower Upper
202 100
200 20.3 16.1 24.1
203 17.3 14.8 22.2




Instrument :
BNA_F
ClientSampleId :
SB-05-COMP

Manual Integrations
APPROVED



mohammad
5/17/2021 11:30:16 AM



#79
Terphenyl-d14
Concen: 32.71 ng
RT: 13.015 min Scan# 1858
Delta R.T. -0.000 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20




Tgt Ion:244 Resp: 205716
Ion Ratio Lower Upper
244 100
212 7.3 5.7 8.5
122 9.9 9.7 14.5




Instrument :
BNA_F
ClientSampleId :
SB-05-COMP

Manual Integrations
APPROVED



mohammad
5/17/2021 11:30:16 AM




#81
Benzo(a)anthracene
Concen: 40.32 ng
RT: 14.062 min Scan# 2036
Delta R.T. 0.006 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20



Tgt Ion:228 Resp: 277710
Ion Ratio Lower Upper
228 100
226 25.9 22.4 33.6
229 21.5 16.0 24.0









#83
Chrysene
Concen: 35.58 ng
RT: 14.104 min Scan# 2043
Delta R.T. 0.006 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20

Tgt Ion:228	Resp:	235750
Ion Ratio	Lower	Upper
228	100	
226	30.1	24.3 36.5
229	21.7	16.2 24.2

Instrument :
BNA_F
ClientSampleId :
SB-05-COMP




Manual Integrations
APPROVED


mohammad
5/17/2021 11:30:16 AM



#86
Perylene-d12
Concen: 20.00 ng
RT: 15.568 min Scan# 2292
Delta R.T. 0.018 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20

Tgt Ion:264	Resp:	68426
Ion Ratio	Lower	Upper
264	100	
260	22.4	18.7 28.1
265	23.2	17.8 26.8









#87
Indeno(1,2,3-cd)pyrene
Concen: 17.12 ng
RT: 17.068 min Scan# 2547
Delta R.T. 0.012 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20

Tgt Ion:	276	Resp:	90469
Ion Ratio	Lower	Upper	
276	100		
138	22.9	16.2	24.2
277	25.5	20.1	30.1

Instrument :
BNA_F
ClientSampleId :
SB-05-COMP




Manual Integrations
APPROVED


mohammad
5/17/2021 11:30:16 AM



#88
Benzo(b)fluoranthene
Concen: 56.04 ng m
RT: 15.133 min Scan# 2218
Delta R.T. 0.018 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20

Tgt Ion:	252	Resp:	296612
Ion Ratio	Lower	Upper	
252	100		
253	20.5	18.1	27.1
125	10.8	6.2	9.4#









#89
Benzo(k)fluoranthene
Concen: 15.09 ng m
RT: 15.156 min Scan# 2222
Delta R.T. 0.006 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20

Tgt Ion:252 Resp: 72778
Ion Ratio Lower Upper
252 100
253 22.9 17.9 26.9
125 8.8 7.0 10.6

Instrument :
BNA_F
ClientSampleId :
SB-05-COMP




Manual Integrations
APPROVED


mohammad
5/17/2021 11:30:16 AM



#90
Benzo(a)pyrene
Concen: 40.87 ng
RT: 15.503 min Scan# 2281
Delta R.T. 0.012 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20


Tgt Ion:252 Resp: 187017
Ion Ratio Lower Upper
252 100
253 23.4 18.4 27.6
125 11.0 8.0 12.0





#91
Dibenzo(a,h)anthracene
Concen: 4.88 ng m
RT: 17.068 min Scan# 2547
Delta R.T. -0.006 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20




Instrument :
BNA_F
ClientSampled :
SB-05-COMP




Tgt Ion:278 Resp: 21700
Ion Ratio Lower Upper
278 100
139 0.0 10.2 15.2#
279 24.6 19.4 29.0

Manual Integrations
APPROVED

mohammad
5/17/2021 11:30:16 AM



#92
Benzo(g,h,i)perylene
Concen: 18.34 ng
RT: 17.521 min Scan# 2624
Delta R.T. 0.012 min
Lab File: BF123949.D
Acq: 15 May 2021 17:20



Tgt Ion:276 Resp: 81845
Ion Ratio Lower Upper
276 100
277 22.1 18.2 27.2
138 22.4 13.8 20.6#

