

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051524\
 Data File : BF137921.D
 Acq On : 15 May 2024 10:28
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 15 10:49:16 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF051424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 15 04:12:16 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.510	0.508	0.4	98	0.00
3	Pyridine	1.144	1.216	-6.3	103	0.00
4	n-Nitrosodimethylamine	0.505	0.505	0.0	100	0.00
5 S	2-Fluorophenol	1.177	1.157	1.7	98	0.00
6	Aniline	1.297	1.379	-6.3	105	0.00
7 S	Phenol-d6	1.450	1.418	2.2	98	0.00
8	2-Chlorophenol	1.274	1.253	1.6	98	0.00
9	Benzaldehyde	0.785	0.777	1.0	97	0.00
10 C	Phenol	1.464	1.442	1.5	98	0.00
11	bis(2-Chloroethyl)ether	1.124	1.091	2.9	97	0.00
12	1,3-Dichlorobenzene	1.474	1.422	3.5	96	0.00
13 C	1,4-Dichlorobenzene	1.480	1.443	2.5	97	0.00
14	1,2-Dichlorobenzene	1.393	1.349	3.2	97	0.00
15	Benzyl Alcohol	0.968	1.006	-3.9	97	0.00
16	2,2'-oxybis(1-Chloropropane	1.382	1.355	2.0	97	0.00
17	2-Methylphenol	1.016	0.995	2.1	97	0.00
18	Hexachloroethane	0.501	0.486	3.0	95	0.00
19 P	n-Nitroso-di-n-propylamine	0.815	0.797	2.2	97	0.00
20	3+4-Methylphenols	1.331	1.309	1.7	98	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
22	Acetophenone	0.448	0.435	2.9	98	0.00
23 S	Nitrobenzene-d5	0.340	0.329	3.2	97	0.00
24	Nitrobenzene	0.345	0.336	2.6	97	0.00
25	Isophorone	0.590	0.575	2.5	97	0.00
26 C	2-Nitrophenol	0.173	0.171	1.2	95	0.00
27	2,4-Dimethylphenol	0.223	0.220	1.3	98	0.00
28	bis(2-Chloroethoxy)methane	0.357	0.343	3.9	97	0.00
29 C	2,4-Dichlorophenol	0.280	0.274	2.1	95	0.00
30	1,2,4-Trichlorobenzene	0.313	0.303	3.2	97	0.00
31	Naphthalene	1.006	0.963	4.3	96	0.00
32	Benzoic acid	0.189	0.185	2.1	95	0.00
33	4-Chloroaniline	0.330	0.335	-1.5	99	0.00
34 C	Hexachlorobutadiene	0.198	0.193	2.5	97	0.00
35	Caprolactam	0.081	0.085	-4.9	100	0.00
36 C	4-Chloro-3-methylphenol	0.285	0.283	0.7	98	0.00
37	2-Methylnaphthalene	0.646	0.625	3.3	97	0.00
38	1-Methylnaphthalene	0.635	0.610	3.9	97	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
40	1,2,4,5-Tetrachlorobenzene	0.544	0.517	5.0	97	0.00
41 P	Hexachlorocyclopentadiene	0.207	0.205	1.0	93	0.00
42 S	2,4,6-Tribromophenol	0.202	0.206	-2.0	97	0.00
43 C	2,4,6-Trichlorophenol	0.357	0.349	2.2	97	0.00
44	2,4,5-Trichlorophenol	0.390	0.388	0.5	97	0.00
45 S	2-Fluorobiphenyl	1.298	1.216	6.3	97	0.00
46	1,1'-Biphenyl	1.403	1.329	5.3	97	0.00
47	2-Chloronaphthalene	1.120	1.074	4.1	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051524\
 Data File : BF137921.D
 Acq On : 15 May 2024 10:28
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 15 10:49:16 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF051424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 15 04:12:16 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.287	0.288	-0.3	96	0.00
49	Acenaphthylene	1.610	1.562	3.0	96	0.00
50	Dimethylphthalate	1.254	1.218	2.9	96	0.00
51	2,6-Dinitrotoluene	0.288	0.286	0.7	96	0.00
52 C	Acenaphthene	1.064	1.040	2.3	98	0.00
53	3-Nitroaniline	0.281	0.290	-3.2	99	0.00
54 P	2,4-Dinitrophenol	0.129	0.137	-6.2	96	0.00
55	Dibenzofuran	1.549	1.498	3.3	97	0.00
56 P	4-Nitrophenol	0.202	0.224	-10.9	100	0.00
57	2,4-Dinitrotoluene	0.372	0.390	-4.8	98	0.00
58	Fluorene	1.225	1.183	3.4	96	0.00
59	2,3,4,6-Tetrachlorophenol	0.306	0.308	-0.7	96	0.00
60	Diethylphthalate	1.171	1.164	0.6	97	0.00
61	4-Chlorophenyl-phenylether	0.604	0.592	2.0	97	0.00
62	4-Nitroaniline	0.267	0.283	-6.0	98	0.00
63	Azobenzene	1.065	1.043	2.1	96	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	0.115	0.117	-1.7	96	0.00
66 c	n-Nitrosodiphenylamine	0.613	0.573	6.5	95	0.00
67	4-Bromophenyl-phenylether	0.219	0.208	5.0	97	0.00
68	Hexachlorobenzene	0.242	0.228	5.8	97	0.00
69	Atrazine	0.167	0.164	1.8	95	0.00
70 C	Pentachlorophenol	0.142	0.149	-4.9	96	0.00
71	Phenanthrene	0.986	0.929	5.8	97	0.00
72	Anthracene	0.978	0.944	3.5	98	0.00
73	Carbazole	0.893	0.887	0.7	100	0.00
74	Di-n-butylphthalate	0.991	0.991	0.0	97	0.00
75 C	Fluoranthene	1.000	0.998	0.2	98	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
77	Benzidine	0.261	0.400	-53.3#	105	0.00
78	Pyrene	1.749	1.682	3.8	100	0.00
79 S	Terphenyl-d14	1.357	1.288	5.1	99	0.00
80	Butylbenzylphthalate	0.528	0.557	-5.5	100	0.00
81	Benzo(a)anthracene	1.265	1.244	1.7	99	0.00
82	3,3'-Dichlorobenzidine	0.351	0.355	-1.1	100	0.00
83	Chrysene	1.210	1.162	4.0	98	0.00
84	Bis(2-ethylhexyl)phthalate	0.663	0.700	-5.6	99	0.00
85 c	Di-n-octyl phthalate	0.896	0.931	-3.9	98	0.00
86 I	Perylene-d12	1.000	1.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	1.285	1.226	4.6	102	0.00
88	Benzo(b)fluoranthene	1.158	1.151	0.6	99	0.00
89	Benzo(k)fluoranthene	1.138	1.119	1.7	103	0.00
90 C	Benzo(a)pyrene	0.989	0.977	1.2	102	0.00
91	Dibenzo(a,h)anthracene	1.062	1.004	5.5	101	0.00
92	Benzo(g,h,i)perylene	1.115	1.057	5.2	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051524\
Data File : BF137921.D
Acq On : 15 May 2024 10:28
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 15 10:49:16 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF051424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 15 04:12:16 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0