

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051525\
 Data File : BF142395.D
 Acq On : 15 May 2025 16:43
 Operator : RC/JU
 Sample : Q2023-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 HEATER-PAD

Quant Time: May 15 17:45:18 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 05 18:41:44 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.898	152	156435	20.000	ng	0.00
21) Naphthalene-d8	8.181	136	603955	20.000	ng	0.00
39) Acenaphthene-d10	9.939	164	322047	20.000	ng	0.00
64) Phenanthrene-d10	11.421	188	565149	20.000	ng	-0.01
76) Chrysene-d12	14.062	240	336374	20.000	ng	0.00
86) Perylene-d12	15.551	264	339653	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.510	112	509705	55.621	ng	0.00
7) Phenol-d6	6.516	99	672211	59.641	ng	0.00
23) Nitrobenzene-d5	7.457	82	412681	41.672	ng	-0.01
42) 2,4,6-Tribromophenol	10.722	330	177354	57.040	ng	-0.01
45) 2-Fluorobiphenyl	9.257	172	952346	42.166	ng	0.00
79) Terphenyl-d14	13.010	244	1013694	44.604	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051525\
Data File : BF142395.D
Acq On : 15 May 2025 16:43
Operator : RC/JU
Sample : Q2023-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
HEATER-PAD

Quant Time: May 15 17:45:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 05 18:41:44 2025
Response via : Initial Calibration

