

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051622\
 Data File : BF128271.D
 Acq On : 16 May 2022 16:42
 Operator : CG\JU
 Sample : N2823-02DL2 25X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SC-2DL2

Quant Time: May 17 06:22:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 14 00:51:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	89213	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	327884	20.000	ng	0.00
39) Acenaphthene-d10	9.928	164	166364	20.000	ng	0.00
64) Phenanthrene-d10	11.416	188	251059	20.000	ng	0.00
76) Chrysene-d12	14.063	240	206787	20.000	ng	0.00
86) Perylene-d12	15.557	264	137817	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.510	112	20519	3.667	ng	0.00
7) Phenol-d6	6.522	99	27763	3.751	ng	0.00
23) Nitrobenzene-d5	7.445	82	17918	2.459	ng	-0.02
42) 2,4,6-Tribromophenol	10.722	330	4482	2.349	ng	0.00
45) 2-Fluorobiphenyl	9.239	172	33250	3.127	ng	-0.01
79) Terphenyl-d14	12.998	244	30101	2.731	ng	-0.01
Target Compounds						
84) Bis(2-ethylhexyl)phtha...	14.027	149	297338	35.303	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051622\
Data File : BF128271.D
Acq On : 16 May 2022 16:42
Operator : CG\JU
Sample : N2823-02DL2 25X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SC-2DL2

Quant Time: May 17 06:22:10 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 14 00:51:21 2022
Response via : Initial Calibration

