

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051622\
 Data File : BF128273.D
 Acq On : 16 May 2022 17:43
 Operator : CG\JU
 Sample : N2823-13DL 10X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SC-13DL

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 05/17/2022
 Supervised By : Jagrut Upadhyay 05/17/2022

Quant Time: May 17 06:22:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 14 00:51:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.887	152	86099	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	321861	20.000	ng	0.00
39) Acenaphthene-d10	9.928	164	161759	20.000	ng	0.00
64) Phenanthrene-d10	11.416	188	257333	20.000	ng	0.00
76) Chrysene-d12	14.063	240	220534	20.000	ng	0.00
86) Perylene-d12	15.557	264	146100	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	57854	10.712	ng	0.00
7) Phenol-d6	6.516	99	77939	10.910	ng	-0.01
23) Nitrobenzene-d5	7.445	82	50974	7.126	ng	-0.02
42) 2,4,6-Tribromophenol	10.716	330	13669	7.368	ng	-0.01
45) 2-Fluorobiphenyl	9.239	172	91062	8.808	ng	-0.01
79) Terphenyl-d14	12.998	244	91857	7.816	ng	-0.01
Target Compounds						Qvalue
71) Phenanthrene	11.445	178	916460	71.523	ng	100
75) Fluoranthene	12.633	202	321853	23.168	ng	98
78) Pyrene	12.863	202	92986	5.815	ng	99
88) Benzo(b)fluoranthene	15.116	252	27138m	2.996	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051622\
Data File : BF128273.D
Acq On : 16 May 2022 17:43
Operator : CG\JU
Sample : N2823-13DL 10X
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SC-13DL

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 05/17/2022
Supervised By : Jagrut Upadhyay 05/17/2022

Quant Time: May 17 06:22:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF050922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 14 00:51:21 2022
Response via : Initial Calibration

