


Data Path : Z:\HPCHEM1\BNA F\DATA\BF051717\
 Data File : BF095187.D
 Acq On : 17 May 2017 16:08
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu May 18 17:07:53 2017



AutoFind: Scans 2315, 2316, 2317; Background Corrected with Scan 2309

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.4	13755	PASS
68	69	0.00	2	0.8	108	PASS
69	198	0.00	100	32.7	13466	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	46.7	19248	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	41221	PASS
199	198	5	9	6.5	2668	PASS
275	198	10	60	23.0	9478	PASS
365	198	1	100	2.5	1017	PASS
441	442	0.01	24	17.1	6013	PASS
442	198	50	100	85.4	35210	PASS
443	442	15	24	19.9	6990	PASS