

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051815\
 Data File : BF079322.D
 Acq On : 19 May 2015 00:07
 Operator : TP/IZ
 Sample : G2270-04 2X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-17

Quant Time: May 19 11:18:20 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 19 01:26:54 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.16	152	47381	20.00	ng	0.00
21) Naphthalene-d8	8.74	136	195474	20.00	ng	0.00
38) Acenaphthene-d10	10.91	164	95386	20.00	ng	0.00
63) Phenanthrene-d10	12.75	188	185335	20.00	ng	0.00
75) Chrysene-d12	16.03	240	196976	20.00	ng	-0.02
86) Perylene-d12	17.74	264	200007	20.00	ng	-0.11
System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	132607m	48.18	ng	0.00
7) Phenol-d6	6.73	99	178052	48.94	ng	-0.01
23) Nitrobenzene-d5	7.86	82	100243	29.47	ng	0.00
41) 2,4,6-Tribromophenol	11.90	330	38489	37.30	ng	0.00
44) 2-Fluorobiphenyl	10.09	172	225268	32.90	ng	0.00
78) Terphenyl-d14	14.74	244	262154	31.86	ng	-0.01
Target Compounds						Qvalue
49) Dimethylphthalate	10.59	163	14408	2.06	ng	98
70) Phenanthrene	12.78	178	24649	2.38	ng	100
74) Fluoranthene	14.26	202	68767	6.36	ng	92
77) Pyrene	14.54	202	70864	6.24	ng	99
80) Benzo(a)anthracene	16.02	228	38035m	3.53	ng	
82) Chrysene	16.07	228	32841	3.17	ng	95
83) Bis(2-ethylhexyl)phthalate	16.08	149	22394	2.98	ng	99
87) Benzo(b)fluoranthene	17.30	252	49061m	4.48	ng	
89) Benzo(a)pyrene	17.68	252	32677	3.24	ng	# 86
91) Benzo(g,h,i)perylene	19.54	276	24049	2.24	ng	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051815\
Data File : BF079322.D
Acq On : 19 May 2015 00:07
Operator : TP/IZ
Sample : G2270-04 2X
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-17

Quant Time: May 19 11:18:20 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 19 01:26:54 2015
Response via : Initial Calibration

