


Data Path : Z:\HPCHEM1\BNA F\Data\BF051815\
 Data File : BF079310.D
 Acq On : 18 May 2015 10:52
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF043015.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri May 15 01:57:52 2015



AutoFind: Scans 1044, 1045, 1046; Background Corrected with Scan 1041

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.0	8838	PASS
68	69	0.00	2	1.3	107	PASS
69	198	0.00	100	37.0	7963	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	50.1	10806	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	21548	PASS
199	198	5	9	7.0	1519	PASS
275	198	10	60	25.7	5546	PASS
365	198	1	100	3.5	754	PASS
441	442	0.01	24	15.8	2964	PASS
442	198	50	100	87.0	18737	PASS
443	442	15	24	19.8	3714	PASS