

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051816\
 Data File : BF087149.D
 Acq On : 18 May 2016 17:51
 Operator : UM/SJ
 Sample : H3135-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SEWER-JOB

Quant Time: May 19 04:09:47 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF051716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 17 16:55:01 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.58	152	132572	20.00	ng	-0.02
21) Naphthalene-d8	7.86	136	545896	20.00	ng	-0.03
38) Acenaphthene-d10	9.61	164	265381	20.00	ng	-0.03
63) Phenanthrene-d10	11.08	188	477590	20.00	ng	-0.03
75) Chrysene-d12	13.71	240	383567	20.00	ng	-0.05
86) Perylene-d12	15.06	264	277576	20.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.18	112	1150252	145.84	ng	-0.02
7) Phenol-d6	6.25	99	1388974	131.30	ng	-0.03
23) Nitrobenzene-d5	7.15	82	923469	84.31	ng	-0.03
41) 2,4,6-Tribromophenol	10.41	330	420214	143.29	ng	-0.03
44) 2-Fluorobiphenyl	8.95	172	1508069	94.11	ng	-0.03
78) Terphenyl-d14	12.67	244	1304043	76.91	ng	-0.03
Target Compounds						
49) Dimethylphthalate	9.35	163	290499	14.35	ng	Qvalue 100
74) Fluoranthene	12.30	202	52557	2.03	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051816\
Data File : BF087149.D
Acq On : 18 May 2016 17:51
Operator : UM/SJ
Sample : H3135-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SEWER-JOB

Quant Time: May 19 04:09:47 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF051716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 17 16:55:01 2016
Response via : Initial Calibration

