


Data Path : Z:\HPCHEM1\BNA F\Data\BF051816\
 Data File : BF087143.D
 Acq On : 18 May 2016 13:21
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF051716.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue May 17 16:55:01 2016



AutoFind: Scans 855, 856, 857; Background Corrected with Scan 851

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	55.6	36339	PASS
68	69	0.00	2	2.0	539	PASS
69	198	0.00	100	41.3	26974	PASS
70	69	0.00	2	0.4	102	PASS
127	198	10	80	48.8	31906	PASS
197	198	0.00	2	0.8	506	PASS
198	198	100	100	100.0	65352	PASS
199	198	5	9	6.6	4290	PASS
275	198	10	60	26.2	17105	PASS
365	198	1	100	3.9	2547	PASS
441	442	0.01	24	16.1	9589	PASS
442	198	50	100	91.1	59509	PASS
443	442	15	24	18.8	11216	PASS