

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051817\
 Data File : BF095255.D
 Acq On : 19 May 2017 9:25
 Operator : SJ/MA
 Sample : I3186-08
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-T4-BASE

Manual Integrations
 APPROVED

mohammad
 5/19/2017 3:40:37 PM

Quant Time: May 19 11:41:57 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 18 17:07:53 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	107829	20.00	ng	0.00
21) Naphthalene-d8	9.79	136	422328	20.00	ng	0.00
38) Acenaphthene-d10	12.62	164	166213	20.00	ng	0.00
63) Phenanthrene-d10	15.02	188	272065	20.00	ng	0.00
75) Chrysene-d12	18.69	240	174405	20.00	ng	0.00
86) Perylene-d12	20.37	264	166556	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.74	112	686431	106.13	ng	0.00
7) Phenol-d6	7.31	99	873654	112.58	ng	0.00
23) Nitrobenzene-d5	8.67	82	491704	73.42	ng	0.00
41) 2,4,6-Tribromophenol	13.91	330	151915	111.60	ng	0.00
44) 2-Fluorobiphenyl	11.55	172	844956	73.17	ng	0.00
78) Terphenyl-d14	17.34	244	588654	74.34	ng	-0.01
Target Compounds						
10) Phenol	7.34	94	17525	2.14	ng	Qvalue # 61
49) Dimethylphthalate	12.25	163	217591	15.95	ng	99
70) Phenanthrene	15.07	178	34322	2.20	ng	96
74) Fluoranthene	16.81	202	53098	3.31	ng	97
77) Pyrene	17.11	202	48017	3.68	ng	96
87) Benzo(b)fluoranthene	19.97	252	20872m	2.03	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051817\
Data File : BF095255.D
Acq On : 19 May 2017 9:25
Operator : SJ/MA
Sample : I3186-08
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-T4-BASE

Manual Integrations
APPROVED

mohammad
5/19/2017 3:40:37 PM

Quant Time: May 19 11:41:57 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 18 17:07:53 2017
Response via : Initial Calibration

