

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051817\
 Data File : BF095259.D
 Acq On : 19 May 2017 11:34
 Operator : SJ/MA
 Sample : I3186-14
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-S4-NSW

Manual Integrations
 APPROVED

mohammad
 5/19/2017 3:40:54 PM

Quant Time: May 19 14:57:22 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 18 17:07:53 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	104247	20.00	ng	0.00
21) Naphthalene-d8	9.78	136	399193	20.00	ng	0.00
38) Acenaphthene-d10	12.62	164	192195	20.00	ng	0.00
63) Phenanthrene-d10	15.02	188	311929	20.00	ng	0.00
75) Chrysene-d12	18.69	240	175415	20.00	ng	0.00
86) Perylene-d12	20.36	264	181700	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.74	112	998700	159.72	ng	0.00
7) Phenol-d6	7.32	99	1213962	161.81	ng	0.00
23) Nitrobenzene-d5	8.67	82	648491	102.44	ng	0.00
41) 2,4,6-Tribromophenol	13.91	330	228396	145.10	ng	0.00
44) 2-Fluorobiphenyl	11.56	172	1242899	93.08	ng	0.00
78) Terphenyl-d14	17.34	244	837569	105.17	ng	0.00
Target Compounds						
10) Phenol	7.33	94	29046	3.67	ng	Qvalue # 80
49) Dimethylphthalate	12.25	163	316702	20.07	ng	99
70) Phenanthrene	15.06	178	158292	8.83	ng	98
71) Anthracene	15.15	178	39796m	2.17	ng	
74) Fluoranthene	16.80	202	238036	12.95	ng	99
77) Pyrene	17.11	202	206477	15.74	ng	97
80) Benzo(a)anthracene	18.68	228	62102m	6.08	ng	
82) Chrysene	18.73	228	63380	6.42	ng	97
85) Indeno(1,2,3-cd)pyrene	21.72	276	29406	3.67	ng	96
87) Benzo(b)fluoranthene	19.96	252	67720m	6.03	ng	
88) Benzo(k)fluoranthene	19.99	252	29460m	2.72	ng	
89) Benzo(a)pyrene	20.31	252	56492	5.45	ng	93
91) Benzo(a,h,i)perylene	22.11	276	29259	3.48	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051817\
Data File : BF095259.D
Acq On : 19 May 2017 11:34
Operator : SJ/MA
Sample : I3186-14
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-S4-NSW

Manual Integrations
APPROVED

mohammad
5/19/2017 3:40:54 PM

Quant Time: May 19 14:57:22 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 18 17:07:53 2017
Response via : Initial Calibration

