

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051823\
 Data File : BF133449.D
 Acq On : 18 May 2023 20:45
 Operator : CG\JU
 Sample : 02825-01 5X
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OK-01-051723

Integration Parameters: rteint.p

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M

Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

Signal : TIC: BF133449.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	3.640	259	264	269	rBV2	10999	18388	1.22%	0.159%
2	4.875	471	474	480	rBV	147987	164447	10.90%	1.422%
3	5.322	546	550	565	rBV	652525	665471	44.12%	5.755%
4	6.357	723	726	739	rBV	697728	698704	46.32%	6.043%
5	6.710	783	786	797	rBV	1350996	1220227	80.90%	10.553%
6	7.275	879	882	894	rBV	461093	418372	27.74%	3.618%
7	7.998	1001	1005	1011	rBV	1897031	1508270	100.00%	13.044%
8	9.069	1184	1187	1197	rBV	891110	790769	52.43%	6.839%
9	9.686	1290	1292	1296	rVB2	20535	18748	1.24%	0.162%
10	9.751	1299	1303	1306	rBV	1704995	1430365	94.83%	12.371%
11	10.463	1421	1424	1427	rVB	50897	44906	2.98%	0.388%
12	10.539	1433	1437	1449	rBV	370138	372047	24.67%	3.218%
13	10.663	1455	1458	1465	rVB4	29062	50282	3.33%	0.435%
14	11.139	1536	1539	1543	rVB4	20683	24463	1.62%	0.212%
15	11.233	1551	1555	1558	rBV	1591620	1330264	88.20%	11.505%
16	11.769	1643	1646	1651	rBV7	32338	46454	3.08%	0.402%
17	12.157	1710	1712	1718	rBV7	11563	20072	1.33%	0.174%
18	12.286	1731	1734	1736	rBV2	17947	19893	1.32%	0.172%
19	12.445	1758	1761	1764	rBV	39858	35854	2.38%	0.310%
20	12.675	1797	1800	1802	rBV	53730	47822	3.17%	0.414%
21	12.816	1821	1824	1832	rVB	894947	735801	48.78%	6.364%
22	13.874	2000	2004	2010	rBV	1240474	1163778	77.16%	10.065%
23	15.298	2242	2246	2250	rBV	645279	737251	48.88%	6.376%




Sum of corrected areas: 11562648

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051823\
Data File : BF133449.D
Acq On : 18 May 2023 20:45
Operator : CG\JU
Sample : 02825-01 5X
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OK-01-051723

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051823\
Data File : BF133449.D
Acq On : 18 May 2023 20:45
Operator : CG\JU
Sample : 02825-01 5X
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OK-01-051723










Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

Peak Number 1 2-Pentanone, 4-hydroxy-4-me... Concentration Rank 1

R.T.	EstConc	Area	Relative to ISTD	R.T.
4.875	2.70 ng	164447	1,4-Dichlorobenzene-d4	6.716

Hit#	of	5	Tentative ID	MW	MolForm	CAS#	Qual
1	2-Pentanone, 4-hydroxy-4-methyl-	116	C6H12O2	000123-42-2	42		
2	Hydrazine, 1,1-bis(1-methylethyl)-	116	C6H16N2	000921-14-2	37		
3	3-Octanol	130	C8H18O	000589-98-0	33		
4	Acetic acid, cyano-, 1,1-dimethy...	141	C7H11NO2	001116-98-9	23		
5	1,8-Nonanediol, 8-methyl-	174	C10H22O2	054725-73-4	23		



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF051823\
Data File : BF133449.D
Acq On : 18 May 2023 20:45
Operator : CG\JU
Sample : 02825-01 5X
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OK-01-051723

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc
2-Pentanone, 4-...	4.875	2.7	ng	164447	1	6.716	1220230	20.0