

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051915\
 Data File : BF079362.D
 Acq On : 20 May 2015 3:03
 Operator : TP/IZ
 Sample : G2281-01
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-01

Quant Time: May 20 06:15:53 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 20 05:08:22 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.13	152	45130	20.00	ng	0.00
21) Naphthalene-d8	8.71	136	185912	20.00	ng	0.00
38) Acenaphthene-d10	10.87	164	93416	20.00	ng	-0.01
63) Phenanthrene-d10	12.71	188	181002	20.00	ng	-0.01
75) Chrysene-d12	16.00	240	204689	20.00	ng	-0.02
86) Perylene-d12	17.70	264	223238	20.00	ng	-0.11
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	265102m	101.12	ng	0.00
7) Phenol-d6	6.70	99	348563	100.58	ng	-0.01
23) Nitrobenzene-d5	7.82	82	200411	61.95	ng	-0.01
41) 2,4,6-Tribromophenol	11.85	330	109037	107.89	ng	-0.01
44) 2-Fluorobiphenyl	10.04	172	421366	62.84	ng	-0.01
78) Terphenyl-d14	14.71	244	522510	61.12	ng	-0.01
Target Compounds						Qvalue
31) Naphthalene	8.73	128	57162	6.45	ng	98
37) 2-Methylnaphthalene	9.59	142	13511	2.20	ng	96
48) Acenaphthylene	10.70	152	78186	8.22	ng	98
49) Dimethylphthalate	10.56	163	30822	4.50	ng	# 99
51) Acenaphthene	10.91	154	15435	2.72	ng	99
57) Fluorene	11.55	166	21105	3.14	ng	99
70) Phenanthrene	12.74	178	266983	26.37	ng	99
71) Anthracene	12.81	178	119988	12.08	ng	99
74) Fluoranthene	14.22	202	545034	51.66	ng	98
77) Pyrene	14.50	202	533712	45.25	ng	99
80) Benzo(a)anthracene	15.99	228	362583m	32.35	ng	
82) Chrysene	16.03	228	264555	24.54	ng	98
85) Indeno(1,2,3-cd)pyrene	19.07	276	247709	18.03	ng	95
87) Benzo(b)fluoranthene	17.27	252	407720m	33.37	ng	
88) Benzo(k)fluoranthene	17.28	252	180762m	15.28	ng	
89) Benzo(a)pyrene	17.65	252	345568	30.71	ng	# 98
90) Dibenzo(a,h)anthracene	19.09	278	51060m	4.27	ng	
91) Benzo(g,h,i)perylene	19.47	276	237091	19.78	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051915\
Data File : BF079362.D
Acq On : 20 May 2015 3:03
Operator : TP/IZ
Sample : G2281-01
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-01

Quant Time: May 20 06:15:53 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 20 05:08:22 2015
Response via : Initial Calibration

