

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051915\
 Data File : BF079366.D
 Acq On : 20 May 2015 4:58
 Operator : TP/IZ
 Sample : G2281-05
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-03

Quant Time: May 20 06:53:17 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 20 05:08:22 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.13	152	47193	20.00	ng	0.00
21) Naphthalene-d8	8.71	136	188935	20.00	ng	0.00
38) Acenaphthene-d10	10.88	164	90745	20.00	ng	0.00
63) Phenanthrene-d10	12.72	188	175225	20.00	ng	0.00
75) Chrysene-d12	16.00	240	191024	20.00	ng	-0.02
86) Perylene-d12	17.71	264	218557	20.00	ng	-0.10
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	262482m	95.74	ng	0.00
7) Phenol-d6	6.71	99	354447	97.81	ng	0.00
23) Nitrobenzene-d5	7.83	82	199442	60.67	ng	0.00
41) 2,4,6-Tribromophenol	11.86	330	100897	102.78	ng	0.00
44) 2-Fluorobiphenyl	10.06	172	414320	63.61	ng	0.00
78) Terphenyl-d14	14.71	244	471144	59.05	ng	-0.01
Target Compounds						Qvalue
31) Naphthalene	8.73	128	24213	2.69	ng	99
48) Acenaphthylene	10.70	152	33552	3.63	ng	98
49) Dimethylphthalate	10.56	163	28264	4.24	ng	99
70) Phenanthrene	12.74	178	102789	10.49	ng	99
71) Anthracene	12.81	178	39673	4.13	ng	99
74) Fluoranthene	14.22	202	209792	20.54	ng	99
77) Pyrene	14.50	202	211841	19.24	ng	99
80) Benzo(a)anthracene	15.99	228	137682m	13.16	ng	
82) Chrysene	16.03	228	111496	11.08	ng	98
85) Indeno(1,2,3-cd)pyrene	19.09	276	93836	7.32	ng	95
87) Benzo(b)fluoranthene	17.28	252	173124m	14.47	ng	
88) Benzo(k)fluoranthene	17.30	252	59701m	5.16	ng	
89) Benzo(a)pyrene	17.65	252	129760	11.78	ng	# 97
91) Benzo(a,h,i)perylene	19.50	276	91676	7.81	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051915\
Data File : BF079366.D
Acq On : 20 May 2015 4:58
Operator : TP/IZ
Sample : G2281-05
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampled :
SB-03

Quant Time: May 20 06:53:17 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 20 05:08:22 2015
Response via : Initial Calibration

