

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051915\
 Data File : BF079367.D
 Acq On : 20 May 2015 5:26
 Operator : TP/IZ
 Sample : G2289-01
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-34D

Quant Time: May 20 06:56:32 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 20 05:08:22 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.13	152	47043	20.00	ng	0.00
21) Naphthalene-d8	8.71	136	191810	20.00	ng	0.00
38) Acenaphthene-d10	10.88	164	91221	20.00	ng	0.00
63) Phenanthrene-d10	12.72	188	175741	20.00	ng	0.00
75) Chrysene-d12	16.00	240	196910	20.00	ng	-0.02
86) Perylene-d12	17.69	264	207419	20.00	ng	-0.13
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	386905m	141.57	ng	0.00
7) Phenol-d6	6.71	99	521253	144.30	ng	0.00
23) Nitrobenzene-d5	7.83	82	282766	84.73	ng	0.00
41) 2,4,6-Tribromophenol	11.86	330	144191	146.11	ng	0.00
44) 2-Fluorobiphenyl	10.06	172	379361	57.94	ng	0.00
78) Terphenyl-d14	14.71	244	401434	48.81	ng	-0.01
Target Compounds						Qvalue
31) Naphthalene	8.73	128	20277	2.22	ng	99
49) Dimethylphthalate	10.56	163	41439	6.19	ng	99
51) Acenaphthene	10.91	154	15302	2.76	ng	98
57) Fluorene	11.55	166	16934	2.58	ng	99
70) Phenanthrene	12.74	178	202870	20.63	ng	100
71) Anthracene	12.81	178	47641	4.94	ng	98
74) Fluoranthene	14.22	202	204877	20.00	ng	99
77) Pyrene	14.50	202	188474	16.61	ng	98
80) Benzo(a)anthracene	15.99	228	96944	8.99	ng	97
82) Chrysene	16.03	228	69554	6.71	ng	97
85) Indeno(1,2,3-cd)pyrene	19.06	276	49351	3.73	ng	95
87) Benzo(b)fluoranthene	17.26	252	103305m	9.10	ng	
88) Benzo(k)fluoranthene	17.29	252	31386m	2.86	ng	
89) Benzo(a)pyrene	17.63	252	77575	7.42	ng	98
91) Benzo(a,h,i)perylene	19.46	276	46538	4.18	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF051915\
Data File : BF079367.D
Acq On : 20 May 2015 5:26
Operator : TP/IZ
Sample : G2289-01
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-34D

Quant Time: May 20 06:56:32 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 20 05:08:22 2015
Response via : Initial Calibration

