

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF051917\
 Data File : BF095274.D
 Acq On : 19 May 2017 21:07
 Operator : SJ/MA
 Sample : I3203-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E2-CONCRETE-BATCH-11-BULK

Quant Time: May 20 01:57:20 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 18 17:07:53 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	81412	20.00	ng	0.00
21) Naphthalene-d8	9.78	136	353358	20.00	ng	0.00
38) Acenaphthene-d10	12.61	164	171162	20.00	ng	0.00
63) Phenanthrene-d10	15.02	188	334325	20.00	ng	0.00
75) Chrysene-d12	18.69	240	265135	20.00	ng	0.00
86) Perylene-d12	20.37	264	160335	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.75	112	253496	51.91	ng	0.01
7) Phenol-d6	7.30	99	548225	93.57	ng	-0.01
23) Nitrobenzene-d5	8.67	82	407171	72.66	ng	0.00
41) 2,4,6-Tribromophenol	13.92	330	45795	32.67	ng	0.00
44) 2-Fluorobiphenyl	11.55	172	900253	75.70	ng	0.00
78) Terphenyl-d14	17.34	244	824887	68.53	ng	-0.01
Target Compounds						Qvalue
10) Phenol	7.33	94	92202	14.93	ng	93
17) 2-Methylphenol	8.26	107	157315	34.59	ng	92
20) 3+4-Methylphenols	8.54	107	215015	34.79	ng	# 61
27) 2,4-Dimethylphenol	9.37	122	136878	26.71	ng	93
49) Dimethylphthalate	12.26	163	108611	7.73	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF051917\
Data File : BF095274.D
Acq On : 19 May 2017 21:07
Operator : SJ/MA
Sample : I3203-02
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E2-CONCRETE-BATCH-11-BULK

Quant Time: May 20 01:57:20 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 18 17:07:53 2017
Response via : Initial Calibration

