

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052016\
 Data File : BF087230.D
 Acq On : 20 May 2016 17:59
 Operator : UM/SJ
 Sample : PB90703BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB90703BL

Quant Time: May 21 00:23:31 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF051716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 21 00:19:36 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.71	152	405273	20.00	ng	0.15
21) Naphthalene-d8	7.87	136	637	20.00	ng	0.02
38) Acenaphthene-d10	9.60	164	288	20.00	ng	0.01
63) Phenanthrene-d10	11.08	188	1052	20.00	ng	0.01
75) Chrysene-d12	13.71	240	283	20.00	ng	0.01
86) Perylene-d12	0.00	264	0	0.00	ng	-15.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.15	112	956847	39.68	ng	0.02
7) Phenol-d6	6.23	99	1340531	41.45	ng	0.00
23) Nitrobenzene-d5	7.13	82	787303	61597.94	ng	-0.01
41) 2,4,6-Tribromophenol	10.38	330	330365	103805.50	ng	-0.01
44) 2-Fluorobiphenyl	8.92	172	1531223	88053.79	ng	0.00
78) Terphenyl-d14	12.65	244	1355221	108326.45	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052016\
Data File : BF087230.D
Acq On : 20 May 2016 17:59
Operator : UM/SJ
Sample : PB90703BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB90703BL

Quant Time: May 21 00:23:31 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF051716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 21 00:19:36 2016
Response via : Initial Calibration

