

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052021\
 Data File : BF124012.D
 Acq On : 20 May 2021 16:44
 Operator : JU/CG
 Sample : M2457-02DL 10X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-4DL

Manual Integrations
 APPROVED

Quant Time: May 20 17:07:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF051421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 14 19:05:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.904	152	56426	20.00	ng	0.00
21) Naphthalene-d8	8.192	136	228123	20.00	ng	0.00
39) Acenaphthene-d10	9.945	164	139217	20.00	ng	0.00
64) Phenanthrene-d10	11.433	188	233317	20.00	ng	0.00
76) Chrysene-d12	14.080	240	169735	20.00	ng	0.01
86) Perylene-d12	15.574	264	106714	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.510	112	30527	7.60	ng	0.00
7) Phenol-d6	6.522	99	39080	7.50	ng	0.00
23) Nitrobenzene-d5	7.463	82	31804	6.01	ng	0.00
42) 2,4,6-Tribromophenol	10.733	330	12936	7.74	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	54339	4.80	ng	0.00
79) Terphenyl-d14	13.015	244	48867	4.36	ng	0.00
Target Compounds						
31) Naphthalene	8.210	128	39784	2.95	ng	Qvalue 96
52) Acenaphthene	9.980	154	32879	3.34	ng	96
55) Dibenzofuran	10.151	168	41566	2.94	ng	95
58) Fluorene	10.492	166	39864	3.74	ng	94
71) Phenanthrene	11.463	178	614569	41.56	ng	99
72) Anthracene	11.510	178	158105	10.55	ng	97
73) Carbazole	11.663	167	55927	4.27	ng	# 95
75) Fluoranthene	12.651	202	606606	39.62	ng	98
78) Pyrene	12.880	202	478384	31.75	ng	99
81) Benzo(a)anthracene	14.068	228	229220	18.67	ng	97
83) Chrysene	14.104	228	182448	15.45	ng	98
87) Indeno(1,2,3-cd)pyrene	17.080	276	52993	6.43	ng	# 94
88) Benzo(b)fluoranthene	15.133	252	165512m	20.05	ng	
89) Benzo(k)fluoranthene	15.156	252	53831m	7.16	ng	
90) Benzo(a)pyrene	15.509	252	107041	15.00	ng	97
91) Dibenzo(a,h)anthracene	17.092	278	16378m	2.36	ng	
92) Benzo(g,h,i)perylene	17.539	276	49107	7.06	ng	94

mohammad

5/21/2021 3:21:30 PM


(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052021\
Data File : BF124012.D
Acq On : 20 May 2021 16:44
Operator : JU/CG
Sample : M2457-02DL 10X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-4DL

Manual Integrations
APPROVED

Quant Time: May 20 17:07:58 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF051421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 14 19:05:57 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.904 min Scan# 819
Delta R.T. 0.000 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:152 Resp: 56426
Ion Ratio Lower Upper
152 100
150 156.5 127.0 190.6
115 65.4 54.3 81.5




Instrument :
BNA_F
ClientSampleId :
TP-4DL

Manual Integrations
APPROVED




#5
2-Fluorophenol
Concen: 7.60 ng
RT: 5.510 min Scan# 582
Delta R.T. -0.000 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:112 Resp: 30527
Ion Ratio Lower Upper
112 100
64 77.0 61.4 92.0
63 42.5 33.0 49.4



mohammad

5/21/2021 3:21:30 PM






#7
Phenol-d6
Concen: 7.50 ng
RT: 6.522 min Scan# 754
Delta R.T. -0.000 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:	99	Resp:	39080
Ion	Ratio	Lower	Upper
99	100		
42	23.7	19.8	29.8
71	39.8	31.1	46.7


Instrument :
BNA_F
ClientSampleId :
TP-4DL

Manual Integrations
APPROVED






mohammad


5/21/2021 3:21:30 PM



#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.192 min Scan# 1038
Delta R.T. 0.006 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:	136	Resp:	228123
Ion	Ratio	Lower	Upper
136	100		
137	10.5	8.7	13.1
54	10.9	9.8	14.6
68	5.5	4.1	6.1








#23
Nitrobenzene-d5
Concen: 6.01 ng
RT: 7.463 min Scan# 914
Delta R.T. -0.000 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion: 82 Resp: 31804

Ion	Ratio	Lower	Upper
82	100		
128	38.5	29.4	44.2
54	73.4	54.1	81.1


Instrument :
BNA_F
ClientSampleId :
TP-4DL

Manual Integrations
APPROVED



mohammad




5/21/2021 3:21:30 PM




#31
Naphthalene
Concen: 2.95 ng
RT: 8.210 min Scan# 1041
Delta R.T. -0.000 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44


Tgt Ion: 128 Resp: 39784

Ion	Ratio	Lower	Upper
128	100		
129	10.5	8.9	13.3
127	11.3	10.8	16.2







#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.945 min Scan# 1336
Delta R.T. 0.000 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44



Tgt Ion:164 Resp: 139217
Ion Ratio Lower Upper
164 100
162 102.7 80.7 121.1
160 44.9 37.3 55.9




Instrument :
BNA_F
ClientSampleId :
TP-4DL


Manual Integrations
APPROVED

mohammad



5/21/2021 3:21:30 PM




#42
2,4,6-Tribromophenol
Concen: 7.74 ng
RT: 10.733 min Scan# 1470
Delta R.T. 0.006 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44



Tgt Ion:330 Resp: 12936
Ion Ratio Lower Upper
330 100
332 92.4 77.1 115.7
141 42.0 27.9 41.9#









#45
2-Fluorobiphenyl
Concen: 4.80 ng
RT: 9.263 min Scan# 1220
Delta R.T. -0.000 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:	172	Resp:	54339
Ion	Ratio	Lower	Upper
172	100		
171	38.5	28.1	42.1
170	21.7	19.2	28.8




Instrument :
BNA_F
ClientSampleId :
TP-4DL

Manual Integrations
APPROVED




#52
Acenaphthene
Concen: 3.34 ng
RT: 9.980 min Scan# 1342
Delta R.T. 0.006 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:	154	Resp:	32879
Ion	Ratio	Lower	Upper
154	100		
153	111.1	88.1	132.1
152	45.9	43.4	65.0



mohammad

5/21/2021 3:21:30 PM






#55
Dibenzenofuran
Concen: 2.94 ng
RT: 10.151 min Scan# 1371
Delta R.T. 0.006 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:168 Resp: 41566
Ion Ratio Lower Upper
168 100
139 37.4 33.2 49.8
169 12.6 10.8 16.2


Instrument :
BNA_F
ClientSampleId :
TP-4DL

Manual Integrations
APPROVED






mohammad


5/21/2021 3:21:30 PM



#58
Fluorene
Concen: 3.74 ng
RT: 10.492 min Scan# 1429
Delta R.T. -0.000 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44


Tgt Ion:166 Resp: 39864
Ion Ratio Lower Upper
166 100
165 97.7 73.5 110.3
167 12.4 10.7 16.1







#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.433 min Scan# 1589
Delta R.T. 0.006 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Instrument :
BNA_F
ClientSampleId :
TP-4DL




Tgt Ion:188 Resp: 233317
Ion Ratio Lower Upper
188 100
94 8.5 6.9 10.3
80 11.9 8.2 12.2

Manual Integrations
APPROVED




mohammad



5/21/2021 3:21:30 PM




#71
Phenanthrene
Concen: 41.56 ng
RT: 11.463 min Scan# 1594
Delta R.T. 0.012 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44




Tgt Ion:178 Resp: 614569
Ion Ratio Lower Upper
178 100
176 18.4 15.6 23.4
179 15.5 12.5 18.7







#72
Anthracene
Concen: 10.55 ng
RT: 11.510 min Scan# 1602
Delta R.T. 0.006 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44



Tgt Ion:178 Resp: 158105
Ion Ratio Lower Upper
178 100
176 18.0 15.4 23.0
179 14.6 12.6 18.8




Instrument :
BNA_F
ClientSampleId :
TP-4DL


Manual Integrations
APPROVED

mohammad



5/21/2021 3:21:30 PM




#73
Carbazole
Concen: 4.27 ng
RT: 11.663 min Scan# 1628
Delta R.T. 0.006 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44



Tgt Ion:167 Resp: 55927
Ion Ratio Lower Upper
167 100
166 22.8 17.0 25.6
139 16.4 10.4 15.6#








#75
Fluoranthene
Concen: 39.62 ng
RT: 12.651 min Scan# 1796
Delta R.T. 0.012 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:202	Resp: 606606
Ion Ratio	Lower Upper
202 100	
101 12.1	0.0 30.9
203 17.3	0.0 37.5


Instrument :
BNA_F
ClientSampleId :
TP-4DL

Manual Integrations
APPROVED






mohammad


5/21/2021 3:21:30 PM



#76
Chrysene-d12
Concen: 20.00 ng
RT: 14.080 min Scan# 2039
Delta R.T. 0.012 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:240	Resp: 169735
Ion Ratio	Lower Upper
240 100	
120 7.4	7.2 10.8
236 22.3	19.0 28.4








#78
Pyrene
Concen: 31.75 ng
RT: 12.880 min Scan# 1835
Delta R.T. 0.012 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:202 Resp: 478384
Ion Ratio Lower Upper
202 100
200 20.2 16.1 24.1
203 17.9 14.8 22.2


Instrument :
BNA_F
ClientSampleId :
TP-4DL

Manual Integrations
APPROVED






mohammad


5/21/2021 3:21:30 PM



#79
Terphenyl-d14
Concen: 4.36 ng
RT: 13.015 min Scan# 1858
Delta R.T. -0.000 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:244 Resp: 48867
Ion Ratio Lower Upper
244 100
212 8.4 5.7 8.5
122 9.0 9.7 14.5#









#81
Benzo(a)anthracene
Concen: 18.67 ng
RT: 14.068 min Scan# 2037
Delta R.T. 0.012 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:228 Resp: 229220
Ion Ratio Lower Upper
228 100
226 26.0 22.4 33.6
229 19.7 16.0 24.0




Instrument :
BNA_F
ClientSampleId :
TP-4DL

Manual Integrations
APPROVED




#83
Chrysene
Concen: 15.45 ng
RT: 14.104 min Scan# 2043
Delta R.T. 0.006 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:228 Resp: 182448
Ion Ratio Lower Upper
228 100
226 30.0 24.3 36.5
229 22.5 16.2 24.2



mohammad

5/21/2021 3:21:30 PM






#86
Perylene-d12
Concen: 20.00 ng
RT: 15.574 min Scan# 2293
Delta R.T. 0.023 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:	264	Resp:	106714
Ion Ratio	Lower	Upper	
264	100		
260	20.8	18.7	28.1
265	21.0	17.8	26.8


Instrument :
BNA_F
ClientSampleId :
TP-4DL

Manual Integrations
APPROVED






mohammad


5/21/2021 3:21:30 PM



#87
Indeno(1,2,3-cd)pyrene
Concen: 6.43 ng
RT: 17.080 min Scan# 2549
Delta R.T. 0.023 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:	276	Resp:	52993
Ion Ratio	Lower	Upper	
276	100		
138	24.5	16.2	24.2#
277	26.4	20.1	30.1








#88
Benzo(b)fluoranthene
Concen: 20.05 ng m
RT: 15.133 min Scan# 2218
Delta R.T. 0.018 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:252 Resp: 165512
Ion Ratio Lower Upper
252 100
253 20.8 18.1 27.1
125 11.1 6.2 9.4#


Instrument :
BNA_F
ClientSampleId :
TP-4DL

Manual Integrations
APPROVED






mohammad


5/21/2021 3:21:30 PM



#89
Benzo(k)fluoranthene
Concen: 7.16 ng m
RT: 15.156 min Scan# 2222
Delta R.T. 0.006 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:252 Resp: 53831
Ion Ratio Lower Upper
252 100
253 26.7 17.9 26.9
125 11.8 7.0 10.6#








#90
Benzo(a)pyrene
Concen: 15.00 ng
RT: 15.509 min Scan# 2282
Delta R.T. 0.018 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:252 Resp: 107041
Ion Ratio Lower Upper
252 100
253 24.0 18.4 27.6
125 11.5 8.0 12.0


Instrument :
BNA_F
ClientSampleId :
TP-4DL

Manual Integrations
APPROVED






mohammad


5/21/2021 3:21:30 PM




#91
Dibenzo(a,h)anthracene
Concen: 2.36 ng m
RT: 17.092 min Scan# 2551
Delta R.T. 0.018 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44

Tgt Ion:278 Resp: 16378
Ion Ratio Lower Upper
278 100
139 18.6 10.2 15.2#
279 20.2 19.4 29.0







#92
Benzo(g,h,i)perylene
Concen: 7.06 ng
RT: 17.539 min Scan# 2627
Delta R.T. 0.029 min
Lab File: BF124012.D
Acq: 20 May 2021 16:44



Tgt Ion:	276	Resp:	49107
Ion Ratio	Lower	Upper	
276	100		
277	19.6	18.2	27.2
138	19.4	13.8	20.6



Instrument :
BNA_F
ClientSampleId :
TP-4DL

Manual Integrations
APPROVED

mohammad

5/21/2021 3:21:30 PM