


Data Path : Z:\HPCHEM1\BNA F\Data\BF052115\
 Data File : BF079421.D
 Acq On : 22 May 2015 1:33
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF043015.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 20 05:08:22 2015



AutoFind: Scans 1036, 1037, 1038; Background Corrected with Scan 1032

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.5	11093	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	36.2	10448	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	50.7	14628	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	28838	PASS
199	198	5	9	6.3	1829	PASS
275	198	10	60	25.2	7268	PASS
365	198	1	100	3.1	886	PASS
441	442	0.01	24	16.2	4361	PASS
442	198	50	100	93.1	26847	PASS
443	442	15	24	19.8	5326	PASS