

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF052120\  
 Data File : BF120386.D  
 Acq On : 21 May 2020 17:36  
 Operator : MA/SJ  
 Sample : L2722-01  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 COMP-1

Quant Time: May 22 01:40:02 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051420.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu May 21 14:56:56 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.59  | 152  | 236782   | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 7.87  | 136  | 951962   | 20.00  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 9.62  | 164  | 425329   | 20.00  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 11.11 | 188  | 867373   | 20.00  | ng    | 0.00     |
| 76) Chrysene-d12            | 13.74 | 240  | 639589   | 20.00  | ng    | 0.00     |
| 87) Perylene-d12            | 15.13 | 264  | 558174   | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.20  | 112  | 1796282  | 147.61 | ng    | 0.00     |
| 7) Phenol-d6                | 6.25  | 99   | 2486304  | 161.56 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.16  | 82   | 1430927  | 102.91 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 10.42 | 330  | 576709   | 166.30 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 8.96  | 172  | 2772570  | 133.92 | ng    | 0.00     |
| 79) Terphenyl-d14           | 12.70 | 244  | 2661661  | 91.94  | ng    | 0.00     |
| Target Compounds            |       |      |          |        |       |          |
| 10) Phenol                  | 6.26  | 94   | 41676    | 2.320  | ng    | 88       |
| 50) Dimethylphthalate       | 9.35  | 163  | 615188   | 22.881 | ng    | 99       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF052120\  
Data File : BF120386.D  
Acq On : 21 May 2020 17:36  
Operator : MA/SJ  
Sample : L2722-01  
Misc :  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
COMP-1

Quant Time: May 22 01:40:02 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF051420.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu May 21 14:56:56 2020  
Response via : Initial Calibration

