

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052121\
 Data File : BF124023.D
 Acq On : 21 May 2021 17:46
 Operator : JU/CG
 Sample : M2366-08
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 JG-SPOILS-5-11-21-B

Quant Time: May 21 18:20:35 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF051421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 14 19:05:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.892	152	47151	20.00	ng	-0.01
21) Naphthalene-d8	8.169	136	187900	20.00	ng	-0.02
39) Acenaphthene-d10	9.927	164	109981	20.00	ng	-0.02
64) Phenanthrene-d10	11.416	188	216575	20.00	ng	-0.01
76) Chrysene-d12	14.057	240	191054	20.00	ng	-0.01
86) Perylene-d12	15.545	264	158938	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	373132	111.11	ng	0.00
7) Phenol-d6	6.516	99	410874	94.31	ng	0.00
23) Nitrobenzene-d5	7.451	82	400239	91.78	ng	-0.01
42) 2,4,6-Tribromophenol	10.721	330	199020	150.66	ng	0.00
45) 2-Fluorobiphenyl	9.251	172	733794	82.10	ng	-0.01
79) Terphenyl-d14	13.010	244	1059267	83.97	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052121\
Data File : BF124023.D
Acq On : 21 May 2021 17:46
Operator : JU/CG
Sample : M2366-08
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
JG-SPOILS-5-11-21-B

Quant Time: May 21 18:20:35 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF051421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 14 19:05:57 2021
Response via : Initial Calibration

