


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052121\
 Data File : BF124014.D
 Acq On : 21 May 2021 11:06
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF051421.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri May 14 19:05:57 2021






AutoFind: Scans 1628, 1629, 1630; Background Corrected with Scan 1624

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	59.6	48275	PASS
68	69	0.00	2	0.7	264	PASS
69	198	0.00	100	46.6	37723	PASS
70	69	0.00	2	0.5	184	PASS
127	198	10	80	48.8	39480	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	80949	PASS
199	198	5	9	6.5	5275	PASS
275	198	10	60	27.0	21819	PASS
365	198	1	100	3.2	2611	PASS
441	198	0.01	100	10.9	8802	PASS
442	442	50	100	100.0	60149	PASS
443	442	15	24	19.2	11569	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052121\
 Data File : BF124014.D
 Acq On : 21 May 2021 11:06
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: May 21 11:39:44 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF051421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 14 19:05:57 2021
 Response via : Initial Calibration



TIC: BF124014.D\data.ms

(77) Benzidine

12.751min (-0.006) 0.00 ng




response 358409

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.10	13.80
183.00	9.80	12.60#
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052121\
Data File : BF124014.D
Acq On : 21 May 2021 11:06
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: May 21 11:39:44 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF051421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 14 19:05:57 2021
Response via : Initial Calibration



TIC: BF124014.D\data.ms

(70) Pentachlorophenol (C)

11.216min (-0.012) 24338.45 ng

response 47843

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.50	53.97
264.00	65.70	56.24
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
5/20/2021	BNA_F	<u>BF124014.D</u>
Compound Name	Response	Retention Time
DDT	144202	13.58
DDD	4760	13.27
DDE	399	12.93
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
5159	149361	3.45