

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052124\
 Data File : BF137976.D
 Acq On : 21 May 2024 13:54
 Operator : CG\JU
 Sample : PB160966TB
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB160966TB

Quant Time: May 21 14:32:28 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF051424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 15 04:12:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.034	152	152768	20.000	ng	-0.01
21) Naphthalene-d8	8.316	136	603272	20.000	ng	-0.01
39) Acenaphthene-d10	10.081	164	352804	20.000	ng	0.00
64) Phenanthrene-d10	11.575	188	621184	20.000	ng	0.00
76) Chrysene-d12	14.221	240	372435	20.000	ng	0.00
86) Perylene-d12	15.786	264	305838	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.669	112	1243973	138.319	ng	0.00
7) Phenol-d6	6.651	99	1496496	135.080	ng	0.00
23) Nitrobenzene-d5	7.598	82	942672	91.960	ng	-0.01
42) 2,4,6-Tribromophenol	10.875	330	532089	149.576	ng	0.00
45) 2-Fluorobiphenyl	9.392	172	2039073	89.029	ng	0.00
79) Terphenyl-d14	13.169	244	2366031	93.629	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052124\
Data File : BF137976.D
Acq On : 21 May 2024 13:54
Operator : CG\JU
Sample : PB160966TB
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB160966TB

Quant Time: May 21 14:32:28 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF051424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 15 04:12:16 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.034 min Scan# 841
Delta R.T. -0.011 min
Lab File: BF137976.D
Acq: 21 May 2024 13:54




Instrument :
BNA_F
ClientSampleId :
PB160966TB


Tgt Ion:152 Resp: 152768
Ion Ratio Lower Upper
152 100
150 152.7 125.7 188.5
115 60.0 47.0 70.6



#5
2-Fluorophenol
Concen: 138.319 ng
RT: 5.669 min Scan# 609
Delta R.T. 0.006 min
Lab File: BF137976.D
Acq: 21 May 2024 13:54

Tgt Ion:112 Resp: 1243973
Ion Ratio Lower Upper
112 100
64 55.5 42.4 63.6
63 28.7 21.4 32.2









#7
Phenol-d6
Concen: 135.080 ng
RT: 6.651 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF137976.D
Acq: 21 May 2024 13:54




Instrument :
BNA_F
ClientSampleId :
PB160966TB

Tgt Ion: 99 Resp: 1496496
Ion Ratio Lower Upper
99 100
42 15.9 12.6 18.8
71 34.4 27.1 40.7



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.316 min Scan# 1059
Delta R.T. -0.012 min
Lab File: BF137976.D
Acq: 21 May 2024 13:54

Tgt Ion: 136 Resp: 603272
Ion Ratio Lower Upper
136 100
137 11.0 8.7 13.1
54 6.9 5.4 8.2
68 5.8 4.6 7.0





#23

Nitrobenzene-d5

Concen: 91.960 ng

RT: 7.598 min Scan# 91

Delta R.T. -0.012 min

Lab File: BF137976.D

Acq: 21 May 2024 13:54

Instrument :

BNA_F

ClientSampleId :

PB160966TB





Tgt Ion: 82 Resp: 942672

Ion Ratio Lower Upper

82 100

128 50.7 40.3 60.5

54 44.4 36.8 55.2



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 10.081 min Scan# 1359

Delta R.T. -0.005 min

Lab File: BF137976.D

Acq: 21 May 2024 13:54




Tgt Ion:164 Resp: 352804

Ion Ratio Lower Upper

164 100

162 100.2 79.4 119.2

160 42.7 33.1 49.7









#42
2,4,6-Tribromophenol
Concen: 149.576 ng
RT: 10.875 min Scan# 1494
Delta R.T. -0.006 min
Lab File: BF137976.D
Acq: 21 May 2024 13:54




Instrument :
BNA_F
ClientSampleId :
PB160966TB


Tgt Ion:330 Resp: 532089
Ion Ratio Lower Upper
330 100
332 95.2 75.8 113.8
141 27.2 23.8 35.6



#45
2-Fluorobiphenyl
Concen: 89.029 ng
RT: 9.392 min Scan# 1242
Delta R.T. -0.006 min
Lab File: BF137976.D
Acq: 21 May 2024 13:54

Tgt Ion:172 Resp: 2039073
Ion Ratio Lower Upper
172 100
171 36.9 28.9 43.3
170 24.1 18.9 28.3









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.575 min Scan# 1
Delta R.T. -0.006 min
Lab File: BF137976.D
Acq: 21 May 2024 13:54




Instrument :
BNA_F
ClientSampleId :
PB160966TB


Tgt Ion:188 Resp: 621184
Ion Ratio Lower Upper
188 100
94 9.1 7.0 10.4
80 9.4 7.7 11.5



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.221 min Scan# 2063
Delta R.T. -0.006 min
Lab File: BF137976.D
Acq: 21 May 2024 13:54

Tgt Ion:240 Resp: 372435
Ion Ratio Lower Upper
240 100
120 11.2 8.9 13.3
236 25.4 21.0 31.4









#79
 Terphenyl-d14
 Concen: 93.629 ng
 RT: 13.169 min Scan# 1883
 Delta R.T. 0.006 min
 Lab File: BF137976.D
 Acq: 21 May 2024 13:54

Instrument :
 BNA_F
 ClientSampleId :
 PB160966TB

Tgt Ion:244 Resp: 2366031
 Ion Ratio Lower Upper
 244 100
 212 6.5 5.6 8.4
 122 9.0 8.8 13.2



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.786 min Scan# 2329
 Delta R.T. 0.000 min
 Lab File: BF137976.D
 Acq: 21 May 2024 13:54

Tgt Ion:264 Resp: 305838
 Ion Ratio Lower Upper
 264 100
 260 22.9 18.3 27.5
 265 20.0 17.6 26.4

