

Data Path : Z:\HPCHEM1\BNA_F\Data\BF052215\
 Data File : BF079452.D
 Acq On : 23 May 2015 00:17
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Quant Time: May 23 06:09:30 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 20 05:08:22 2015
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	0.00	152	0	0.00	ng	-7.13
21) Naphthalene-d8	0.00	136	0	0.00	ng	-8.71
38) Acenaphthene-d10	0.00	164	0	0.00	ng	-10.88
63) Phenanthrene-d10	12.63	188	228980	20.00	ng	-0.09
75) Chrysene-d12	15.91	240	226332	20.00	ng	-0.11
86) Perylene-d12	17.60	264	221390	20.00	ng	-0.22
System Monitoring Compounds						
5) 2-Fluorophenol	0.00	112	0	0.00	ng	
7) Phenol-d6	0.00	99	0	0.00	ng	
23) Nitrobenzene-d5	0.00	82	0	0.00	ng	
41) 2,4,6-Tribromophenol	0.00	330	0	0.00	ng	
44) 2-Fluorobiphenyl	0.00	172	0	0.00	ng	
78) Terphenyl-d14	14.63	244	819487	86.69	ng	-0.09
Target Compounds						Qvalue
64) 4,6-Dinitro-2-methylphenol	11.56	198	37758	41.77	ng	86
65) n-Nitrosodiphenylamine	11.63	169	336302	44.10	ng	99
66) 4-Bromophenyl-phenylether	12.09	248	105270	44.11	ng	# 89
67) Hexachlorobenzene	12.15	284	119257	43.72	ng	# 77
68) Atrazine	12.32	200	93064	41.97	ng	92
69) Pentachlorophenol	12.41	266	44076	32.10	ng	99
70) Phenanthrene	12.66	178	543674	42.44	ng	99
71) Anthracene	12.72	178	553096	44.01	ng	99
72) Carbazole	12.94	167	519403	43.36	ng	100
73) Di-n-butylphthalate	13.39	149	629578	43.83	ng	99
74) Fluoranthene	14.14	202	579133	43.39	ng	94
76) Benzidine	14.32	184	283116	46.41	ng	98
77) Pyrene	14.41	202	578282	44.34	ng	99
79) Butylbenzylphthalate	15.25	149	257975	45.25	ng	94
80) Benzo(a)anthracene	15.90	228	518733	41.85	ng	100
81) 3,3'-Dichlorobenzidine	15.89	252	190298	43.10	ng	98
82) Chrysene	15.94	228	496419	41.64	ng	99
83) Bis(2-ethylhexyl)phthalate	15.97	149	372638	43.15	ng	99
84) Di-n-octyl phthalate	16.74	149	626485	41.19	ng	97
85) Indeno(1,2,3-cd)pyrene	18.93	276	619840	40.81	ng	99
87) Benzo(b)fluoranthene	17.17	252	495458	40.89	ng	97
88) Benzo(k)fluoranthene	17.17	252	495458	42.23	ng	# 98
89) Benzo(a)pyrene	17.53	252	490354	43.95	ng	# 98
90) Dibenzo(a,h)anthracene	18.94	278	516691	43.57	ng	98
91) Benzo(g,h,i)perylene	19.31	276	517348	43.53	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\Data\BF052215\
Data File : BF079452.D
Acq On : 23 May 2015 00:17
Operator : TP/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Quant Time: May 23 06:09:30 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 20 05:08:22 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 0.00 ng

Expected RT: 7.13 min

Lab File: BF079452.D

Acq: 23 May 2015 00:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC



Tgt Ion: 152

Sig Exp Ratio

152 100

150 154.8

115 61.7



#5

2-Fluorophenol

Concen: 0.00 ng

Expected RT: 5.42 min

Lab File: BF079452.D

Acq: 23 May 2015 00:17


Tgt Ion: 112


Sig Exp Ratio

112 100

64 55.2

63 28.4





#7

Phenol-d6

Concen: 0.00 ng

Expected RT: 6.71 min

Lab File: BF079452.D

Acq: 23 May 2015 00:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC



Tgt Ion: 99

Sig Exp Ratio

99 100

42 21.4

71 32.9



#21

Naphthalene-d8

Concen: 0.00 ng

Expected RT: 8.71 min

Lab File: BF079452.D

Acq: 23 May 2015 00:17

Tgt Ion: 136


Sig Exp Ratio


136 100

137 11.0

54 8.6

68 6.1







#23
 Nitrobenzene-d5
 Concen: 0.00 ng
 Expected RT: 7.83 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17


Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion	Exp Ratio
82	100
128	37.1
54	57.8



#38
 Acenaphthene-d10
 Concen: 0.00 ng
 Expected RT: 10.88 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17

Tgt Ion	Exp Ratio
164	100
162	101.7
160	45.7







#41
 2,4,6-Tribromophenol
 Concen: 0.00 ng
 Expected RT: 11.86 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17


Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion: 330
 Sig Exp Ratio
 330 100
 332 96.1
 141 34.8



#44
 2-Fluorobiphenyl
 Concen: 0.00 ng
 Expected RT: 10.06 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17

Tgt Ion: 172
 Sig Exp Ratio
 172 100
 171 35.4
 170 24.3









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 12.63 min Scan# 1001
Delta R.T. -0.09 min
Lab File: BF079452.D
Acq: 23 May 2015 00:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion:188 Resp: 228980
Ion Ratio Lower Upper
188 100
94 12.0 9.4 14.0
80 12.3 9.8 14.6




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BF0
Ion 80.00 (79.70 to 80.70): BF0






#64
4,6-Dinitro-2-methylphenol
Concen: 41.77 ng
RT: 11.56 min Scan# 908
Delta R.T. -0.08 min
Lab File: BF079452.D
Acq: 23 May 2015 00:17

Tgt Ion:198 Resp: 37758
Ion Ratio Lower Upper
198 100
51 36.3 7.0 47.0
105 39.5 13.7 53.7



Abundance Ion 198.00 (197.70 to 198.70): E
Ion 51.00 (50.70 to 51.70): BF0
Ion 105.00 (104.70 to 105.70): E









#65
 n-Nitrosodiphenylamine
 Concen: 44.10 ng
 RT: 11.63 min Scan# 914
 Delta R.T. -0.08 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:169 Resp: 336302
 Ion Ratio Lower Upper
 169 100
 168 67.9 54.1 81.1
 167 38.1 30.1 45.1



#66
 4-Bromophenyl-phenylether
 Concen: 44.11 ng
 RT: 12.09 min Scan# 954
 Delta R.T. -0.08 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17

Tgt Ion:248 Resp: 105270
 Ion Ratio Lower Upper
 248 100
 250 96.9 78.0 117.0
 141 70.4 72.6 109.0#









#67
Hexachlorobenzene
Concen: 43.72 ng
RT: 12.15 min Scan# 959
Delta R.T. -0.09 min
Lab File: BF079452.D
Acq: 23 May 2015 00:17




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 284 Resp: 119257
Ion Ratio Lower Upper
284 100
142 30.8 43.2 64.8#
249 29.8 27.5 41.3



#68
Atrazine
Concen: 41.97 ng
RT: 12.32 min Scan# 974
Delta R.T. -0.08 min
Lab File: BF079452.D
Acq: 23 May 2015 00:17

Tgt Ion: 200 Resp: 93064
Ion Ratio Lower Upper
200 100
173 24.6 10.9 50.9
215 47.0 23.2 63.2









#69
 Pentachlorophenol
 Concen: 32.10 ng
 RT: 12.41 min Scan# 982
 Delta R.T. -0.08 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion: 266 Resp: 44076
 Ion Ratio Lower Upper
 266 100
 268 67.1 55.0 82.6
 264 63.1 50.9 76.3




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E






#70
 Phenanthrene
 Concen: 42.44 ng
 RT: 12.66 min Scan# 1004
 Delta R.T. -0.08 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17

Tgt Ion: 178 Resp: 543674
 Ion Ratio Lower Upper
 178 100
 176 20.3 15.9 23.9
 179 15.8 12.2 18.4



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 44.01 ng
 RT: 12.72 min Scan# 1009
 Delta R.T. -0.09 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:178 Resp: 553096
 Ion Ratio Lower Upper
 178 100
 176 19.6 15.3 22.9
 179 15.3 12.6 19.0




Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E






#72
 Carbazole
 Concen: 43.36 ng
 RT: 12.94 min Scan# 1028
 Delta R.T. -0.08 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17

Tgt Ion:167 Resp: 519403
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.5 26.3
 139 13.5 11.0 16.4



Abundance Ion 167.00 (166.70 to 167.70): E
 Ion 166.00 (165.70 to 166.70): E
 Ion 139.00 (138.70 to 139.70): E





#73

Di-n-butylphthalate

Concen: 43.83 ng

RT: 13.39 min Scan# 1068

Delta R.T. -0.08 min

Lab File: BF079452.D

Acq: 23 May 2015 00:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:149 Resp: 629578

Ion Ratio Lower Upper

149 100

150 9.1 7.0 10.4




104 4.2 3.3 4.9



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 43.39 ng

RT: 14.14 min Scan# 1133

Delta R.T. -0.09 min

Lab File: BF079452.D

Acq: 23 May 2015 00:17


Tgt Ion:202 Resp: 579133

Ion Ratio Lower Upper

202 100

101 9.7 0.0 34.7



203 17.7 0.0 37.9




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 15.91 min Scan# 1288

Delta R.T. -0.11 min

Lab File: BF079452.D

Acq: 23 May 2015 00:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:240 Resp: 226332

Ion Ratio Lower Upper




240 100

120 12.9 8.6 12.8#

236 26.5 21.0 31.6



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 46.41 ng

RT: 14.32 min Scan# 1149

Delta R.T. -0.09 min

Lab File: BF079452.D

Acq: 23 May 2015 00:17


Tgt Ion:184 Resp: 283116

Ion Ratio Lower Upper



184 100


185 14.8 12.8 19.2

183 11.4 9.0 13.4



Abundance Ion 184.00 (183.70 to 184.70): E
Ion 185.00 (184.70 to 185.70): E
Ion 183.00 (182.70 to 183.70): E









#77
 Pyrene
 Concen: 44.34 ng
 RT: 14.41 min Scan# 1157
 Delta R.T. -0.09 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion: 202 Resp: 578282
 Ion Ratio Lower Upper
 202 100
 200 22.0 17.5 26.3
 203 17.9 14.7 22.1




Abundance Ion 202.00 (201.70 to 202.70): E
 Ion 200.00 (199.70 to 200.70): E
 Ion 203.00 (202.70 to 203.70): E






#78
 Terphenyl-d14
 Concen: 86.69 ng
 RT: 14.63 min Scan# 1176
 Delta R.T. -0.09 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17

Tgt Ion: 244 Resp: 819487
 Ion Ratio Lower Upper
 244 100
 212 6.7 5.6 8.4
 122 10.7 10.5 15.7



Abundance Ion 244.00 (243.70 to 244.70): E
 Ion 212.00 (211.70 to 212.70): E
 Ion 122.00 (121.70 to 122.70): E









#79
Butylbenzylphthalate
Concen: 45.25 ng
RT: 15.25 min Scan# 1230
Delta R.T. -0.09 min
Lab File: BF079452.D
Acq: 23 May 2015 00:17




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion	Ratio	Lower	Upper
149	100		
91	75.2	56.2	84.2
206	16.7	15.2	22.8



#80
Benzo(a)anthracene
Concen: 41.85 ng
RT: 15.90 min Scan# 1287
Delta R.T. -0.11 min
Lab File: BF079452.D
Acq: 23 May 2015 00:17

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.0	22.3	33.5
229	19.8	16.1	24.1









#81
 3,3'-Dichlorobenzidine
 Concen: 43.10 ng
 RT: 15.89 min Scan# 1286
 Delta R.T. -0.10 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:252 Resp: 190298
 Ion Ratio Lower Upper
 252 100
 254 63.5 52.2 78.2
 126 8.5 6.6 9.8



#82
 Chrysene
 Concen: 41.64 ng
 RT: 15.94 min Scan# 1291
 Delta R.T. -0.11 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17

Tgt Ion:228 Resp: 496419
 Ion Ratio Lower Upper
 228 100
 226 31.6 24.8 37.2
 229 20.1 16.1 24.1









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 43.15 ng
 RT: 15.97 min Scan# 1293
 Delta R.T. -0.10 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion	Ratio	Lower	Upper
149	100		
167	26.0	21.2	31.8
279	3.2	2.6	4.0



#84
 Di-n-octyl phthalate
 Concen: 41.19 ng
 RT: 16.74 min Scan# 1361
 Delta R.T. -0.15 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17

Tgt Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	9.9	9.0	13.6









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 40.81 ng
 RT: 18.93 min Scan# 1552
 Delta R.T. -0.29 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:276 Resp: 619840
 Ion Ratio Lower Upper
 276 100
 138 25.9 21.1 31.7
 277 24.9 19.9 29.9



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 17.60 min Scan# 1436
 Delta R.T. -0.22 min
 Lab File: BF079452.D
 Acq: 23 May 2015 00:17

Tgt Ion:264 Resp: 221390
 Ion Ratio Lower Upper
 264 100
 260 23.7 19.2 28.8
 265 21.7 17.2 25.8





#87

Benzo(b)fluoranthene

Concen: 40.89 ng

RT: 17.17 min Scan# 1398

Delta R.T. -0.18 min





Lab File: BF079452.D

Acq: 23 May 2015 00:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion: 252 Resp: 495458

Ion	Ratio	Lower	Upper
252	100		
253	22.0	18.1	27.1
125	9.5	9.4	14.2



#88

Benzo(k)fluoranthene

Concen: 42.23 ng

RT: 17.17 min Scan# 1398




Delta R.T. -0.22 min


Lab File: BF079452.D

Acq: 23 May 2015 00:17

Tgt Ion: 252 Resp: 495458

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.4	26.0
125	9.5	6.1	9.1#





#89

Benzo(a)pyrene

Concen: 43.95 ng

RT: 17.53 min Scan# 1430

Delta R.T. -0.22 min

Lab File: BF079452.D

Acq: 23 May 2015 00:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:252 Resp: 490354

Ion Ratio Lower Upper

252 100

253 22.6 18.0 27.0




125 12.4 8.2 12.2#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 43.57 ng

RT: 18.94 min Scan# 1553

Delta R.T. -0.31 min

Lab File: BF079452.D

Acq: 23 May 2015 00:17


Tgt Ion:278 Resp: 516691

Ion Ratio Lower Upper

278 100

139 18.7 13.1 19.7



279 23.8 19.1 28.7






Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 43.53 ng

RT: 19.31 min Scan# 1586

Delta R.T. -0.31 min

Lab File: BF079452.D

Acq: 23 May 2015 00:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC

Tgt Ion: 276 Resp: 517348

Ion Ratio Lower Upper

276 100

277 23.5 18.8 28.2

138 19.7 15.2 22.8

