

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052215\
 Data File : BF079436.D
 Acq On : 22 May 2015 16:40
 Operator : TP/IZ
 Sample : G2255-20
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 LF001MW001-150518

Manual Integrations
 APPROVED

apatel
 5/27/2015 8:15:36 AM

Quant Time: May 26 11:30:49 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 22 22:53:44 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.06	152	54209m	20.00	ng	0.01
21) Naphthalene-d8	8.64	136	223286	20.00	ng	0.01
38) Acenaphthene-d10	10.81	164	114141	20.00	ng	0.01
63) Phenanthrene-d10	12.64	188	212890	20.00	ng	0.00
75) Chrysene-d12	15.93	240	216071	20.00	ng	-0.03
86) Perylene-d12	17.71	264	216104	20.00	ng	-0.14
System Monitoring Compounds						
5) 2-Fluorophenol	5.35	112	23858m	7.58	ng	0.02
7) Phenol-d6	6.71	99	37137m	8.92	ng	0.08
23) Nitrobenzene-d5	7.76	82	17188	4.42	ng	0.01
41) 2,4,6-Tribromophenol	11.78	330	21516	17.42	ng	0.00
44) 2-Fluorobiphenyl	9.99	172	102096	12.46	ng	0.01
78) Terphenyl-d14	14.63	244	98745	10.94	ng	0.00
Target Compounds						Qvalue
10) Phenol	6.72	94	141358m	29.27	ng	
12) 1,3-Dichlorobenzene	6.98	146	145625	36.27	ng	# 95
13) 1,4-Dichlorobenzene	7.08	146	386515	93.55	ng	98
14) 1,2-Dichlorobenzene	7.27	146	1699718	441.91	ng	97
17) 2-Methylphenol	7.43	107	108214m	37.65	ng	
20) 3+4-Methylphenols	7.62	107	328897m	85.88	ng	
27) 2,4-Dimethylphenol	8.30	122	127412m	39.95	ng	
30) 1,2,4-Trichlorobenzene	8.58	180	10882	3.49	ng	97
31) Naphthalene	8.67	128	1002535	94.23	ng	99
37) 2-Methylnaphthalene	9.53	142	93323	12.66	ng	# 95
45) 1,1'-Biphenyl	10.11	154	38534	4.25	ng	90
49) Dimethylphthalate	10.50	163	17146	2.05	ng	# 89
59) Diethylphthalate	11.37	149	120612	14.53	ng	98
83) Bis(2-ethylhexyl)phthalate	15.99	149	58802	7.13	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052215\
Data File : BF079436.D
Acq On : 22 May 2015 16:40
Operator : TP/IZ
Sample : G2255-20
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
LF001MW001-150518

Manual Integrations
APPROVED

apatel
5/27/2015 8:15:36 AM

Quant Time: May 26 11:30:49 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 22 22:53:44 2015
Response via : Initial Calibration

