

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052215\
 Data File : BF079440.D
 Acq On : 22 May 2015 18:34
 Operator : TP/IZ
 Sample : G2355-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FRACTANKSAMPLE1

Manual Integrations
 APPROVED

apatel
 5/27/2015 8:16:02 AM

Quant Time: May 22 23:52:44 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 22 22:53:44 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.05	152	66391	20.00	ng	0.00
21) Naphthalene-d8	8.63	136	281485	20.00	ng	0.00
38) Acenaphthene-d10	10.80	164	137168	20.00	ng	0.00
63) Phenanthrene-d10	12.63	188	254320	20.00	ng	-0.01
75) Chrysene-d12	15.92	240	245692	20.00	ng	-0.05
86) Perylene-d12	17.67	264	236731	20.00	ng	-0.18
System Monitoring Compounds						
5) 2-Fluorophenol	5.34	112	181957m	47.18	ng	0.01
7) Phenol-d6	6.64	99	145295m	28.50	ng	0.01
23) Nitrobenzene-d5	7.75	82	524889	107.17	ng	0.00
41) 2,4,6-Tribromophenol	11.78	330	158139	106.57	ng	0.00
44) 2-Fluorobiphenyl	9.98	172	1033085	104.93	ng	0.00
78) Terphenyl-d14	14.63	244	1092148	106.42	ng	0.00
Target Compounds						
9) Benzaldehyde	6.50	77	90268m	26.28	ng	Qvalue
22) Acetophenone	7.56	105	138601	21.79	ng	# 86
31) Naphthalene	8.65	128	74073	5.52	ng	98
32) Benzoic acid	8.35	122	33773	17.96	ng	93
73) Di-n-butylphthalate	13.39	149	551590	34.57	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052215\
Data File : BF079440.D
Acq On : 22 May 2015 18:34
Operator : TP/IZ
Sample : G2355-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FRACTANKSAMPLE1

Manual Integrations
APPROVED

apatel
5/27/2015 8:16:02 AM

Quant Time: May 22 23:52:44 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 22 22:53:44 2015
Response via : Initial Calibration

