

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052215\
 Data File : BF079456.D
 Acq On : 23 May 2015 5:02
 Operator : TP/IZ
 Sample : G2289-13
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-31S

Manual Integrations
 APPROVED

apatel
 5/27/2015 8:17:17 AM

Quant Time: May 23 06:52:05 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 23 06:12:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.05	152	71860	20.00	ng	0.00
21) Naphthalene-d8	8.63	136	289070	20.00	ng	0.00
38) Acenaphthene-d10	10.80	164	137394	20.00	ng	0.00
63) Phenanthrene-d10	12.63	188	266384	20.00	ng	0.00
75) Chrysene-d12	15.91	240	256740	20.00	ng	0.00
86) Perylene-d12	17.60	264	266670	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.34	112	669941m	160.48	ng	0.01
7) Phenol-d6	6.63	99	842428m	152.67	ng	0.00
23) Nitrobenzene-d5	7.75	82	469737	93.39	ng	0.00
41) 2,4,6-Tribromophenol	11.77	330	245749	165.34	ng	0.00
44) 2-Fluorobiphenyl	9.98	172	934019	94.71	ng	0.00
78) Terphenyl-d14	14.63	244	1023968	95.49	ng	0.00
Target Compounds						
49) Dimethylphthalate	10.49	163	73632	7.30	ng	Qvalue # 95
70) Phenanthrene	12.66	178	219967	14.76	ng	99
71) Anthracene	12.72	178	65762	4.50	ng	100
74) Fluoranthene	14.14	202	411461	26.50	ng	94
77) Pyrene	14.41	202	361239	24.42	ng	99
80) Benzo(a)anthracene	15.90	228	187877	13.36	ng	99
82) Chrysene	15.94	228	138390	10.23	ng	98
85) Indeno(1,2,3-cd)pyrene	18.93	276	85131	4.94	ng	94
87) Benzo(b)fluoranthene	17.17	252	190326m	13.04	ng	
88) Benzo(k)fluoranthene	17.20	252	58427m	4.13	ng	
89) Benzo(a)pyrene	17.53	252	137723	10.25	ng	# 95
91) Benzo(a,h,i)perylene	19.31	276	83854	5.86	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052215\
Data File : BF079456.D
Acq On : 23 May 2015 5:02
Operator : TP/IZ
Sample : G2289-13
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-31S

Manual Integrations
APPROVED

apatel
5/27/2015 8:17:17 AM

Quant Time: May 23 06:52:05 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 23 06:12:59 2015
Response via : Initial Calibration

