

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052215\
 Data File : BF079460.D
 Acq On : 23 May 2015 6:57
 Operator : TP/IZ
 Sample : G2255-10 10X
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 KY032MW006-150512

Quant Time: May 26 12:42:57 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 25 06:56:43 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev (Min)
1) 1,4-Dichlorobenzene-d4	7.05	152	63254	20.00	ng	0.00
21) Naphthalene-d8	8.64	136	260119	20.00	ng	0.01
38) Acenaphthene-d10	10.80	164	139065	20.00	ng	0.00
63) Phenanthrene-d10	12.63	188	268533	20.00	ng	0.00
75) Chrysene-d12	15.91	240	269125	20.00	ng	0.00
86) Perylene-d12	17.62	264	280033	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.36	112	3185m	0.87	ng	0.03
7) Phenol-d6	0.00	99	0	0.00	ng	
23) Nitrobenzene-d5	7.76	82	1386	0.31	ng	0.01
41) 2,4,6-Tribromophenol	11.78	330	1106	0.74	ng	0.00
44) 2-Fluorobiphenyl	9.99	172	9389	0.94	ng	0.01
78) Terphenyl-d14	14.63	244	10900	0.97	ng	0.00
Target Compounds						
2) 1,4-Dioxane	1.90	88	401449m	251.27	ng	97
10) Phenol	6.70	94	152567m	27.08	ng	
11) bis(2-Chloroethyl)ether	6.75	93	165903m	40.16	ng	
14) 1,2-Dichlorobenzene	7.26	146	27893	6.21	ng	
17) 2-Methylphenol	7.42	107	72449m	21.60	ng	
20) 3+4-Methylphenols	7.62	107	401094m	89.75	ng	99
27) 2,4-Dimethylphenol	8.28	122	174210m	46.88	ng	
83) Bis(2-ethylhexyl)phthalate	15.97	149	26780	2.61	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052215\
Data File : BF079460.D
Acq On : 23 May 2015 6:57
Operator : TP/IZ
Sample : G2255-10 10X
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
KY032MW006-150512

Quant Time: May 26 12:42:57 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 25 06:56:43 2015
Response via : Initial Calibration

