

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052215\
 Data File : BF079466.D
 Acq On : 23 May 2015 9:49
 Operator : TP/IZ
 Sample : G2319-01 2X
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 GAS-LINE-WASTE-COMP

Manual Integrations
 APPROVED

apatel
 5/27/2015 8:18:15 AM

Quant Time: May 25 07:38:36 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 25 06:56:43 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.05	152	70826	20.00	ng	0.00
21) Naphthalene-d8	8.63	136	286570	20.00	ng	0.00
38) Acenaphthene-d10	10.80	164	136456	20.00	ng	0.00
63) Phenanthrene-d10	12.63	188	255045	20.00	ng	0.00
75) Chrysene-d12	15.92	240	256146	20.00	ng	0.01
86) Perylene-d12	17.69	264	252794	20.00	ng	0.08
System Monitoring Compounds						
5) 2-Fluorophenol	5.36	112	247518m	60.16	ng	0.03
7) Phenol-d6	6.65	99	332893m	61.21	ng	0.02
23) Nitrobenzene-d5	7.75	82	124819	25.03	ng	0.00
41) 2,4,6-Tribromophenol	11.78	330	76224	51.64	ng	0.00
44) 2-Fluorobiphenyl	9.98	172	369969	37.77	ng	0.00
78) Terphenyl-d14	14.63	244	391182	36.56	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	10.49	163	34749	3.47	ng	98
70) Phenanthrene	12.66	178	116792	8.19	ng	99
71) Anthracene	12.73	178	39645	2.83	ng	96
74) Fluoranthene	14.14	202	214705	14.44	ng	96
77) Pyrene	14.41	202	183080	12.40	ng	99
80) Benzo(a)anthracene	15.91	228	103181m	7.36	ng	
82) Chrysene	15.95	228	80751	5.99	ng	96
85) Indeno(1,2,3-cd)pyrene	19.05	276	45722m	2.66	ng	
87) Benzo(b)fluoranthene	17.23	252	104711m	7.57	ng	
88) Benzo(k)fluoranthene	17.27	252	39066m	2.92	ng	
89) Benzo(a)pyrene	17.63	252	74214	5.83	ng	# 93
91) Benzo(a,h,i)perylene	19.45	276	50005	3.68	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052215\
Data File : BF079466.D
Acq On : 23 May 2015 9:49
Operator : TP/IZ
Sample : G2319-01 2X
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
GAS-LINE-WASTE-COMP

Manual Integrations
APPROVED

apatel
5/27/2015 8:18:15 AM

Quant Time: May 25 07:38:36 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF043015.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 25 06:56:43 2015
Response via : Initial Calibration

