

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF052216\  
 Data File : BF087279.D  
 Acq On : 22 May 2016 9:33  
 Operator : UM/SJ  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 SSTDCCC040

Quant Time: May 23 03:34:34 2016  
 Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF051716.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon May 23 01:56:40 2016  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|---------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 6.56  | 152  | 149471   | 20.00 | ng    | 0.00     |
| 21) Naphthalene-d8        | 7.85  | 136  | 640576   | 20.00 | ng    | 0.00     |
| 38) Acenaphthene-d10      | 9.60  | 164  | 293690   | 20.00 | ng    | 0.01     |
| 63) Phenanthrene-d10      | 11.07 | 188  | 520857   | 20.00 | ng    | 0.00     |
| 75) Chrysene-d12          | 13.70 | 240  | 322837   | 20.00 | ng    | 0.00     |
| 86) Perylene-d12          | 15.05 | 264  | 289201   | 20.00 | ng    | 0.01     |

## System Monitoring Compounds

|                          |       |     |         |       |    |      |
|--------------------------|-------|-----|---------|-------|----|------|
| 5) 2-Fluorophenol        | 5.14  | 112 | 718217  | 80.77 | ng | 0.01 |
| 7) Phenol-d6             | 6.24  | 99  | 924421  | 77.51 | ng | 0.01 |
| 23) Nitrobenzene-d5      | 7.14  | 82  | 929952  | 72.35 | ng | 0.00 |
| 41) 2,4,6-Tribromophenol | 10.39 | 330 | 209893  | 64.67 | ng | 0.00 |
| 44) 2-Fluorobiphenyl     | 8.92  | 172 | 1271137 | 71.68 | ng | 0.00 |
| 78) Terphenyl-d14        | 12.65 | 244 | 1272403 | 89.16 | ng | 0.00 |

## Target Compounds

|                                |      |     |         |       |    | Qvalue |
|--------------------------------|------|-----|---------|-------|----|--------|
| 2) 1,4-Dioxane                 | 1.96 | 88  | 177540  | 38.81 | ng | # 73   |
| 3) Pyridine                    | 2.58 | 79  | 397140  | 35.89 | ng | # 66   |
| 4) n-Nitrosodimethylamine      | 2.56 | 42  | 312887  | 39.90 | ng | # 49   |
| 6) Aniline                     | 6.23 | 93  | 556597  | 36.74 | ng | 97     |
| 8) 2-Chlorophenol              | 6.35 | 128 | 412308  | 38.11 | ng | 99     |
| 9) Benzaldehyde                | 6.11 | 77  | 238415  | 36.30 | ng | 97     |
| 10) Phenol                     | 6.25 | 94  | 572478  | 38.18 | ng | 85     |
| 11) bis(2-Chloroethyl)ether    | 6.31 | 93  | 442759  | 39.33 | ng | 89     |
| 12) 1,3-Dichlorobenzene        | 6.49 | 146 | 447458  | 38.91 | ng | # 90   |
| 13) 1,4-Dichlorobenzene        | 6.73 | 146 | 406876  | 34.58 | ng | 95     |
| 14) 1,2-Dichlorobenzene        | 6.73 | 146 | 410230  | 39.82 | ng | 99     |
| 15) Benzyl Alcohol             | 6.73 | 79  | 367572  | 41.40 | ng | 89     |
| 16) 2,2'-oxybis(1-Chloropropan | 6.84 | 45  | 896995  | 38.76 | ng | 88     |
| 17) 2-Methylphenol             | 6.86 | 107 | 334954  | 37.70 | ng | # 80   |
| 18) Hexachloroethane           | 7.06 | 117 | 179192  | 39.83 | ng | 98     |
| 19) n-Nitroso-di-n-propylamine | 7.00 | 70  | 344721  | 38.02 | ng | # 84   |
| 20) 3+4-Methylphenols          | 7.02 | 107 | 440675  | 37.60 | ng | # 61   |
| 22) Acetophenone               | 6.99 | 105 | 658548  | 42.01 | ng | 96     |
| 24) Nitrobenzene               | 7.16 | 77  | 526368  | 37.51 | ng | 92     |
| 25) Isophorone                 | 7.40 | 82  | 903659  | 38.41 | ng | 97     |
| 26) 2-Nitrophenol              | 7.47 | 139 | 215758  | 37.11 | ng | # 57   |
| 27) 2,4-Dimethylphenol         | 7.54 | 122 | 358633  | 36.15 | ng | 93     |
| 28) bis(2-Chloroethoxy)methane | 7.62 | 93  | 513671  | 39.77 | ng | 99     |
| 29) 2,4-Dichlorophenol         | 7.72 | 162 | 343342  | 39.27 | ng | 91     |
| 30) 1,2,4-Trichlorobenzene     | 7.79 | 180 | 370920  | 38.24 | ng | 95     |
| 31) Naphthalene                | 7.87 | 128 | 1247994 | 39.87 | ng | 99     |
| 32) Benzoic acid               | 7.71 | 122 | 154481  | 26.12 | ng | # 81   |
| 33) 4-Chloroaniline            | 7.94 | 127 | 460060  | 35.36 | ng | 98     |
| 34) Hexachlorobutadiene        | 7.99 | 225 | 235129  | 42.71 | ng | 99     |
| 35) Caprolactam                | 8.34 | 113 | 109744  | 37.70 | ng | # 72   |
| 36) 4-Chloro-3-methylphenol    | 8.44 | 107 | 390886  | 37.07 | ng | 94     |
| 37) 2-Methylnaphthalene        | 8.66 | 142 | 721958  | 36.06 | ng | 95     |
| 39) 1,2,4,5-Tetrachlorobenzene | 8.73 | 216 | 367897  | 42.89 | ng | 99     |
| 40) Hexachlorocyclopentadiene  | 8.71 | 237 | 78659   | 29.27 | ng | 97     |

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF052216\  
 Data File : BF087279.D  
 Acq On : 22 May 2016 9:33  
 Operator : UM/SJ  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1


Instrument :  
 BNA\_F  
 ClientSampleId :  
 SSTDCCC040

Quant Time: May 23 03:34:34 2016  
 Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF051716.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon May 23 01:56:40 2016  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|--------------------------------|-------|------|----------|--------|-------|----------|
| 42) 2,4,6-Trichlorophenol      | 8.86  | 196  | 216577   | 38.93  | ng    | 93       |
| 43) 2,4,5-Trichlorophenol      | 8.90  | 196  | 223598   | 38.69  | ng #  | 86       |
| 45) 1,1'-Biphenyl              | 9.03  | 154  | 1070492  | 43.96  | ng    | 98       |
| 46) 2-Chloronaphthalene        | 9.05  | 162  | 762865   | 40.81  | ng    | 98       |
| 47) 2-Nitroaniline             | 9.16  | 65   | 288055   | 40.21  | ng #  | 81       |
| 48) Acenaphthylene             | 9.46  | 152  | 1093216  | 38.30  | ng    | 99       |
| 49) Dimethylphthalate          | 9.34  | 163  | 842843   | 37.62  | ng    | 99       |
| 50) 2,6-Dinitrotoluene         | 9.40  | 165  | 167314   | 34.39  | ng #  | 68       |
| 51) Acenaphthene               | 9.63  | 154  | 667524   | 37.43  | ng    | 98       |
| 52) 3-Nitroaniline             | 9.58  | 138  | 178397   | 33.88  | ng #  | 73       |
| 53) 2,4-Dinitrophenol          | 9.70  | 184  | 59582    | 24.85  | ng #  | 82       |
| 54) Dibenzofuran               | 9.80  | 168  | 904809   | 39.23  | ng    | 99       |
| 55) 4-Nitrophenol              | 9.80  | 139  | 460404   | 100.28 | ng #  | 23       |
| 56) 2,4-Dinitrotoluene         | 9.82  | 165  | 249226   | 40.00  | ng #  | 82       |
| 57) Fluorene                   | 10.14 | 166  | 681852   | 36.80  | ng    | 99       |
| 58) 2,3,4,6-Tetrachlorophenol  | 9.93  | 232  | 166349   | 36.96  | ng    | 97       |
| 59) Diethylphthalate           | 10.03 | 149  | 847870   | 39.64  | ng    | 99       |
| 60) 4-Chlorophenyl-phenylether | 10.14 | 204  | 371304   | 42.27  | ng    | 94       |
| 61) 4-Nitroaniline             | 10.19 | 138  | 163342   | 31.41  | ng #  | 63       |
| 62) Azobenzene                 | 10.30 | 77   | 990771   | 40.20  | ng    | 87       |
| 64) 4,6-Dinitro-2-methylphenol | 10.23 | 198  | 102098   | 29.76  | ng    | 88       |
| 65) n-Nitrosodiphenylamine     | 10.26 | 169  | 630663   | 38.48  | ng    | 98       |
| 66) 4-Bromophenyl-phenylether  | 10.62 | 248  | 213352   | 39.83  | ng #  | 83       |
| 67) Hexachlorobenzene          | 10.68 | 284  | 222492   | 35.83  | ng #  | 73       |
| 68) Atrazine                   | 10.80 | 200  | 211795   | 39.61  | ng    | 94       |
| 69) Pentachlorophenol          | 10.89 | 266  | 56961    | 24.42  | ng    | 96       |
| 70) Phenanthrene               | 11.10 | 178  | 1098348  | 38.88  | ng    | 100      |
| 71) Anthracene                 | 11.14 | 178  | 1061383  | 37.15  | ng    | 99       |
| 72) Carbazole                  | 11.31 | 167  | 925731   | 35.47  | ng    | 98       |
| 73) Di-n-butylphthalate        | 11.64 | 149  | 1317200  | 43.94  | ng #  | 98       |
| 74) Fluoranthene               | 12.27 | 202  | 977701   | 34.54  | ng    | 97       |
| 76) Benzidine                  | 12.41 | 184  | 315235   | 32.65  | ng    | 98       |
| 77) Pyrene                     | 12.50 | 202  | 1044360  | 44.78  | ng    | 100      |
| 79) Butylbenzylphthalate       | 13.13 | 149  | 498918   | 50.67  | ng #  | 88       |
| 80) Benzo(a)anthracene         | 13.69 | 228  | 716325   | 38.37  | ng    | 99       |
| 81) 3,3'-Dichlorobenzidine     | 13.66 | 252  | 465279   | 39.17  | ng    | 99       |
| 82) Chrysene                   | 13.72 | 228  | 749563   | 41.50  | ng    | 99       |
| 83) Bis(2-ethylhexyl)phthalate | 13.69 | 149  | 625790   | 52.22  | ng #  | 99       |
| 84) Di-n-octyl phthalate       | 14.30 | 149  | 979377   | 47.45  | ng #  | 84       |
| 85) Indeno(1,2,3-cd)pyrene     | 16.27 | 276  | 670101   | 42.27  | ng    | 95       |
| 87) Benzo(b)fluoranthene       | 14.69 | 252  | 1286142  | 67.04  | ng    | 99       |
| 88) Benzo(k)fluoranthene       | 14.69 | 252  | 1286536  | 90.07  | ng    | 98       |
| 89) Benzo(a)pyrene             | 14.99 | 252  | 613319   | 38.24  | ng #  | 96       |
| 90) Dibenzo(a,h)anthracene     | 16.29 | 278  | 560621   | 41.52  | ng #  | 93       |
| 91) Benzo(g,h,i)perylene       | 16.64 | 276  | 552035   | 40.48  | ng #  | 93       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 6.56 min Scan# 435

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion:152 Resp: 149471

Ion Ratio Lower Upper

152 100

150 154.9 125.8 188.6




115 62.1 51.5 77.3



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 38.81 ng

RT: 1.96 min Scan# 33

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion: 88 Resp: 177540

Ion Ratio Lower Upper

88 100

58 94.6 59.6 89.4#



43 45.2 21.8 32.6#




Abundance Ion 88.00 (87.70 to 88.70): BF0

Ion 58.00 (57.70 to 58.70): BF0

Ion 43.00 (42.70 to 43.70): BF0





#3

Pyridine

Concen: 35.89 ng

RT: 2.58 min Scan# 87

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 79 Resp: 397140

Ion Ratio Lower Upper

79 100

52 100.4 55.9 83.9#

51 51.9 28.1 42.1#



#4

n-Nitrosodimethylamine

Concen: 39.90 ng

RT: 2.56 min Scan# 85

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33




Tgt Ion: 42 Resp: 312887


Ion Ratio Lower Upper

42 100

74 85.4 123.4 185.0#

44 6.0 5.8 8.6





#5

2-Fluorophenol

Concen: 80.77 ng

RT: 5.14 min Scan# 311

Delta R.T. 0.01 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040




|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 112   | Resp: | 718217 |
| Ion      | Ratio | Lower | Upper  |
| 112      | 100   |       |        |
| 64       | 60.9  | 52.1  | 78.1   |
| 63       | 33.9  | 25.7  | 38.5   |



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 36.74 ng


RT: 6.23 min Scan# 406

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33



|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 93    | Resp: | 556597 |
| Ion      | Ratio | Lower | Upper  |
| 93       | 100   |       |        |
| 66       | 51.2  | 39.0  | 58.6   |
| 65       | 24.4  | 20.6  | 31.0   |




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 77.51 ng

RT: 6.24 min Scan# 407

Delta R.T. 0.01 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 99 Resp: 924421

Ion Ratio Lower Upper

99 100

42 26.3 13.6 20.4#




71 37.7 24.6 36.8#



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#8

2-Chlorophenol

Concen: 38.11 ng

RT: 6.35 min Scan# 417

Delta R.T. 0.01 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion: 128 Resp: 412308

Ion Ratio Lower Upper

128 100

130 33.4 12.5 52.5



64 47.6 27.3 67.3




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF0









#9  
Benzaldehyde  
Concen: 36.30 ng  
RT: 6.11 min Scan# 396  
Delta R.T. 0.01 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33




Instrument :  
BNA\_F  
ClientSampleId :  
SSTDCCC040


Tgt Ion: 77 Resp: 238415  
Ion Ratio Lower Upper  
77 100  
105 96.8 72.7 112.7  
106 91.7 70.8 110.8



#10  
Phenol  
Concen: 38.18 ng  
RT: 6.25 min Scan# 408  
Delta R.T. 0.01 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33

Tgt Ion: 94 Resp: 572478  
Ion Ratio Lower Upper  
94 100  
65 33.1 7.2 47.2  
66 44.9 14.9 54.9





#11

bis(2-Chloroethyl)ether

Concen: 39.33 ng

RT: 6.31 min Scan# 413

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 93 Resp: 442759

Ion Ratio Lower Upper

93 100

63 91.5 58.0 98.0




95 32.4 11.9 51.9



Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 63.00 (62.70 to 63.70): BF0

Ion 95.00 (94.70 to 95.70): BF0



#12

1,3-Dichlorobenzene

Concen: 38.91 ng

RT: 6.49 min Scan# 429

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion: 146 Resp: 447458

Ion Ratio Lower Upper

146 100

148 63.0 52.4 78.6



75 42.5 22.8 34.2#




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 75.00 (74.70 to 75.70): BF0









#13  
 1,4-Dichlorobenzene  
 Concen: 34.58 ng  
 RT: 6.73 min Scan# 450  
 Delta R.T. 0.16 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33




Instrument :  
 BNA\_F  
 ClientSampleId :  
 SSTDCCC040


Tgt Ion:146 Resp: 406876  
 Ion Ratio Lower Upper  
 146 100  
 148 64.3 52.2 78.4  
 111 46.0 30.7 46.1



#14  
 1,2-Dichlorobenzene  
 Concen: 39.82 ng  
 RT: 6.73 min Scan# 450  
 Delta R.T. -0.00 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33

Tgt Ion:146 Resp: 410230  
 Ion Ratio Lower Upper  
 146 100  
 148 64.3 51.0 76.6  
 111 46.0 36.2 54.4





#15

Benzyl Alcohol

Concen: 41.40 ng

RT: 6.73 min Scan# 450

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 79 Resp: 367572

Ion Ratio Lower Upper

79 100

108 69.4 68.4 102.6




77 64.2 50.6 76.0



Abundance Ion 79.00 (78.70 to 79.70): BF0

Ion 108.00 (107.70 to 108.70): BF0

Ion 77.00 (76.70 to 77.70): BF0



#16

2,2'-oxybis(1-chloropropane)

Concen: 38.76 ng

RT: 6.84 min Scan# 460

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion: 45 Resp: 896995

Ion Ratio Lower Upper

45 100

77 11.4 0.0 36.4



79 8.6 0.0 33.4




Abundance Ion 45.00 (44.70 to 45.70): BF0

Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0









#17  
2-Methylphenol  
Concen: 37.70 ng  
RT: 6.86 min Scan# 461  
Delta R.T. -0.00 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33




Instrument :  
BNA\_F  
ClientSampleId :  
SSTDCCC040



| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 107     | 100   |       |       |
| 108     | 110.8 | 93.7  | 140.5 |
| 77      | 69.6  | 33.4  | 50.0# |
| 79      | 64.0  | 34.3  | 51.5# |



#18  
Hexachloroethane  
Concen: 39.83 ng  
RT: 7.06 min Scan# 479  
Delta R.T. -0.00 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 117     | 100   |       |       |
| 119     | 96.2  | 76.5  | 114.7 |
| 201     | 82.7  | 63.0  | 94.6  |









#19  
n-Nitroso-di-n-propylamine  
Concen: 38.02 ng  
RT: 7.00 min Scan# 474  
Delta R.T. 0.01 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33


Instrument :  
BNA\_F  
ClientSampleId :  
SSTDCCC040


Tgt Ion: 70 Resp: 344721  
Ion Ratio Lower Upper  
70 100  
42 71.6 43.1 64.7#  
101 10.3 8.1 12.1  
130 23.8 18.6 28.0



#20  
3+4-Methylphenols  
Concen: 37.60 ng  
RT: 7.02 min Scan# 475  
Delta R.T. -0.00 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33

Tgt Ion: 107 Resp: 440675  
Ion Ratio Lower Upper  
107 100  
108 87.6 68.5 108.5  
77 39.4 101.6 141.6#  
79 32.2 7.4 47.4





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 7.85 min Scan# 548

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040

Tgt Ion:136 Resp: 640576


Ion Ratio Lower Upper

136 100

137 10.7 8.8 13.2

54 9.5 5.0 7.4#

68 4.9 4.4 6.6






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0



#22

Acetophenone

Concen: 42.01 ng

RT: 6.99 min Scan# 473

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Tgt Ion:105 Resp: 658548


Ion Ratio Lower Upper

105 100

71 6.8 4.8 7.2

51 36.7 32.6 49.0

120 20.7 16.5 24.7






Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF0

Ion 51.00 (50.70 to 51.70): BF0

Ion 120.00 (119.70 to 120.70): E





#23

Nitrobenzene-d5

Concen: 72.35 ng

RT: 7.14 min Scan# 486

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 82 Resp: 929952

Ion Ratio Lower Upper

82 100

128 38.8 40.6 61.0#




54 69.7 38.1 57.1#



Abundance Ion 82.00 (81.70 to 82.70): BF0

Ion 128.00 (127.70 to 128.70): BF0

Ion 54.00 (53.70 to 54.70): BF0



#24

Nitrobenzene

Concen: 37.51 ng

RT: 7.16 min Scan# 488

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion: 77 Resp: 526368

Ion Ratio Lower Upper

77 100

123 47.5 33.0 49.4



65 14.6 10.8 16.2




Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 123.00 (122.70 to 123.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#25

Isophorone

Concen: 38.41 ng

RT: 7.40 min Scan# 509

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 82 Resp: 903659

Ion Ratio Lower Upper

82 100

95 7.6 5.1 7.7

138 16.4 12.4 18.6



#26

2-Nitrophenol

Concen: 37.11 ng

RT: 7.47 min Scan# 515

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33




Tgt Ion: 139 Resp: 215758


Ion Ratio Lower Upper

139 100

109 37.4 19.5 29.3#

65 81.0 37.5 56.3#









#27  
2,4-Dimethylphenol  
Concen: 36.15 ng  
RT: 7.54 min Scan# 521  
Delta R.T. 0.01 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33




Instrument :  
BNA\_F  
ClientSampleId :  
SSTDCCC040


Tgt Ion:122 Resp: 358633  
Ion Ratio Lower Upper  
122 100  
107 112.0 82.0 123.0  
121 59.9 46.1 69.1



#28  
bis(2-Chloroethoxy)methane  
Concen: 39.77 ng  
RT: 7.62 min Scan# 528  
Delta R.T. -0.00 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33

Tgt Ion: 93 Resp: 513671  
Ion Ratio Lower Upper  
93 100  
95 32.8 26.0 39.0  
123 11.1 9.5 14.3





#29

2,4-Dichlorophenol

Concen: 39.27 ng

RT: 7.72 min Scan# 537

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion:162 Resp: 343342

Ion Ratio Lower Upper

162 100

164 61.5 42.2 82.2




98 51.9 19.0 59.0



Abundance Ion 162.00 (161.70 to 162.70): E

Ion 164.00 (163.70 to 164.70): E

Ion 98.00 (97.70 to 98.70): BF0



#30

1,2,4-Trichlorobenzene

Concen: 38.24 ng

RT: 7.79 min Scan# 543

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion:180 Resp: 370920

Ion Ratio Lower Upper

180 100

182 93.3 78.0 117.0



145 33.9 24.2 36.2




Abundance Ion 180.00 (179.70 to 180.70): E

Ion 182.00 (181.70 to 182.70): E

Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 39.87 ng

RT: 7.87 min Scan# 550

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion:128 Resp: 1247994

Ion Ratio Lower Upper

128 100

129 11.2 8.6 13.0




127 13.8 10.8 16.2



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 26.12 ng

RT: 7.71 min Scan# 536

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion:122 Resp: 154481

Ion Ratio Lower Upper

122 100

105 127.1 92.6 132.6



77 104.1 60.0 100.0#




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0





#33

4-Chloroaniline

Concen: 35.36 ng

RT: 7.94 min Scan# 556

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040

Tgt Ion:127 Resp: 460060


Ion Ratio Lower Upper

127 100

129 32.7 26.2 39.4

65 31.6 23.0 34.4

92 19.5 15.1 22.7






Abundance Ion 127.00 (126.70 to 127.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 65.00 (64.70 to 65.70): BF0

Ion 92.00 (91.70 to 92.70): BF0



#34

Hexachlorobutadiene

Concen: 42.71 ng

RT: 7.99 min Scan# 560

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion:225 Resp: 235129

Ion Ratio Lower Upper

225 100

223 63.4 51.5 77.3



227 64.5 52.2 78.2




Abundance Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E





#35

Caprolactam

Concen: 37.70 ng

RT: 8.34 min Scan# 591

Delta R.T. 0.01 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion:113 Resp: 109744

Ion Ratio Lower Upper

113 100

55 224.2 162.0 202.0#




56 167.0 115.3 155.3#



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BF0

Ion 56.00 (55.70 to 56.70): BF0



#36

4-Chloro-3-methylphenol

Concen: 37.07 ng

RT: 8.44 min Scan# 600

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion:107 Resp: 390886

Ion Ratio Lower Upper

107 100

144 23.9 20.7 31.1



142 73.7 63.2 94.8




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E





#37

2-Methylnaphthalene

Concen: 36.06 ng

RT: 8.66 min Scan# 619

Delta R.T. 0.10 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion:142 Resp: 721958

Ion Ratio Lower Upper

142 100

141 92.9 71.0 106.6




115 36.5 26.6 39.8



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 9.60 min Scan# 701

Delta R.T. 0.01 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion:164 Resp: 293690

Ion Ratio Lower Upper

164 100

162 105.3 82.8 124.2



160 48.1 36.9 55.3




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#39  
 1,2,4,5-Tetrachlorobenzene  
 Concen: 42.89 ng  
 RT: 8.73 min Scan# 625  
 Delta R.T. -0.00 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33

Instrument :  
 BNA\_F  
 ClientSampleId :  
 SSTDCCC040

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 216   | Resp: | 367897 |
| Ion      | Ratio | Lower | Upper  |
| 216      | 100   |       |        |
| 214      | 78.2  | 62.6  | 93.8   |
| 179      | 23.5  | 18.2  | 27.4   |
| 108      | 21.0  | 17.5  | 26.3   |




Abundance Ion 216.00 (215.70 to 216.70): E  
 Ion 214.00 (213.70 to 214.70): E  
 Ion 179.00 (178.70 to 179.70): E  
 Ion 108.00 (107.70 to 108.70): E





#40  
 Hexachlorocyclopentadiene  
 Concen: 29.27 ng  
 RT: 8.71 min Scan# 623  
 Delta R.T. -0.00 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 237   | Resp: | 78659 |
| Ion      | Ratio | Lower | Upper |
| 237      | 100   |       |       |
| 235      | 64.3  | 47.2  | 87.2  |
| 272      | 11.8  | 0.0   | 31.1  |



Abundance Ion 236.80 (236.50 to 237.50): E  
 Ion 234.90 (234.60 to 235.60): E  
 Ion 271.75 (271.45 to 272.45): E









#41  
 2,4,6-Tribromophenol  
 Concen: 64.67 ng  
 RT: 10.39 min Scan# 770  
 Delta R.T. -0.00 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33

Instrument :  
 BNA\_F  
 ClientSampleId :  
 SSTDCCC040

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 330     | 100   |       |       |
| 332     | 95.8  | 79.0  | 118.4 |
| 141     | 39.8  | 31.2  | 46.8  |




Abundance Ion 329.80 (329.50 to 330.50): E  
 Ion 331.80 (331.50 to 332.50): E  
 Ion 141.00 (140.70 to 141.70): E






#42  
 2,4,6-Trichlorophenol  
 Concen: 38.93 ng  
 RT: 8.86 min Scan# 636  
 Delta R.T. -0.00 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 196     | 100   |       |       |
| 198     | 100.0 | 74.7  | 112.1 |
| 200     | 32.7  | 23.8  | 35.6  |



Abundance Ion 196.00 (195.70 to 196.70): E  
 Ion 198.00 (197.70 to 198.70): E  
 Ion 200.00 (199.70 to 200.70): E









#43  
 2,4,5-Trichlorophenol  
 Concen: 38.69 ng  
 RT: 8.90 min Scan# 640  
 Delta R.T. -0.00 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33

Instrument :  
 BNA\_F  
 ClientSampleId :  
 SSTDCCC040

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 196     | 100   |       |       |
| 198     | 100.5 | 77.5  | 116.3 |
| 97      | 67.2  | 34.8  | 52.2  |
| 132     | 35.0  | 23.0  | 34.4  |




Abundance Ion 196.00 (195.70 to 196.70): E  
 Ion 198.00 (197.70 to 198.70): E  
 Ion 97.00 (96.70 to 97.70): BF0  
 Ion 132.00 (131.70 to 132.70): E






#44  
 2-Fluorobiphenyl  
 Concen: 71.68 ng  
 RT: 8.92 min Scan# 642  
 Delta R.T. -0.00 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 172     | 100   |       |       |
| 171     | 36.9  | 29.0  | 43.6  |
| 170     | 24.8  | 19.8  | 29.8  |



Abundance Ion 172.00 (171.70 to 172.70): E  
 Ion 171.00 (170.70 to 171.70): E  
 Ion 170.00 (169.70 to 170.70): E





#45

1,1'-Biphenyl

Concen: 43.96 ng

RT: 9.03 min Scan# 651

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion:154 Resp: 1070492

Ion Ratio Lower Upper

154 100

153 41.3 20.3 60.3




76 11.5 0.0 30.2



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BF0



#46

2-Chloronaphthalene

Concen: 40.81 ng

RT: 9.05 min Scan# 653

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion:162 Resp: 762865

Ion Ratio Lower Upper

162 100

127 41.5 31.8 47.6



164 33.2 27.0 40.6




Abundance Ion 162.00 (161.70 to 162.70): E

Ion 127.00 (126.70 to 127.70): E

Ion 164.00 (163.70 to 164.70): E









#47  
2-Nitroaniline  
Concen: 40.21 ng  
RT: 9.16 min Scan# 663  
Delta R.T. -0.00 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33




Instrument :  
BNA\_F  
ClientSampleId :  
SSTDCCC040


Tgt Ion: 65 Resp: 288055  
Ion Ratio Lower Upper  
65 100  
92 61.7 57.4 86.2  
138 87.9 90.1 135.1#



#48  
Acenaphthylene  
Concen: 38.30 ng  
RT: 9.46 min Scan# 689  
Delta R.T. 0.01 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33

Tgt Ion: 152 Resp: 1093216  
Ion Ratio Lower Upper  
152 100  
151 20.5 16.8 25.2  
153 13.4 10.9 16.3





#49

Dimethylphthalate

Concen: 37.62 ng

RT: 9.34 min Scan# 678

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion:163 Resp: 842843

Ion Ratio Lower Upper

163 100

194 3.0 2.6 4.0




164 10.7 8.2 12.4



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 34.39 ng

RT: 9.40 min Scan# 684

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion:165 Resp: 167314

Ion Ratio Lower Upper

165 100

63 100.0 51.8 77.8#



89 78.1 50.7 76.1#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0





#51

Acenaphthene

Concen: 37.43 ng

RT: 9.63 min Scan# 704

Delta R.T. 0.01 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion:154 Resp: 667524

Ion Ratio Lower Upper

154 100

153 113.9 89.8 134.6




152 51.1 43.4 65.0



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 33.88 ng

RT: 9.58 min Scan# 699

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion:138 Resp: 178397

Ion Ratio Lower Upper

138 100

108 11.8 7.8 11.6#



92 140.2 87.8 131.8#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53  
 2,4-Dinitrophenol  
 Concen: 24.85 ng  
 RT: 9.70 min Scan# 710  
 Delta R.T. 0.01 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33




Instrument :  
 BNA\_F  
 ClientSampleId :  
 SSTDCCC040




Tgt Ion:184 Resp: 59582  
 Ion Ratio Lower Upper  
 184 100  
 63 93.4 55.8 83.8#  
 154 71.3 62.4 93.6



#54  
 Dibenzofuran  
 Concen: 39.23 ng  
 RT: 9.80 min Scan# 719  
 Delta R.T. 0.01 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33

Tgt Ion:168 Resp: 904809  
 Ion Ratio Lower Upper  
 168 100  
 139 38.7 31.4 47.0  
 169 12.9 10.7 16.1





#55

4-Nitrophenol

Concen: 100.28 ng

RT: 9.80 min Scan# 719

Delta R.T. 0.03 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040





Tgt Ion:139 Resp: 460404

Ion Ratio Lower Upper

139 100

109 5.0 36.9 76.9#

65 10.4 64.0 104.0#



#56

2,4-Dinitrotoluene

Concen: 40.00 ng

RT: 9.82 min Scan# 720

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Tgt Ion:165 Resp: 249226


Ion Ratio Lower Upper


165 100

63 45.6 44.3 66.5

89 67.6 70.0 105.0#

182 2.4 1.8 2.6





#57

Fluorene

Concen: 36.80 ng

RT: 10.14 min Scan# 748

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion:166 Resp: 681852

Ion Ratio Lower Upper

166 100

165 99.2 79.0 118.6




167 14.4 10.6 15.8



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#58

2,3,4,6-Tetrachlorophenol

Concen: 36.96 ng

RT: 9.93 min Scan# 730

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Tgt Ion:232 Resp: 166349


Ion Ratio Lower Upper

232 100

131 52.8 41.9 62.9

130 2.6 2.2 3.4

166 29.6 26.8 40.2






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E









#59  
Diethylphthalate  
Concen: 39.64 ng  
RT: 10.03 min Scan# 739  
Delta R.T. -0.00 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33




Instrument :  
BNA\_F  
ClientSampleId :  
SSTDCCC040


Tgt Ion:149 Resp: 847870  
Ion Ratio Lower Upper  
149 100  
177 21.4 16.6 24.8  
150 13.0 10.0 15.0



#60  
4-Chlorophenyl-phenylether  
Concen: 42.27 ng  
RT: 10.14 min Scan# 748  
Delta R.T. -0.00 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33

Tgt Ion:204 Resp: 371304  
Ion Ratio Lower Upper  
204 100  
206 32.5 25.4 38.0  
141 70.2 61.6 92.4





#61

4-Nitroaniline

Concen: 31.41 ng

RT: 10.19 min Scan# 753

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion: 138 Resp: 163342

Ion Ratio Lower Upper

138 100

92 59.7 24.7 64.7




108 93.4 37.4 77.4#



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF0

Ion 108.00 (107.70 to 108.70): E



#62

Azobenzene

Concen: 40.20 ng

RT: 10.30 min Scan# 762

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Tgt Ion: 77 Resp: 990771


Ion Ratio Lower Upper

77 100

182 20.9 0.0 38.8

105 20.6 3.2 43.2

51 42.3 8.8 48.8






Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BF0









#63  
Phenanthrene-d10  
Concen: 20.00 ng  
RT: 11.07 min Scan# 830  
Delta R.T. -0.00 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33




Instrument :  
BNA\_F  
ClientSampleId :  
SSTDCCC040


Tgt Ion:188 Resp: 520857  
Ion Ratio Lower Upper  
188 100  
94 7.5 6.8 10.2  
80 7.9 7.1 10.7



#64  
4,6-Dinitro-2-methylphenol  
Concen: 29.76 ng  
RT: 10.23 min Scan# 756  
Delta R.T. -0.00 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33

Tgt Ion:198 Resp: 102098  
Ion Ratio Lower Upper  
198 100  
51 50.1 20.5 60.5  
105 38.3 24.5 64.5





#65

n-Nitrosodiphenylamine

Concen: 38.48 ng

RT: 10.26 min Scan# 759

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion:169 Resp: 630663

Ion Ratio Lower Upper

169 100

168 68.3 53.8 80.6




167 37.9 29.5 44.3



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 39.83 ng

RT: 10.62 min Scan# 790

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion:248 Resp: 213352

Ion Ratio Lower Upper

248 100

250 97.9 78.6 117.8



141 128.5 76.0 114.0#




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E









#67  
Hexachlorobenzene  
Concen: 35.83 ng  
RT: 10.68 min Scan# 796  
Delta R.T. -0.00 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33




Instrument :  
BNA\_F  
ClientSampleId :  
SSTDCCC040


Tgt Ion: 284 Resp: 222492  
Ion Ratio Lower Upper  
284 100  
142 40.6 16.7 25.1#  
249 33.9 21.3 31.9#



#68  
Atrazine  
Concen: 39.61 ng  
RT: 10.80 min Scan# 806  
Delta R.T. -0.00 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33

Tgt Ion: 200 Resp: 211795  
Ion Ratio Lower Upper  
200 100  
173 25.7 8.5 48.5  
215 42.5 27.2 67.2









#69  
 Pentachlorophenol  
 Concen: 24.42 ng  
 RT: 10.89 min Scan# 814  
 Delta R.T. -0.00 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33




Instrument :  
 BNA\_F  
 ClientSampleId :  
 SSTDCCC040


Tgt Ion:266 Resp: 56961  
 Ion Ratio Lower Upper  
 266 100  
 268 62.2 51.5 77.3  
 264 62.3 52.9 79.3



#70  
 Phenanthrene  
 Concen: 38.88 ng  
 RT: 11.10 min Scan# 832  
 Delta R.T. -0.00 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33

Tgt Ion:178 Resp: 1098348  
 Ion Ratio Lower Upper  
 178 100  
 176 20.1 16.0 24.0  
 179 15.6 12.6 18.8





#71

Anthracene

Concen: 37.15 ng

RT: 11.14 min Scan# 836

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion:178 Resp: 1061383

Ion Ratio Lower Upper

178 100

176 19.2 15.6 23.4




179 16.2 12.7 19.1



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Carbazole

Concen: 35.47 ng

RT: 11.31 min Scan# 851

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion:167 Resp: 925731

Ion Ratio Lower Upper

167 100

166 22.2 17.7 26.5



139 14.9 10.8 16.2




Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E





#73

Di-n-butylphthalate

Concen: 43.94 ng

RT: 11.64 min Scan# 880

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 1317200

Ion Ratio Lower Upper

149 100

150 9.7 7.6 11.4




104 5.9 3.8 5.6#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 34.54 ng

RT: 12.27 min Scan# 935

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion:202 Resp: 977701

Ion Ratio Lower Upper

202 100

101 12.7 0.0 35.4



203 18.2 0.0 38.1




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 13.70 min Scan# 1060

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion:240 Resp: 322837

Ion Ratio Lower Upper

240 100

120 7.6 11.2 16.8#



236 26.6 21.2 31.8



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 32.65 ng

RT: 12.41 min Scan# 947

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion:184 Resp: 315235

Ion Ratio Lower Upper

184 100

185 15.8 13.6 20.4



183 11.5 9.8 14.6




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E









#77  
 Pyrene  
 Concen: 44.78 ng  
 RT: 12.50 min Scan# 955  
 Delta R.T. -0.00 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33

Instrument :  
 BNA\_F  
 ClientSampleId :  
 SSTDCCC040

Tgt Ion:202 Resp: 1044360  
 Ion Ratio Lower Upper  
 202 100  
 200 22.2 17.7 26.5  
 203 17.6 14.2 21.4




Abundance Ion 202.00 (201.70 to 202.70): E  
 Ion 200.00 (199.70 to 200.70): E  
 Ion 203.00 (202.70 to 203.70): E





#78  
 Terphenyl-d14  
 Concen: 89.16 ng  
 RT: 12.65 min Scan# 968  
 Delta R.T. -0.00 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33

Tgt Ion:244 Resp: 1272403  
 Ion Ratio Lower Upper  
 244 100  
 212 7.8 6.2 9.2  
 122 12.0 12.0 18.0



Abundance Ion 244.00 (243.70 to 244.70): E  
 Ion 212.00 (211.70 to 212.70): E  
 Ion 122.00 (121.70 to 122.70): E





#79

Butylbenzylphthalate

Concen: 50.67 ng

RT: 13.13 min Scan# 1010

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 498918

Ion Ratio Lower Upper

149 100

91 72.2 48.0 72.0#




206 19.3 15.8 23.6



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF0

Ion 206.00 (205.70 to 206.70): E



#80

Benzo(a)anthracene

Concen: 38.37 ng

RT: 13.69 min Scan# 1059

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion:228 Resp: 716325

Ion Ratio Lower Upper

228 100

226 27.7 22.5 33.7



229 19.7 16.6 25.0




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#81  
 3,3'-Dichlorobenzidine  
 Concen: 39.17 ng  
 RT: 13.66 min Scan# 1056  
 Delta R.T. -0.00 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33




Instrument :  
 BNA\_F  
 ClientSampleId :  
 SSTDCCC040

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 252     | 100   |       |       |
| 254     | 62.8  | 50.7  | 76.1  |
| 126     | 15.8  | 12.7  | 19.1  |



#82  
 Chrysene  
 Concen: 41.50 ng  
 RT: 13.72 min Scan# 1062  
 Delta R.T. -0.00 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 228     | 100   |       |       |
| 226     | 31.1  | 24.4  | 36.6  |
| 229     | 19.5  | 15.8  | 23.8  |





#83

Bis(2-ethylhexyl)phthalate

Concen: 52.22 ng

RT: 13.69 min Scan# 1059

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion:149 Resp: 625790

Ion Ratio Lower Upper

149 100

167 26.8 21.8 32.6




279 4.0 2.6 4.0#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#84

Di-n-octyl phthalate

Concen: 47.45 ng

RT: 14.30 min Scan# 1112

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion:149 Resp: 979377

Ion Ratio Lower Upper

149 100

167 1.5 1.2 1.8



43 17.1 8.2 12.4#




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BF0









#85  
 Indeno(1,2,3-cd)pyrene  
 Concen: 42.27 ng  
 RT: 16.27 min Scan# 1285  
 Delta R.T. -0.00 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33




Instrument :  
 BNA\_F  
 ClientSampleId :  
 SSTDCCC040


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 276   | Resp: | 670101 |
| Ion      | Ratio | Lower | Upper  |
| 276      | 100   |       |        |
| 138      | 27.0  | 25.6  | 38.4   |
| 277      | 25.1  | 20.5  | 30.7   |



#86  
 Perylene-d12  
 Concen: 20.00 ng  
 RT: 15.05 min Scan# 1178  
 Delta R.T. 0.01 min  
 Lab File: BF087279.D  
 Acq: 22 May 2016 9:33

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 264   | Resp: | 289201 |
| Ion      | Ratio | Lower | Upper  |
| 264      | 100   |       |        |
| 260      | 24.5  | 19.5  | 29.3   |
| 265      | 21.7  | 17.3  | 25.9   |





#87

Benzo(b)fluoranthene

Concen: 67.04 ng

RT: 14.69 min Scan# 1146

Delta R.T. -0.00 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040


Tgt Ion:252 Resp: 1286142

Ion Ratio Lower Upper

252 100

253 22.5 17.8 26.6




125 13.1 11.4 17.0



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 90.07 ng

RT: 14.69 min Scan# 1146

Delta R.T. -0.02 min

Lab File: BF087279.D

Acq: 22 May 2016 9:33


Tgt Ion:252 Resp: 1286536

Ion Ratio Lower Upper

252 100

253 22.5 17.8 26.6



125 13.1 9.1 13.7




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#89  
Benzo(a)pyrene  
Concen: 38.24 ng  
RT: 14.99 min Scan# 1173  
Delta R.T. -0.00 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33




Instrument :  
BNA\_F  
ClientSampleId :  
SSTDCCC040


Tgt Ion: 252 Resp: 613319  
Ion Ratio Lower Upper  
252 100  
253 22.3 17.2 25.8  
125 14.8 9.0 13.6#



#90  
Dibenzo(a,h)anthracene  
Concen: 41.52 ng  
RT: 16.29 min Scan# 1286  
Delta R.T. 0.01 min  
Lab File: BF087279.D  
Acq: 22 May 2016 9:33

Tgt Ion: 278 Resp: 560621  
Ion Ratio Lower Upper  
278 100  
139 14.1 16.6 24.8#  
279 23.9 19.6 29.4





#91

Benzo(g,h,i)perylene

Concen: 40.48 ng

RT: 16.64 min Scan# 1317

Delta R.T. 0.01 min

Lab File: BF087279.D


Acq: 22 May 2016 9:33

Instrument :

BNA\_F

ClientSampleId :

SSTDCCC040



Tgt Ion: 276 Resp: 552035

Ion Ratio Lower Upper

276 100

277 23.9 19.6 29.4

138 23.3 23.6 35.4#

Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

