

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052216\
 Data File : BF087279.D
 Acq On : 22 May 2016 9:33
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 23 03:42:50 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF051716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 23 01:56:40 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	-1.07#
2	1,4-Dioxane	0.612	0.594	2.9	77	0.03
3	Pyridine	1.481	1.328	10.3	67	0.02
4	n-Nitrosodimethylamine	1.049	1.047	0.2	73	0.03
5 S	2-Fluorophenol	1.190	1.201	-0.9	81	-0.40
6	Aniline	2.027	1.862	8.1	73	-0.90#
7 S	Phenol-d6	1.596	1.546	3.1	77	-0.94#
8	2-Chlorophenol	1.448	1.379	4.8	75	-0.96#
9	Benzaldehyde	0.908	0.798	12.1	78	-0.83#
10 C	Phenol	2.006	1.915	4.5	73	-0.95#
11	bis(2-Chloroethyl)ether	1.506	1.481	1.7	88	-0.97#
12	1,3-Dichlorobenzene	1.539	1.497	2.7	76	-1.04#
13 C	1,4-Dichlorobenzene	1.575	1.515	3.8	77	-1.10#
14	1,2-Dichlorobenzene	1.378	1.372	0.4	86	-1.17#
15	Benzyl Alcohol	1.188	1.230	-3.5	79	-1.20#
16	2,2'-oxybis(1-Chloropropane	3.097	3.001	3.1	81	-1.27#
17	2-Methylphenol	1.189	1.120	5.8	74	-1.28#
18	Hexachloroethane	0.602	0.599	0.5	78	-1.35#
19 P	n-Nitroso-di-n-propylamine	1.213	1.153	4.9	73	-1.35#
20	3+4-Methylphenols	1.568	1.474	6.0	72	-1.38#
21 I	Naphthalene-d8	1.000	1.000	0.0	80	-1.82#
22	Acetophenone	0.489	0.514	-5.1	83	-1.33#
23 S	Nitrobenzene-d5	0.401	0.363	9.5	75	-1.42#
24	Nitrobenzene	0.438	0.411	6.2	72	-1.42#
25	Isophorone	0.735	0.705	4.1	78	-1.58#
26 C	2-Nitrophenol	0.182	0.168	7.7	76	-1.61#
27	2,4-Dimethylphenol	0.310	0.280	9.7	74	-1.68#
28	bis(2-Chloroethoxy)methane	0.403	0.401	0.5	75	-1.73#
29 C	2,4-Dichlorophenol	0.273	0.268	1.8	81	-1.77#
30	1,2,4-Trichlorobenzene	0.303	0.290	4.3	82	-1.79#
31	Naphthalene	0.977	0.974	0.3	83	-1.83#
32	Benzoic acid	0.185	0.121	34.6#	51	-1.83#
33	4-Chloroaniline	0.406	0.359	11.6	69	-1.90#
34 C	Hexachlorobutadiene	0.172	0.184	-7.0	92	-1.92#
35	Caprolactam	0.091	0.086	5.5	75	-2.15#
36 C	4-Chloro-3-methylphenol	0.329	0.305	7.3	77	-2.25#
37	2-Methylnaphthalene	0.625	0.571	8.6	77	-2.26#
38 I	Acenaphthene-d10	1.000	1.000	0.0	79	-2.90#
39	1,2,4,5-Tetrachlorobenzene	0.584	0.626	-7.2	92	-2.37#
40 P	Hexachlorocyclopentadiene	0.182	0.134	26.4#	49#	-2.37#
41 S	2,4,6-Tribromophenol	0.221	0.179	19.0	59	-3.41#
42 C	2,4,6-Trichlorophenol	0.379	0.369	2.6	76	-2.47#
43	2,4,5-Trichlorophenol	0.394	0.381	3.3	74	-2.50#
44 S	2-Fluorobiphenyl	1.208	1.082	10.4	68	-2.53#

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052216\
 Data File : BF087279.D
 Acq On : 22 May 2016 9:33
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 23 03:42:50 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF051716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 23 01:56:40 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.658	1.822	-9.9	92	-2.57#
46	2-Chloronaphthalene	1.273	1.299	-2.0	86	-2.56#
47	2-Nitroaniline	0.488	0.490	-0.4	76	-2.66#
48	Acenaphthylene	1.944	1.861	4.3	82	-2.81#
49	Dimethylphthalate	1.526	1.435	6.0	75	-2.81#
50	2,6-Dinitrotoluene	0.331	0.285	13.9	63	-2.83#
51 C	Acenaphthene	1.215	1.136	6.5	83	-2.91#
52	3-Nitroaniline	0.359	0.304	15.3	64	-2.94#
53 P	2,4-Dinitrophenol	0.166	0.101	39.2#	41#	-3.01#
54	Dibenzofuran	1.571	1.540	2.0	85	-3.03#
55 P	4-Nitrophenol	0.155	0.050#	67.7#	21#	-3.11#
56	2,4-Dinitrotoluene	0.424	0.424	0.0	75	-3.09#
57	Fluorene	1.262	1.161	8.0	80	-3.26#
58	2,3,4,6-Tetrachlorophenol	0.307	0.283	7.8	71	-3.14#
59	Diethylphthalate	1.457	1.443	1.0	73	-3.26#
60	4-Chlorophenyl-phenylether	0.598	0.632	-5.7	77	-3.28#
61	4-Nitroaniline	0.354	0.278	21.5	54	-3.31#
62	Azobenzene	1.678	1.687	-0.5	77	-3.38#
63 I	Phenanthrene-d10	1.000	1.000	0.0	74	-3.83#
64	4,6-Dinitro-2-methylphenol	0.132	0.098	25.8#	51	-3.34#
65 c	n-Nitrosodiphenylamine	0.629	0.605	3.8	68	-3.37#
66	4-Bromophenyl-phenylether	0.206	0.205	0.5	81	-3.59#
67	Hexachlorobenzene	0.238	0.214	10.1	63	-3.59#
68	Atrazine	0.205	0.203	1.0	71	-3.75#
69 C	Pentachlorophenol	0.090	0.055	38.9#	41#	-3.75#
70	Phenanthrene	1.085	1.054	2.9	70	-3.84#
71	Anthracene	1.097	1.019	7.1	78	-3.87#
72	Carbazole	1.002	0.889	11.3	64	-3.99#
73	Di-n-butylphthalate	1.151	1.264	-9.8	77	-4.23#
74 C	Fluoranthene	1.087	0.939	13.6	66	-4.41#
75 I	Chrysene-d12	1.000	1.000	0.0	60	-4.88#
76	Benzidine	0.526	0.488	7.2	52	-4.49#
77	Pyrene	1.445	1.617	-11.9	66	-4.48#
78 S	Terphenyl-d14	0.884	0.985	-11.4	75	-4.58#
79	Butylbenzylphthalate	0.610	0.773	-26.7#	69	-4.75#
80	Benzo(a)anthracene	1.156	1.109	4.1	61	-4.87#
81	3,3'-Dichlorobenzidine	0.736	0.721	2.0	61	-4.89#
82	Chrysene	1.119	1.161	-3.8	63	-4.88#
83	Bis(2-ethylhexyl)phthalate	0.742	0.969	-30.6#	77	-4.94#
84 c	Di-n-octyl phthalate	1.279	1.517	-18.6	73	-5.11#
85	Indeno(1,2,3-cd)pyrene	0.982	1.038	-5.7	64	-5.24#
86 I	Perylene-d12	1.000	1.000	0.0	63	-5.19#
87	Benzo(b)fluoranthene	1.327	1.049	20.9	58	-5.14#

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052216\
Data File : BF087279.D
Acq On : 22 May 2016 9:33
Operator : UM/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 23 03:42:50 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF051716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 23 01:56:40 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	0.988	1.152	-16.6	63	-5.16#
89 C	Benzo(a)pyrene	1.109	1.060	4.4	60	-5.19#
90	Dibenzo(a,h)anthracene	0.934	0.969	-3.7	65	-5.24#
91	Benzo(a,h,i)perylene	0.943	0.954	-1.2	62	-5.24#

(#) = Out of Range

SPCC's out = 1 CCC's out = 1