

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF052317\
 Data File : BF095349.D
 Acq On : 23 May 2017 14:05
 Operator : SJ/MA
 Sample : I3240-08
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EP2-WC5(10-24)

Quant Time: May 24 02:16:17 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 18 17:07:53 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	85741	20.00	ng	-0.01
21) Naphthalene-d8	9.78	136	380760	20.00	ng	0.00
38) Acenaphthene-d10	12.61	164	187832	20.00	ng	-0.01
63) Phenanthrene-d10	15.01	188	324811	20.00	ng	-0.01
75) Chrysene-d12	18.68	240	266489	20.00	ng	-0.01
86) Perylene-d12	20.37	264	165411	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.72	112	624067	121.35	ng	-0.01
7) Phenol-d6	7.30	99	701018	113.61	ng	-0.02
23) Nitrobenzene-d5	8.65	82	438748	72.66	ng	-0.02
41) 2,4,6-Tribromophenol	13.90	330	182287	118.50	ng	-0.02
44) 2-Fluorobiphenyl	11.54	172	952367	72.98	ng	-0.02
78) Terphenyl-d14	17.33	244	739229	61.10	ng	-0.02
Target Compounds						
49) Dimethylphthalate	12.25	163	118014	7.65	ng	Qvalue 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF052317\
Data File : BF095349.D
Acq On : 23 May 2017 14:05
Operator : SJ/MA
Sample : I3240-08
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EP2-WC5(10-24)

Quant Time: May 24 02:16:17 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 18 17:07:53 2017
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.75 min Scan# 989
Delta R.T. -0.01 min
Lab File: BF095349.D
Acq: 23 May 2017 14:05

Instrument :
BNA_F
ClientSampleId :
EP2-WC5(10-24)

Tgt Ion:152 Resp: 85741
Ion Ratio Lower Upper
152 100
150 159.5 126.2 189.2
115 60.4 52.2 78.2




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 121.35 ng
RT: 5.72 min Scan# 644
Delta R.T. -0.01 min
Lab File: BF095349.D
Acq: 23 May 2017 14:05

Tgt Ion:112 Resp: 624067
Ion Ratio Lower Upper
112 100
64 54.4 46.0 69.0
63 27.2 23.4 35.0



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF0
Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 113.61 ng

RT: 7.30 min Scan# 911

Delta R.T. -0.02 min

Lab File: BF095349.D

Acq: 23 May 2017 14:05

Instrument :

BNA_F

ClientSampleId :

EP2-WC5(10-24)


Tgt Ion: 99 Resp: 701018

Ion Ratio Lower Upper

99 100

42 16.3 13.5 20.3




71 31.6 24.5 36.7



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 9.78 min Scan# 1333

Delta R.T. -0.01 min

Lab File: BF095349.D

Acq: 23 May 2017 14:05

Tgt Ion: 136 Resp: 380760


Ion Ratio Lower Upper

136 100

137 11.0 8.5 12.7

54 6.5 4.9 7.3

68 4.7 4.1 6.1






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0









#23
 Nitrobenzene-d5
 Concen: 72.66 ng
 RT: 8.65 min Scan# 1142
 Delta R.T. -0.02 min
 Lab File: BF095349.D
 Acq: 23 May 2017 14:05




Instrument :
 BNA_F
 ClientSampleId :
 EP2-WC5(10-24)


Tgt Ion: 82 Resp: 438748
 Ion Ratio Lower Upper
 82 100
 128 46.6 34.0 51.0
 54 46.7 42.2 63.2



#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 12.61 min Scan# 1814
 Delta R.T. -0.01 min
 Lab File: BF095349.D
 Acq: 23 May 2017 14:05

Tgt Ion: 164 Resp: 187832
 Ion Ratio Lower Upper
 164 100
 162 105.9 82.6 123.8
 160 44.7 36.2 54.2









#41
 2,4,6-Tribromophenol
 Concen: 118.50 ng
 RT: 13.90 min Scan# 2034
 Delta R.T. -0.02 min
 Lab File: BF095349.D
 Acq: 23 May 2017 14:05

Instrument :
 BNA_F
 ClientSampleId :
 EP2-WC5(10-24)

Tgt Ion: 330 Resp: 182287
 Ion Ratio Lower Upper
 330 100
 332 96.6 76.6 115.0
 141 40.1 31.4 47.2




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 72.98 ng
 RT: 11.54 min Scan# 1633
 Delta R.T. -0.02 min
 Lab File: BF095349.D
 Acq: 23 May 2017 14:05

Tgt Ion: 172 Resp: 952367
 Ion Ratio Lower Upper
 172 100
 171 36.4 29.6 44.4
 170 23.4 19.8 29.8



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#49

Dimethylphthalate

Concen: 7.65 ng

RT: 12.25 min Scan# 1753

Delta R.T. -0.01 min

Lab File: BF095349.D

Acq: 23 May 2017 14:05

Instrument :

BNA_F

ClientSampleId :

EP2-WC5(10-24)


Tgt Ion:163 Resp: 118014

Ion Ratio Lower Upper

163 100

194 3.6 2.6 4.0




164 10.0 8.4 12.6



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 15.01 min Scan# 2223

Delta R.T. -0.01 min

Lab File: BF095349.D

Acq: 23 May 2017 14:05


Tgt Ion:188 Resp: 324811

Ion Ratio Lower Upper

188 100

94 9.9 9.4 14.2



80 10.5 10.2 15.2




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF0

Ion 80.00 (79.70 to 80.70): BF0





#75

Chrysene-d12

Concen: 20.00 ng

RT: 18.68 min Scan# 2847

Delta R.T. -0.01 min

Lab File: BF095349.D

Acq: 23 May 2017 14:05

Instrument :

BNA_F

ClientSampleId :

EP2-WC5(10-24)


Tgt Ion:240 Resp: 266489

Ion Ratio Lower Upper

240 100

120 14.1 5.8 8.8#




236 25.4 20.5 30.7



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#78

Terphenyl-d14

Concen: 61.10 ng

RT: 17.33 min Scan# 2617

Delta R.T. -0.02 min

Lab File: BF095349.D

Acq: 23 May 2017 14:05


Tgt Ion:244 Resp: 739229

Ion Ratio Lower Upper

244 100

212 7.0 6.0 9.0



122 13.8 10.3 15.5




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#86
Perylene-d12
Concen: 20.00 ng
RT: 20.37 min Scan# 3133
Delta R.T. 0.00 min
Lab File: BF095349.D
Acq: 23 May 2017 14:05

Instrument :
BNA_F
ClientSampleId :
EP2-WC5(10-24)

Tgt Ion: 264 Resp: 165411
Ion Ratio Lower Upper
264 100
260 23.9 19.5 29.3
265 21.6 17.2 25.8

