


Data Path : Z:\HPCHEM1\BNA_F\Data\BF052317\
 Data File : BF095347.D
 Acq On : 23 May 2017 12:23
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu May 18 17:07:53 2017



AutoFind: Scans 2312, 2313, 2314; Background Corrected with Scan 2307

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.9	14629	PASS
68	69	0.00	2	0.9	127	PASS
69	198	0.00	100	33.0	14266	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.1	20356	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	43176	PASS
199	198	5	9	6.3	2719	PASS
275	198	10	60	23.6	10171	PASS
365	198	1	100	2.2	953	PASS
441	442	0.01	24	15.6	5442	PASS
442	198	50	100	81.0	34984	PASS
443	442	15	24	19.8	6928	PASS