

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052423\
 Data File : BF133478.D
 Acq On : 24 May 2023 19:07
 Operator : CG\JU
 Sample : 02892-01DL 5X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-30DL

Quant Time: May 25 05:21:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 24 02:41:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.839	152	122754	20.000	ng	0.00
21) Naphthalene-d8	8.122	136	421777	20.000	ng	# 0.00
39) Acenaphthene-d10	9.880	164	183792	20.000	ng	0.00
64) Phenanthrene-d10	11.369	188	368123	20.000	ng	0.00
76) Chrysene-d12	14.004	240	304668	20.000	ng	0.00
86) Perylene-d12	15.462	264	268485	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.434	112	135022	17.980	ng	-0.01
7) Phenol-d6	6.445	99	168415	18.656	ng	-0.02
23) Nitrobenzene-d5	7.398	82	95243	15.612	ng	-0.01
42) 2,4,6-Tribromophenol	10.669	330	31927	18.805	ng	0.00
45) 2-Fluorobiphenyl	9.198	172	175891	13.464	ng	0.00
79) Terphenyl-d14	12.951	244	213724	13.412	ng	0.00
Target Compounds						
						Qvalue
31) Naphthalene	8.145	128	49453	2.387	ng	# 74
37) 2-Methylnaphthalene	8.839	142	617132	44.434	ng	99
38) 1-Methylnaphthalene	8.939	142	534106	41.270	ng	99
52) Acenaphthene	9.916	154	46525	4.125	ng	# 77
58) Fluorene	10.433	166	84882	6.986	ng	# 90
71) Phenanthrene	11.392	178	282853	14.839	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052423\
Data File : BF133478.D
Acq On : 24 May 2023 19:07
Operator : CG\JU
Sample : 02892-01DL 5X
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-30DL

Quant Time: May 25 05:21:54 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 24 02:41:39 2023
Response via : Initial Calibration

