

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052423\
 Data File : BF133481.D
 Acq On : 24 May 2023 20:43
 Operator : CG\JU
 Sample : 02899-09 10X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC14

Quant Time: May 25 05:22:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 24 02:41:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.840	152	132228	20.000	ng	0.00
21) Naphthalene-d8	8.122	136	511853	20.000	ng	0.00
39) Acenaphthene-d10	9.881	164	244621	20.000	ng	0.00
64) Phenanthrene-d10	11.363	188	429806	20.000	ng	0.00
76) Chrysene-d12	14.004	240	284835	20.000	ng	0.00
86) Perylene-d12	15.463	264	205074	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.434	112	63544	7.855	ng	-0.01
7) Phenol-d6	6.451	99	78927	8.117	ng	-0.01
23) Nitrobenzene-d5	7.398	82	38030	5.137	ng	-0.01
42) 2,4,6-Tribromophenol	10.663	330	15144	6.702	ng	0.00
45) 2-Fluorobiphenyl	9.198	172	94202	5.418	ng	0.00
79) Terphenyl-d14	12.951	244	87896	5.900	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052423\
Data File : BF133481.D
Acq On : 24 May 2023 20:43
Operator : CG\JU
Sample : 02899-09 10X
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC14

Quant Time: May 25 05:22:47 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 24 02:41:39 2023
Response via : Initial Calibration

