

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052516\
 Data File : BF087394.D
 Acq On : 26 May 2016 4:11
 Operator : UM/SJ
 Sample : H3212-08
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 VE3-1

Quant Time: May 26 08:38:18 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF052316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 25 16:26:52 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.52	152	103041	20.00	ng	0.00
21) Naphthalene-d8	9.55	136	389347	20.00	ng	0.00
38) Acenaphthene-d10	12.39	164	205117	20.00	ng	0.00
63) Phenanthrene-d10	14.80	188	233284	20.00	ng	0.02
75) Chrysene-d12	18.50	240	212693	20.00	ng	0.02
86) Perylene-d12	20.16	264	143423	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	388710	66.29	ng	0.03
7) Phenol-d6	7.10	99	324383	40.94	ng	0.02
23) Nitrobenzene-d5	8.44	82	706752	91.48	ng	0.00
41) 2,4,6-Tribromophenol	13.71	330	170722	86.61	ng	0.03
44) 2-Fluorobiphenyl	11.35	172	1056824	72.64	ng	0.01
78) Terphenyl-d14	17.15	244	688386	68.81	ng	0.02
Target Compounds						Qvalue
17) 2-Methylphenol	8.07	107	35804	6.11	ng	# 79
20) 3+4-Methylphenols	8.33	107	94265	13.30	ng	# 57
27) 2,4-Dimethylphenol	9.14	122	52971	9.15	ng	# 84
37) 2-Methylnaphthalene	10.72	142	422582	34.67	ng	# 93
51) Acenaphthene	12.45	154	60338	4.69	ng	96
57) Fluorene	13.29	166	96664	6.40	ng	# 94
70) Phenanthrene	14.85	178	156939	12.15	ng	# 92
83) Bis(2-ethylhexyl)phthalate	18.56	149	95913	9.68	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052516\
Data File : BF087394.D
Acq On : 26 May 2016 4:11
Operator : UM/SJ
Sample : H3212-08
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
VE3-1

Quant Time: May 26 08:38:18 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF052316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 25 16:26:52 2016
Response via : Initial Calibration

