


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF052516\
 Data File : BF087366.D
 Acq On : 25 May 2016 10:39
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF052316.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 25 16:26:52 2016



AutoFind: Scans 1201, 1202, 1203; Background Corrected with Scan 1198

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	64.3	19646	PASS
68	69	0.00	2	1.3	182	PASS
69	198	0.00	100	47.0	14370	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	55.4	16927	PASS
197	198	0.00	2	0.4	130	PASS
198	198	100	100	100.0	30546	PASS
199	198	5	9	6.8	2088	PASS
275	198	10	60	24.3	7432	PASS
365	198	1	100	3.1	939	PASS
441	442	0.01	24	16.6	3423	PASS
442	198	50	100	67.7	20678	PASS
443	442	15	24	17.8	3682	PASS