

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052615\
 Data File : BF079505.D
 Acq On : 27 May 2015 6:12
 Operator : TP/IZ
 Sample : G2255-10DL 25X
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 KY032MW006-150512DL

Manual Integrations
 APPROVED

apatel
 5/27/2015 4:23:19 PM

Quant Time: May 27 11:35:20 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF052615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 27 01:32:46 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.02	152	71398	20.00	ng	0.00
21) Naphthalene-d8	8.60	136	285492	20.00	ng	0.01
38) Acenaphthene-d10	10.77	164	167232	20.00	ng	0.00
63) Phenanthrene-d10	12.61	188	332709	20.00	ng	0.01
75) Chrysene-d12	15.89	240	366773	20.00	ng	0.01
86) Perylene-d12	17.60	264	378203	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	0.00	112	0	0.00	ng	
7) Phenol-d6	6.63	99	618	0.12	ng	0.01
23) Nitrobenzene-d5	7.72	82	1618	0.33	ng	0.01
41) 2,4,6-Tribromophenol	11.75	330	263	0.14	ng	0.00
44) 2-Fluorobiphenyl	9.94	172	2632	0.22	ng	0.00
78) Terphenyl-d14	14.59	244	3830	0.25	ng	0.00
Target Compounds						
2) 1,4-Dioxane	1.84	88	111053m	56.71	ng	
10) Phenol	6.64	94	43927m	7.11	ng	
11) bis(2-Chloroethyl)ether	6.72	93	51030	11.42	ng	98
17) 2-Methylphenol	7.38	107	21987	5.84	ng	99
20) 3+4-Methylphenols	7.59	107	215164	41.59	ng	100
27) 2,4-Dimethylphenol	8.24	122	52721	12.73	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052615\
Data File : BF079505.D
Acq On : 27 May 2015 6:12
Operator : TP/IZ
Sample : G2255-10DL 25X
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
KY032MW006-150512DL

Manual Integrations
APPROVED

apatel
5/27/2015 4:23:19 PM

Quant Time: May 27 11:35:20 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF052615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 27 01:32:46 2015
Response via : Initial Calibration

