

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF052616\
 Data File : BF087437.D
 Acq On : 27 May 2016 9:38
 Operator : UM/SJ
 Sample : H3218-06DL 5X
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 E-4(12.5-13)DL

Quant Time: May 27 11:56:34 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF052316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 25 16:26:52 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.51	152	106945	20.00	ng	-0.01
21) Naphthalene-d8	9.54	136	443115	20.00	ng	-0.01
38) Acenaphthene-d10	12.37	164	203231	20.00	ng	-0.02
63) Phenanthrene-d10	14.77	188	367150	20.00	ng	-0.01
75) Chrysene-d12	18.47	240	258862	20.00	ng	-0.01
86) Perylene-d12	20.14	264	220594	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	161585	26.55	ng	0.01
7) Phenol-d6	7.06	99	236165	28.72	ng	-0.01
23) Nitrobenzene-d5	8.43	82	138067	15.70	ng	-0.01
41) 2,4,6-Tribromophenol	13.67	330	44443	22.76	ng	-0.01
44) 2-Fluorobiphenyl	11.31	172	219564	15.23	ng	-0.02
78) Terphenyl-d14	17.12	244	178392	14.65	ng	-0.01
Target Compounds						Qvalue
31) Naphthalene	9.58	128	87027	3.92	ng	99
49) Dimethylphthalate	12.01	163	88964	5.46	ng	99
51) Acenaphthene	12.42	154	119329	9.37	ng	97
57) Fluorene	13.27	166	78348	5.24	ng	97
70) Phenanthrene	14.80	178	471416	23.18	ng	99
71) Anthracene	14.90	178	121420	5.91	ng	98
74) Fluoranthene	16.57	202	157269	7.71	ng	96
77) Pyrene	16.88	202	221639	11.89	ng	98
80) Benzo(a)anthracene	18.46	228	56746	3.85	ng	98
82) Chrysene	18.51	228	53541	3.66	ng	96
89) Benzo(a)pyrene	20.10	252	34332	2.82	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF052616\
Data File : BF087437.D
Acq On : 27 May 2016 9:38
Operator : UM/SJ
Sample : H3218-06DL 5X
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
E-4(12.5-13)DL

Quant Time: May 27 11:56:34 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF052316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 25 16:26:52 2016
Response via : Initial Calibration

