


Data Path : Z:\HPCHEM1\BNA F\DATA\BF052616\
 Data File : BF087429.D
 Acq On : 27 May 2016 3:47
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF052316.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri May 27 17:03:59 2016



AutoFind: Scans 1200, 1201, 1202; Background Corrected with Scan 1197

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	62.2	17714	PASS
68	69	0.00	2	1.1	142	PASS
69	198	0.00	100	45.7	13027	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	54.3	15469	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	28477	PASS
199	198	5	9	6.8	1934	PASS
275	198	10	60	24.4	6958	PASS
365	198	1	100	3.3	941	PASS
441	442	0.01	24	15.9	2832	PASS
442	198	50	100	62.7	17864	PASS
443	442	15	24	18.5	3299	PASS