

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052617\
 Data File : BF095456.D
 Acq On : 26 May 2017 12:37
 Operator : SJ/MA
 Sample : I3241-15
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EP1-WC4(0-9.5)

Manual Integrations
 APPROVED

mohammad
 5/30/2017 1:35:51 PM

Quant Time: May 26 16:53:16 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 26 16:50:16 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	91075	20.00	ng	0.00
21) Naphthalene-d8	9.77	136	369835	20.00	ng	0.00
38) Acenaphthene-d10	12.59	164	170285	20.00	ng	0.00
63) Phenanthrene-d10	15.00	188	333103	20.00	ng	0.00
75) Chrysene-d12	18.68	240	290078	20.00	ng	0.00
86) Perylene-d12	20.35	264	192845	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.71	112	611531	111.95	ng	0.00
7) Phenol-d6	7.29	99	762464	116.33	ng	0.00
23) Nitrobenzene-d5	8.64	82	451713	77.02	ng	-0.01
41) 2,4,6-Tribromophenol	13.89	330	167596	120.18	ng	0.00
44) 2-Fluorobiphenyl	11.53	172	817827	69.13	ng	0.00
78) Terphenyl-d14	17.32	244	746007	56.64	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	12.24	163	105465	7.54	ng	100
70) Phenanthrene	15.04	178	61414	3.21	ng	96
74) Fluoranthene	16.79	202	110911	5.65	ng	97
77) Pyrene	17.09	202	101171	4.66	ng	96
80) Benzo(a)anthracene	18.67	228	36306	2.15	ng	97
82) Chrysene	18.71	228	44740	2.74	ng	98
87) Benzo(b)fluoranthene	19.96	252	33099m	2.78	ng	
89) Benzo(a)pyrene	20.31	252	23583	2.14	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052617\
Data File : BF095456.D
Acq On : 26 May 2017 12:37
Operator : SJ/MA
Sample : I3241-15
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EP1-WC4(0-9.5)

Manual Integrations
APPROVED

mohammad
5/30/2017 1:35:51 PM

Quant Time: May 26 16:53:16 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 26 16:50:16 2017
Response via : Initial Calibration

