

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052617\
 Data File : BF095464.D
 Acq On : 26 May 2017 16:54
 Operator : SJ/MA
 Sample : I3222-17
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 BLIND-DUPE-X

Quant Time: May 26 18:04:10 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 26 16:50:16 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	91929	20.00	ng	0.00
21) Naphthalene-d8	9.77	136	367528	20.00	ng	0.00
38) Acenaphthene-d10	12.60	164	164216	20.00	ng	0.00
63) Phenanthrene-d10	15.00	188	248529	20.00	ng	0.00
75) Chrysene-d12	18.68	240	194803	20.00	ng	0.00
86) Perylene-d12	20.35	264	153117	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.71	112	647265	117.39	ng	0.00
7) Phenol-d6	7.29	99	774057	117.00	ng	0.00
23) Nitrobenzene-d5	8.64	82	422821	72.55	ng	-0.01
41) 2,4,6-Tribromophenol	13.89	330	123912	92.14	ng	0.00
44) 2-Fluorobiphenyl	11.53	172	842786	73.87	ng	0.00
78) Terphenyl-d14	17.32	244	519198	58.70	ng	0.00
Target Compounds						
49) Dimethylphthalate	12.24	163	112935	8.38	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052617\
Data File : BF095464.D
Acq On : 26 May 2017 16:54
Operator : SJ/MA
Sample : I3222-17
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
BLIND-DUPE-X

Quant Time: May 26 18:04:10 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 26 16:50:16 2017
Response via : Initial Calibration

