

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052617\
 Data File : BF095472.D
 Acq On : 26 May 2017 21:15
 Operator : SJ/MA
 Sample : I3244-02
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 NM3-COMP-6-18

Manual Integrations
 APPROVED

mohammad
 5/30/2017 1:36:35 PM

Quant Time: May 27 04:25:20 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 26 16:50:16 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	95626	20.00	ng	0.00
21) Naphthalene-d8	9.77	136	364710	20.00	ng	0.00
38) Acenaphthene-d10	12.60	164	160055	20.00	ng	0.00
63) Phenanthrene-d10	14.99	188	223128	20.00	ng	0.00
75) Chrysene-d12	18.67	240	199904	20.00	ng	0.00
86) Perylene-d12	20.35	264	143940	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.71	112	618958	107.91	ng	0.00
7) Phenol-d6	7.29	99	752490	109.34	ng	0.00
23) Nitrobenzene-d5	8.65	82	407848	70.52	ng	0.00
41) 2,4,6-Tribromophenol	13.89	330	110661	84.42	ng	0.00
44) 2-Fluorobiphenyl	11.54	172	720682	64.81	ng	0.00
78) Terphenyl-d14	17.32	244	435700	48.01	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	12.24	163	91459	6.96	ng	100
70) Phenanthrene	15.04	178	514174	40.11	ng	97
71) Anthracene	15.12	178	92164	7.04	ng	97
74) Fluoranthene	16.78	202	763560	58.06	ng	99
77) Pyrene	17.08	202	680131	45.49	ng	98
80) Benzo(a)anthracene	18.66	228	328748	28.26	ng	98
82) Chrysene	18.71	228	304218	27.04	ng	97
85) Indeno(1,2,3-cd)pyrene	21.69	276	78412	8.58	ng	92
87) Benzo(b)fluoranthene	19.95	252	247857m	27.87	ng	
88) Benzo(k)fluoranthene	19.97	252	98718m	11.51	ng	
89) Benzo(a)pyrene	20.29	252	153241	18.67	ng	97
90) Dibenzo(a,h)anthracene	21.68	278	21397	3.16	ng	# 89
91) Benzo(a,h,i)perylene	22.08	276	65146	9.79	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052617\
Data File : BF095472.D
Acq On : 26 May 2017 21:15
Operator : SJ/MA
Sample : I3244-02
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
NM3-COMP-6-18

Manual Integrations
APPROVED

mohammad
5/30/2017 1:36:35 PM

Quant Time: May 27 04:25:20 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 26 16:50:16 2017
Response via : Initial Calibration

