

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052622\
 Data File : BF128484.D
 Acq On : 26 May 2022 20:26
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF052622

Quant Time: May 27 02:02:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 27 01:51:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	92	0.00
2	1,4-Dioxane	40.000	39.312	1.7	92	0.00
3	Pyridine	40.000	39.309	1.7	93	0.00
4	n-Nitrosodimethylamine	40.000	39.985	0.0	92	0.00
5 S	2-Fluorophenol	80.000	77.089	3.6	93	0.00
6	Aniline	40.000	39.996	0.0	91	0.00
7 S	Phenol-d6	80.000	77.483	3.1	94	0.00
8	2-Chlorophenol	40.000	39.186	2.0	94	0.00
9	Benzaldehyde	40.000	47.564	-18.9	102	0.00
10 C	Phenol	40.000	39.157	2.1	95	0.00
11	bis(2-Chloroethyl)ether	40.000	38.267	4.3	93	0.00
12	1,3-Dichlorobenzene	40.000	39.164	2.1	94	0.00
13 C	1,4-Dichlorobenzene	40.000	39.246	1.9	94	0.00
14	1,2-Dichlorobenzene	40.000	39.029	2.4	95	0.00
15	Benzyl Alcohol	40.000	39.924	0.2	94	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.580	3.6	92	0.00
17	2-Methylphenol	40.000	39.059	2.4	93	0.00
18	Hexachloroethane	40.000	40.252	-0.6	96	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.771	5.6	93	0.00
20	3+4-Methylphenols	40.000	38.867	2.8	95	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	95	0.00
22	Acetophenone	40.000	38.316	4.2	95	0.00
23 S	Nitrobenzene-d5	80.000	82.802	-3.5	97	0.00
24	Nitrobenzene	40.000	40.752	-1.9	96	0.00
25	Isophorone	40.000	38.471	3.8	93	0.00
26 C	2-Nitrophenol	40.000	44.895	-12.2	101	0.00
27	2,4-Dimethylphenol	40.000	39.136	2.2	95	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.337	4.2	94	0.00
29 C	2,4-Dichlorophenol	40.000	39.121	2.2	94	0.00
30	1,2,4-Trichlorobenzene	40.000	38.771	3.1	94	0.00
31	Naphthalene	40.000	38.721	3.2	96	0.00
32	Benzoic acid	40.000	45.937	-14.8	101	0.00
33	4-Chloroaniline	40.000	38.263	4.3	92	0.00
34 C	Hexachlorobutadiene	40.000	38.980	2.6	94	0.00
35	Caprolactam	40.000	39.891	0.3	94	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.216	2.0	95	0.00
37	2-Methylnaphthalene	40.000	38.819	3.0	96	0.00
38	1-Methylnaphthalene	40.000	38.113	4.7	95	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	94	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.227	1.9	97	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.997	-2.5	94	0.00
42 S	2,4,6-Tribromophenol	80.000	81.075	-1.3	97	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.232	-0.6	96	0.00
44	2,4,5-Trichlorophenol	40.000	39.184	2.0	94	0.00
45 S	2-Fluorobiphenyl	80.000	75.762	5.3	96	0.00
46	1,1'-Biphenyl	40.000	38.643	3.4	96	0.00
47	2-Chloronaphthalene	40.000	38.878	2.8	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052622\
 Data File : BF128484.D
 Acq On : 26 May 2022 20:26
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF052622

Quant Time: May 27 02:02:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 27 01:51:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.249	-13.1	102	0.00
49	Acenaphthylene	40.000	39.062	2.3	96	0.00
50	Dimethylphthalate	40.000	38.726	3.2	95	0.00
51	2,6-Dinitrotoluene	40.000	44.957	-12.4	101	0.00
52 C	Acenaphthene	40.000	39.214	2.0	96	0.00
53	3-Nitroaniline	40.000	44.184	-10.5	100	0.00
54 P	2,4-Dinitrophenol	40.000	41.573	-3.9	105	0.00
55	Dibenzofuran	40.000	39.347	1.6	96	0.00
56 P	4-Nitrophenol	40.000	44.277	-10.7	98	0.00
57	2,4-Dinitrotoluene	40.000	46.944	-17.4	100	0.00
58	Fluorene	40.000	38.261	4.3	96	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.270	-0.7	97	0.00
60	Diethylphthalate	40.000	39.471	1.3	95	0.00
61	4-Chlorophenyl-phenylether	40.000	38.480	3.8	96	0.00
62	4-Nitroaniline	40.000	44.258	-10.6	98	0.00
63	Azobenzene	40.000	39.195	2.0	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	95	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.500	-3.8	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.851	2.9	95	0.00
67	4-Bromophenyl-phenylether	40.000	39.894	0.3	97	0.00
68	Hexachlorobenzene	40.000	38.915	2.7	95	0.00
69	Atrazine	40.000	39.064	2.3	95	0.00
70 C	Pentachlorophenol	40.000	41.708	-4.3	95	0.00
71	Phenanthrene	40.000	38.387	4.0	95	0.00
72	Anthracene	40.000	38.850	2.9	97	0.00
73	Carbazole	40.000	38.817	3.0	95	0.00
74	Di-n-butylphthalate	40.000	39.280	1.8	95	0.00
75 C	Fluoranthene	40.000	39.024	2.4	95	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	95	0.00
77	Benzydine	40.000	33.683	15.8	75	0.00
78	Pyrene	40.000	40.043	-0.1	97	0.00
79 S	Terphenyl-d14	80.000	78.181	2.3	96	0.00
80	Butylbenzylphthalate	40.000	41.315	-3.3	96	0.00
81	Benzo(a)anthracene	40.000	39.020	2.4	97	0.00
82	3,3'-Dichlorobenzidine	40.000	38.094	4.8	91	0.00
83	Chrysene	40.000	39.412	1.5	95	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.661	-1.7	97	0.00
85 c	Di-n-octyl phthalate	40.000	40.308	-0.8	95	0.00
86 I	Perylene-d12	20.000	20.000	0.0	93	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.836	0.4	94	0.00
88	Benzo(b)fluoranthene	40.000	41.345	-3.4	93	0.00
89	Benzo(k)fluoranthene	40.000	38.612	3.5	94	0.00
90 C	Benzo(a)pyrene	40.000	39.251	1.9	92	0.00
91	Dibenzo(a,h)anthracene	40.000	39.491	1.3	93	0.00
92	Benzo(g,h,i)perylene	40.000	39.246	1.9	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052622\
Data File : BF128484.D
Acq On : 26 May 2022 20:26
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF052622

Quant Time: May 27 02:02:15 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 27 01:51:59 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0