

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF052715\
 Data File : BF079535.D
 Acq On : 27 May 2015 23:21
 Operator : TP/IZ
 Sample : G2359-10
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-31BS

Quant Time: May 28 00:59:30 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF052615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 28 00:48:25 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.00	152	74529	20.00	ng	0.00
21) Naphthalene-d8	8.58	136	309619	20.00	ng	0.00
38) Acenaphthene-d10	10.74	164	151095	20.00	ng	0.00
63) Phenanthrene-d10	12.58	188	287649	20.00	ng	0.00
75) Chrysene-d12	15.86	240	283863	20.00	ng	-0.08
86) Perylene-d12	17.57	264	329324	20.00	ng	-0.32

System Monitoring Compounds						
5) 2-Fluorophenol	5.28	112	388301	90.37	ng	0.00
7) Phenol-d6	6.59	99	553161	101.00	ng	0.00
23) Nitrobenzene-d5	7.70	82	336498	62.82	ng	0.00
41) 2,4,6-Tribromophenol	11.73	330	169301	102.01	ng	-0.01
44) 2-Fluorobiphenyl	9.93	172	669323	63.20	ng	0.00
78) Terphenyl-d14	14.58	244	700477	57.91	ng	-0.01


Target Compounds				Qvalue		
31) Naphthalene	8.60	128	147869	9.95	ng	98
37) 2-Methylnaphthalene	9.46	142	46860	4.54	ng	# 93
48) Acenaphthylene	10.57	152	61392	3.97	ng	98
49) Dimethylphthalate	10.43	163	26334	2.37	ng	98
51) Acenaphthene	10.79	154	129615	14.65	ng	100
54) Dibenzofuran	11.01	168	139243	10.67	ng	92
55) 4-Nitrophenol	11.01	139	52890	32.97	ng	# 14
57) Fluorene	11.43	166	162212	15.08	ng	97
70) Phenanthrene	12.62	178	1451517	91.98	ng	97
71) Anthracene	12.67	178	471761	29.45	ng	99
72) Carbazole	12.89	167	174962	11.34	ng	98
73) Di-n-butylphthalate	13.35	149	7865	Below Cal	#	1
74) Fluoranthene	14.09	202	1651468	97.59	ng	94
77) Pyrene	14.37	202	1546248	87.92	ng	# 95
80) Benzo(a)anthracene	15.85	228	915995	56.09	ng	98
82) Chrysene	15.90	228	734201	47.30	ng	96
85) Indeno(1,2,3-cd)pyrene	19.26	276	430887	21.58	ng	96
87) Benzo(b)fluoranthene	17.13	252	1315150	70.02	ng	# 96
88) Benzo(k)fluoranthene	17.13	252	1317617	78.45	ng	# 94
89) Benzo(a)pyrene	17.51	252	768889	46.76	ng	98
90) Dibenzo(a,h)anthracene	19.04	278	112149	5.16	ng	96
91) Benzo(g,h,i)perylene	19.26	276	430887	24.70	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF052715\
Data File : BF079535.D
Acq On : 27 May 2015 23:21
Operator : TP/IZ
Sample : G2359-10
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-31BS

Quant Time: May 28 00:59:30 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF052615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 28 00:48:25 2015
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.00 min Scan# 509
Delta R.T. 0.00 min
Lab File: BF079535.D
Acq: 27 May 2015 23:21

Instrument :
BNA_F
ClientSampleId :
SB-31BS




Tgt Ion:152 Resp: 74529
Ion Ratio Lower Upper
152 100
150 152.1 124.0 186.0
115 50.6 49.6 74.4




#5

2-Fluorophenol
Concen: 90.37 ng
RT: 5.28 min Scan# 358
Delta R.T. 0.00 min
Lab File: BF079535.D
Acq: 27 May 2015 23:21

Tgt Ion:112 Resp: 388301
Ion Ratio Lower Upper
112 100
64 61.6 43.5 65.3
63 30.7 21.8 32.8





#7

Phenol-d6

Concen: 101.00 ng

RT: 6.59 min Scan# 473

Delta R.T. 0.00 min

Lab File: BF079535.D

Acq: 27 May 2015 23:21

Instrument :

BNA_F

ClientSampleId :

SB-31BS


Tgt Ion: 99 Resp: 553161

Ion Ratio Lower Upper

99 100

42 18.1 11.9 17.9#




71 31.1 22.1 33.1



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 8.58 min Scan# 647

Delta R.T. 0.00 min

Lab File: BF079535.D

Acq: 27 May 2015 23:21

Tgt Ion: 136 Resp: 309619


Ion Ratio Lower Upper

136 100

137 10.4 8.5 12.7

54 5.6 6.6 9.8#

68 4.6 4.7 7.1#






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0









#23
 Nitrobenzene-d5
 Concen: 62.82 ng
 RT: 7.70 min Scan# 570
 Delta R.T. 0.00 min
 Lab File: BF079535.D
 Acq: 27 May 2015 23:21

Instrument :
 BNA_F
 ClientSampleId :
 SB-31BS

Tgt Ion: 82 Resp: 336498
 Ion Ratio Lower Upper
 82 100
 128 48.2 30.8 46.2#
 54 48.2 42.8 64.2




Abundance Ion 82.00 (81.70 to 82.70): BF0
 Ion 128.00 (127.70 to 128.70): E
 Ion 54.00 (53.70 to 54.70): BF0






#31
 Naphthalene
 Concen: 9.95 ng
 RT: 8.60 min Scan# 649
 Delta R.T. 0.00 min
 Lab File: BF079535.D
 Acq: 27 May 2015 23:21

Tgt Ion: 128 Resp: 147869
 Ion Ratio Lower Upper
 128 100
 129 10.6 9.3 13.9
 127 12.8 10.6 15.8



Abundance Ion 128.00 (127.70 to 128.70): E
 Ion 129.00 (128.70 to 129.70): E
 Ion 127.00 (126.70 to 127.70): E









#37
 2-Methylnaphthalene
 Concen: 4.54 ng
 RT: 9.46 min Scan# 724
 Delta R.T. 0.00 min
 Lab File: BF079535.D
 Acq: 27 May 2015 23:21




Instrument :
 BNA_F
 ClientSampleId :
 SB-31BS


Tgt Ion:142 Resp: 46860
 Ion Ratio Lower Upper
 142 100
 141 88.5 69.5 104.3
 115 42.0 23.8 35.6#



#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 10.74 min Scan# 836
 Delta R.T. 0.00 min
 Lab File: BF079535.D
 Acq: 27 May 2015 23:21

Tgt Ion:164 Resp: 151095
 Ion Ratio Lower Upper
 164 100
 162 100.8 82.6 124.0
 160 45.9 37.7 56.5









#41
 2,4,6-Tribromophenol
 Concen: 102.01 ng
 RT: 11.73 min Scan# 922
 Delta R.T. -0.01 min
 Lab File: BF079535.D
 Acq: 27 May 2015 23:21




Instrument :
 BNA_F
 ClientSampleId :
 SB-31BS


Tgt Ion:330 Resp: 169301
 Ion Ratio Lower Upper
 330 100
 332 95.1 0.0 0.0#
 141 33.7 0.0 0.0#



#44
 2-Fluorobiphenyl
 Concen: 63.20 ng
 RT: 9.93 min Scan# 765
 Delta R.T. 0.00 min
 Lab File: BF079535.D
 Acq: 27 May 2015 23:21

Tgt Ion:172 Resp: 669323
 Ion Ratio Lower Upper
 172 100
 171 34.9 28.5 42.7
 170 23.4 19.7 29.5





#48

Acenaphthylene

Concen: 3.97 ng

RT: 10.57 min Scan# 821

Delta R.T. 0.00 min

Lab File: BF079535.D

Acq: 27 May 2015 23:21

Instrument :

BNA_F

ClientSampleId :

SB-31BS


Tgt Ion:152 Resp: 61392

Ion Ratio Lower Upper

152 100

151 19.3 15.8 23.8




153 14.5 10.7 16.1



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#49

Dimethylphthalate

Concen: 2.37 ng

RT: 10.43 min Scan# 809

Delta R.T. -0.01 min

Lab File: BF079535.D

Acq: 27 May 2015 23:21


Tgt Ion:163 Resp: 26334

Ion Ratio Lower Upper

163 100

194 3.0 2.4 3.6



164 9.5 8.2 12.4




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E





#51

Acenaphthene

Concen: 14.65 ng

RT: 10.79 min Scan# 840

Delta R.T. 0.00 min

Lab File: BF079535.D

Acq: 27 May 2015 23:21

Instrument :

BNA_F

ClientSampleId :

SB-31BS


Tgt Ion:154 Resp: 129615

Ion Ratio Lower Upper

154 100

153 113.7 91.1 136.7




152 55.5 44.6 66.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#54

Dibenzofuran

Concen: 10.67 ng

RT: 11.01 min Scan# 859

Delta R.T. 0.00 min

Lab File: BF079535.D

Acq: 27 May 2015 23:21


Tgt Ion:168 Resp: 139243

Ion Ratio Lower Upper

168 100

139 34.4 32.9 49.3



169 13.3 11.1 16.7




Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E





#55

4-Nitrophenol

Concen: 32.97 ng

RT: 11.01 min Scan# 859

Delta R.T. 0.03 min

Lab File: BF079535.D

Acq: 27 May 2015 23:21

Instrument :

BNA_F

ClientSampled :

SB-31BS


Tgt Ion:139 Resp: 52890

Ion Ratio Lower Upper

139 100

109 1.2 42.8 82.8#




65 2.5 66.3 106.3#



Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF0



#57

Fluorene

Concen: 15.08 ng

RT: 11.43 min Scan# 896

Delta R.T. 0.00 min

Lab File: BF079535.D

Acq: 27 May 2015 23:21


Tgt Ion:166 Resp: 162212

Ion Ratio Lower Upper

166 100

165 101.2 78.6 117.8



167 14.2 10.8 16.2




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E





#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 12.58 min Scan# 997

Delta R.T. 0.00 min

Lab File: BF079535.D

Acq: 27 May 2015 23:21

Instrument :

BNA_F

ClientSampleId :

SB-31BS


Tgt Ion:188 Resp: 287649

Ion Ratio Lower Upper

188 100

94 10.1 9.8 14.6




80 10.7 10.5 15.7



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF0

Ion 80.00 (79.70 to 80.70): BF0



#70

Phenanthrene

Concen: 91.98 ng

RT: 12.62 min Scan# 1000

Delta R.T. 0.00 min

Lab File: BF079535.D

Acq: 27 May 2015 23:21


Tgt Ion:178 Resp: 1451517

Ion Ratio Lower Upper

178 100

176 21.4 15.7 23.5



179 16.4 12.4 18.6




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#71

Anthracene

Concen: 29.45 ng

RT: 12.67 min Scan# 1005

Delta R.T. 0.00 min

Lab File: BF079535.D

Acq: 27 May 2015 23:21

Instrument :

BNA_F

ClientSampled :

SB-31BS


Tgt Ion:178 Resp: 471761

Ion Ratio Lower Upper

178 100

176 19.3 15.3 22.9




179 16.2 12.2 18.4



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Carbazole

Concen: 11.34 ng

RT: 12.89 min Scan# 1024

Delta R.T. 0.00 min

Lab File: BF079535.D

Acq: 27 May 2015 23:21


Tgt Ion:167 Resp: 174962

Ion Ratio Lower Upper

167 100

166 21.1 17.0 25.4



139 12.5 11.2 16.8




Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E





#74

Fluoranthene

Concen: 97.59 ng

RT: 14.09 min Scan# 1129

Delta R.T. 0.00 min

Lab File: BF079535.D

Acq: 27 May 2015 23:21

Instrument :

BNA_F

ClientSampleId :

SB-31BS


Tgt Ion:202 Resp: 1651468

Ion Ratio Lower Upper




202 100

101 13.9 0.0 29.3

203 18.9 0.0 37.6



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E



#75

Chrysene-d12

Concen: 20.00 ng

RT: 15.86 min Scan# 1284

Delta R.T. -0.08 min

Lab File: BF079535.D

Acq: 27 May 2015 23:21


Tgt Ion:240 Resp: 283863

Ion Ratio Lower Upper



240 100


120 15.7 10.8 16.2

236 26.0 21.0 31.6



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E









#77
 Pyrene
 Concen: 87.92 ng
 RT: 14.37 min Scan# 1153
 Delta R.T. 0.00 min
 Lab File: BF079535.D
 Acq: 27 May 2015 23:21




Instrument :
 BNA_F
 ClientSampleId :
 SB-31BS


Tgt Ion:202 Resp: 1546248
 Ion Ratio Lower Upper
 202 100
 200 23.0 17.5 26.3
 203 22.1 14.4 21.6#



#78
 Terphenyl-d14
 Concen: 57.91 ng
 RT: 14.58 min Scan# 1172
 Delta R.T. -0.01 min
 Lab File: BF079535.D
 Acq: 27 May 2015 23:21

Tgt Ion:244 Resp: 700477
 Ion Ratio Lower Upper
 244 100
 212 6.4 5.2 7.8
 122 8.9 8.0 12.0





#80

Benzo(a)anthracene

Concen: 56.09 ng

RT: 15.85 min Scan# 1283

Delta R.T. -0.08 min

Lab File: BF079535.D

Acq: 27 May 2015 23:21

Instrument :

BNA_F

ClientSampleId :

SB-31BS


Tgt Ion:228 Resp: 915995

Ion Ratio Lower Upper

228 100

226 29.7 22.8 34.2




229 21.3 15.9 23.9



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#82

Chrysene

Concen: 47.30 ng

RT: 15.90 min Scan# 1287

Delta R.T. -0.08 min

Lab File: BF079535.D

Acq: 27 May 2015 23:21


Tgt Ion:228 Resp: 734201

Ion Ratio Lower Upper

228 100

226 31.9 25.0 37.4



229 23.3 15.9 23.9




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 21.58 ng
 RT: 19.26 min Scan# 1581
 Delta R.T. -0.05 min
 Lab File: BF079535.D
 Acq: 27 May 2015 23:21




Instrument :
 BNA_F
 ClientSampleId :
 SB-31BS


Tgt Ion:276 Resp: 430887
 Ion Ratio Lower Upper
 276 100
 138 24.4 21.4 32.2
 277 25.8 19.8 29.6



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 17.57 min Scan# 1433
 Delta R.T. -0.32 min
 Lab File: BF079535.D
 Acq: 27 May 2015 23:21

Tgt Ion:264 Resp: 329324
 Ion Ratio Lower Upper
 264 100
 260 25.1 19.8 29.8
 265 23.2 17.8 26.6









#87
Benzo(b)fluoranthene
Concen: 70.02 ng
RT: 17.13 min Scan# 1395
Delta R.T. -0.25 min
Lab File: BF079535.D
Acq: 27 May 2015 23:21




Instrument :
BNA_F
ClientSampleId :
SB-31BS


Tgt Ion:252 Resp: 1315150
Ion Ratio Lower Upper
252 100
253 23.4 17.6 26.4
125 12.2 7.8 11.6#



#88
Benzo(k)fluoranthene
Concen: 78.45 ng
RT: 17.13 min Scan# 1395
Delta R.T. -0.29 min
Lab File: BF079535.D
Acq: 27 May 2015 23:21

Tgt Ion:252 Resp: 1317617
Ion Ratio Lower Upper
252 100
253 23.4 17.2 25.8
125 12.2 6.8 10.2#









#89
Benzo(a)pyrene
Concen: 46.76 ng
RT: 17.51 min Scan# 1428
Delta R.T. -0.30 min
Lab File: BF079535.D
Acq: 27 May 2015 23:21




Instrument :
BNA_F
ClientSampleId :
SB-31BS


Tgt Ion: 252 Resp: 768889
Ion Ratio Lower Upper
252 100
253 23.6 17.8 26.8
125 9.6 7.9 11.9



#90
Dibenzo(a,h)anthracene
Concen: 5.16 ng
RT: 19.04 min Scan# 1562
Delta R.T. -0.29 min
Lab File: BF079535.D
Acq: 27 May 2015 23:21

Tgt Ion: 278 Resp: 112149
Ion Ratio Lower Upper
278 100
139 17.4 12.8 19.2
279 25.8 19.1 28.7





#91

Benzo(g,h,i)perylene

Concen: 24.70 ng

RT: 19.26 min Scan# 1581

Delta R.T. -0.43 min

Lab File: BF079535.D

Acq: 27 May 2015 23:21

Instrument :

BNA_F

ClientSampleId :

SB-31BS


Tgt Ion: 276 Resp: 430887

Ion Ratio Lower Upper

276 100

277 23.8 19.1 28.7

138 26.4 20.1 30.1



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

